

CERTIFICATE

It is hereby confirmed that the

Fume Cupboard T120

of the manufacturer

Renggli AG
Industrie-Ost
Birkenstrasse 31
6343 Rotkreuz
Switzerland

was tested according to

EN 14175

part 3


at the total volume flow rate of **480 m³/h**,
with a nominal face velocity of **0.34 m/s**
With this exhaust volume flow rate the requirements of
- **BG Chemie, Germany, issued 29.07.2003**
are completely fulfilled

The details and data of the complete test procedure
are content of the type test report

FCS/EN/93/2006

which is given to the manufacturer on
12. June 2006

Tested by:
Tintschl BioEnergie und
Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen


Dipl.-Ing. Bernd Schubert
Erlangen, 11.07.2014



BioEnergie und Strömungstechnik AG