

VINNO6



Data sheet
V1.11



Contents

1. System Overview	5
1.1 Architecture	5
1.2 Applications	6
1.3 Imaging features	6
1.4 Standard features	7
1.5 Language support	8
2. Ergonomics	8
2.1 Keyboard	9
2.2 Image display screen	9
2.3 Touch gestures	9
2.4 Comments	9
2.5 Bodymark	9
2.6 Peripherals	9
2.7 Dimensions and Weight	10
2.8 Electrical Power	10
2.9 Operating Environment	10
2.10 Storage & Transportation Environment	10
3. Transducers	10
3.1 Transducer Technology	10
3.2 Transducer types	10
3.3 Transducer selection	10
4. Advanced Imaging controls	15
4.1 VFusion	15
4.2 VSpeckle	15
4.3 VTissue	16
4.4 Tissue Doppler (TD)	16
4.5 Tissue Velocity Imaging (TVI)(optional)*	16
4.6 3D/4D	16
4.6.1 HQ (optional)*	16
4.6.2 Spatio-Temporal Image Correlation (STIC) (optional)*	16
4.6.3 3D Smart Face(optional)*	16
4.7 Smart operation	17
4.7.1 Free View(optional)*	17
4.7.2 Inversion mode(optional)*	17
4.7.3 Magic Cut(optional)*	17
4.8 Smart 3D	17
4.8.1 Volume Measurement(optional)*	17
4.8.2 Auto Follicle(2D/3D)(optional)*	17
4.9 Tissue Velocity M mode(TVM)(optional) *	17



4.11 Elastic Imaging(EI) (optional)*	17
4.12 Next generation RF-based image processing	18
4.13 Stress Echo(optional)*	18
4.14 Strain Imaging(optional)*	18
5. Imaging modes	18
5.1 2D Imaging	18
5.2 Harmonic Imaging	19
5.3 M mode	19
5.4 Color Doppler mode	19
5.5 Power Doppler mode	20
5.6 Pulsed Wave (PW) Doppler	20
5.7 Continuous Wave Doppler (CWD)(optinal)*	21
5.8 Pview(optional)*	21
5.9 Tview	21
5.10 Auto	21
5.11 Tissue Doppler Imaging (TD)	22
5.12 Tissue Velocity Imaging (TVI) (optional)*	22
5.13 Tissue Velocity M mode (TVM) (optional)*	22
5.14 Elastography imaging(optional)*	23
5.15 3D/4D	23
6. Touch Panel Interface	24
6.1 2D mode	24
6.2 M Mode	25
6.3 CF mode	25
6.4 PW/CW mode	25
6.5 3D mode	26
6.6 4D mode	26
7. System Feature	26
7.1 Display modes	26
7.2 Display annotation	27
7.3 Cineloop	27
7.4 Compare	28
7.5 Physio	28
7.6 Archive	28
7.7 Report	28
7.8 Connectivity	28
7.9 Probes/application	29
7.10 Safety Conformance	29
8. Measurement and Analysis	29
8.1 Measurement in different modes	29
8.1.1 Generic Measurement in 2D mode	30
8.1.2 Generic Measurement in M mode	30



8.1.3 Generic Measurement in PW mode	30
8.2 Measurement in different applications	31
8.2.1 Abdominal Measurement	31
8.2.2 Small Part Measurement	31
8.2.3 Vessel Measurement	31
8.2.4 Gynecology Measurement	31
8.2.5 Urology Measurement	31
8.2.6 Pediatric Measurement	31
8.2.7 Obstetrics Measurement	31
8.2.8 Cardiac Measurement	32
8.2.9 Auto NT (Nuchal Translucency measurement)(Optional)*	32
8.2.10 Smart 3D Volume Measurement(Optional)*	32
8.2.11 Live IMT (Intima-Media Thickness) measurement (Optional) *	32
8.2.12 Auto IT (Intracranial translucency) measurement(Optional) *	32
8.2.13 Auto Follicle(2D/3D)(Optional) *	32
8.2.14 Smart 3D Volume Measurement(Optional) *	33
8.2.15 VAim OB measurement (Optional) *	33
8.2.16 VAim Hip measurement	33
(Optional) *	33
8.2.17 VAim Follicle (2D) measurement (Optional) *	33



VINNO⁶

Ultrasound System Specifications

The premium performance of the full functional Portable VINNO 6 provides a fast and easy diagnosis by:

- Ultra-premium contrast and resolution imaging benefited from the first RF platform of the world
- All ranges of features, functions and probes
- Ease of use and ergonomic design



1. System Overview

1.1 Architecture

- The revolutionary RF platform, The First In The World, allows for more accurate information. This platform transfers all RF data for computing without any information loss. It has a much better advantage in detail imaging than current advanced platforms.
- Thanks to the RF platform, it allows the development of many RF-based processing algorithms, which have ultra-premium contrast and resolution imaging

- This unique platform is capable of processing multiple data streams simultaneously
- World-class Up to 23MHz imaging performance
- Directional-enhanced information compiling for more tissue detail and reduction of angle-generated artifacts
- Next generation adaptive image processing for noise and artifact reduction that improves tissue presentation and edge definition
- Fully independent, triplex multiple mode operation for easy in Doppler procedures
- Multi-processors allow simultaneous mode changes and support for advanced system



functionality

- Zone Imaging technology can obtain high resolution and good penetration in the whole image zone through the adaptive dynamic beam control from the near field to the far field
- World-class design to be thinner and lighter
- VLuminous Flow provides the color Doppler flow innovatively in a 3D view with excellent sensitivity, which can help understand the structure of blood flow and small vessels intuitively
- Sync ROI enables the width of 2D scan area is synchronized with the CF ROI, which effectively improves the frame rate
- Diverse customized tools make V6 a truly elite unit, which enhances efficiency dramatically
- Zscore analysis, provide a new way for fetal heart evaluation
- Support to export 3D data for 3D printer
- Support multiple DICOM server configuration
- Background transfer, supports background export without interrupting the actual scan
- VReport, a customer-centric tool for report templates design, makes the whole report procedure more smooth and individual
- Customized user interface, allows user to change the position of buttons

on the touch screen, also the size of ‘probe&app’ UI window is adjustable

- VWork, an intelligent feature, which enables users to configure workflows for every application scenario. This leads to easy and effective adherence to a department protocol and saves operation time to a great extent
- Less 50 sec boot up time for easy mobile ability

1.2 Applications

- Abdomen 1.2
- Obstetric
- Gynecology
- Cardiology 1.1
- Urology
- Vascular 1.3
- TCD
- Small Parts
- Pediatrics
- Intra-operative

Abdominalinis, kardiologinis, kraujagyslių

1.3 Imaging features

- 2D grayscale imaging
- Harmonic imaging both in tissue harmonic and pulse inversion harmonic technologies 3.9
- VFusion, directional-enhanced information compounding
- Vspeckle , specialized and adaptive imaging processing to remove speckle

Harmoninis vaizdavimas tiek audinių harmonikų, tiek impulsų inversijos harmonikų technologijose



noise artifacts and enhance tissue edge for clarity and accuracy

- VTissue, the advanced adaptive image processing to compensate for sound and speed variation in different tissue
- VFlow, adaptive color flow filter to increase the sensitivity of blood flow
- Auto imaging optimization
- Easy Comparative Function to compare previous exam
- M Mode imaging
- Color Doppler imaging
- Power Doppler imaging
- Pulse wave Doppler imaging
- Duplex 2D/PW Doppler **3.7**
- Triplex 2D/Color/PW Doppler **3.8**
- High PRF pulsed wave Doppler
- Continuous wave Doppler
- Zoom
- HPRF
- FULL screen imaging to enlarge imaging size
- Dual real time imaging without compromising imaging size
- PView for panoramic imaging (optional)* Trapecinis vaizdavimas
- TView for trapezoidal imaging **4.1.3**
- Color M-mode(optional)*
- Free 3D(optional)*
- 3D/4D imaging
- HQ (optional)*
- HQ Silhouette (Optional)*
- Spatio Temporal Image Correlation (STIC) (optional)*
- Inversion mode(optional)* Skaitmeninių kanalų skaičius
- Magic cut(optional)* Maksimalus skenavimo gylis 40 m, priklausomai nuo daviklio

- Auto NT(Nuchal translucency)(optional)*
- VLuminous flow, a feature which shows the blood flow in a 3-D view with excellent sensitivity
- Curved M mode, user can draw any curved sample line freely and get corresponding results
- Sync B/C width, the width of B mode interest area is always be the same with the CF mode
- Multi-line Angular M-Mode, Up to 4 sample lines(optional)*
- Live IMT, display intima-media thickness in real time(Optional)*
- VAim(Vinno Artificial Intelligent Measurement) for OB、Follicle、Hip、pelvic(Optional)*
- 3D Smart Face, an intelligent tool for fetal face optimization(Optional)*
- Auto IT, automatic measurement of Intracranial translucency(Optional)*
- Cardiac Guide, an auxiliary scanning function for junior sonographers , which can automatically detect and capture the standard four-chamber view and starts the CF and PW mode during the scanning(Optional)*

1.4 Standard features

- Up to 25Mhz high frequency in system platform. Up to 23MHz' s
- Transmitting channels: 514 000 **2.5**
- Depth of scanning: max. 40 cm **2.6**

probes are supported

- RF platform and RF data processing
- Up to 1000 seconds

cine storage

5.1 • 250GB SSD quick boot up and storage

250 GB kietojo disko talpa

- Patient information database
- Image archive on hard drive
- Quick store to USB memory stick
- Quick store to hard drive
- Report package
- Quick print to B/W and color thermal video printer
- Network storage and printing
- Full measurement and analysis package
- Real time auto wave Doppler track and calculations
- Vascular calculations
- Cardiac calculations
- OB calculations and tables
- Gynecological calculations
- Urological calculations
- Renal calculations
- Volume calculations
- Wireless networking for easy data sharing, storage and printing*(optional)
- Up-to-date connectivity and data management solutions, wireless, LAN, integrated database*(optional)
- Capability to send data to mobile by mail and blue tooth*(optional)
- Total mobile medical solution for remote data transfer and diagnostic
- DICOM compatibility*(optional)

• 2USB ports 5.2

- 6 TGC slides
- Average 4 multiple adjustable frequency in every probe and mode
- Up to 512 line density

1.5 Language support

- Software: Chinese, English, German, Greek, Malay,Portuguese, Romanian, Spanish, Swedish, Norwegian, Danish, Finnish, French, Polish, Russian, Uighur, Italian, Czech, Hungarian, Cambodia
- Keyboard input: Chinese, English, German, Greek, Malay,Portuguese, Romanian, Spanish, Swedish, Polish, Norwegian, Danish, Finnish, French, Russian, Italian, Czech, Cambodia,Polski
- Control panel overlay: English
- User manual: Chinese, English, German, Russian, Portuguese, Spanish, Italian, French

2. Ergonomics

- Unique human oriented design for comfort and convenience
- 15.6-inch high resolution flat panel display with up to 150 degree positioning adjustments
- Easy to carry by integrated handle 2.1.1
- Full integrated probe to reduce overall space
- Integrated touchable alphabetic

Integruota rankena



keyboard

- Integrated capacity touch panel to easy and simplify workflow
- Cart support up to 200 mm up/down operation panel*(optional)

• Operating time from the built-in battery not less than 60 min. (depending on operating mode) **2.1.2**

- USB DVDRW *(optional)

Vidinės baterijos veikimo laikas ne mažiau 60 min. (priklausomai nuo darbo režimo)

2.1 Keyboard

- Highly sensitive 8 inch capacity touch panel
- Intuitive, configurable and touchable interactive operation interface
- Ergonomic hard keys for general ultrasound operations
- One active transducer port
- 6 TGC slides, functionality at any depth
- Backlight keys
- Resolution: 1280*800 pixels

2.2 Image display screen

15,6 colių ekranas

- 15.6 inch high resolution IPS, **2.3.1**

LED ekranas

LED technology, pixel resolution **2.3**

- Big angle tilting capability
- Backlight adjustment

Skiriamoji geba

Resolution: 1920*1080 pixels **2.3.2**

2.3 Touch gestures

- Swipe down/up: display/remove projected image on touch screen
- Swipe horizontally:

page up/down or review images/cine loops one by one

- Swipe from left edge to right: display hidden menu on projected image.

- Image parameter adjustment
- Measurement on projected image

on touch screen

- Zoom in/out the projected image

on touch screen

- Rotate or erase on projected 3D/4D image on touch screen

2.4 Comments

- Supports text input and arrow
- Support freehand marking on touch screen
- Adjustable text size and arrow size
- Supports home position
- Covers various application
- User customizable

2.5 Bodymark

- More than 215 bodymarks for versatile application
- User customizable

2.6 Peripherals

- B&W thermal video printer: Sony



UP-D898MD(optional)

- Color thermal video printer: Sony

UP-D25MD (optional)

- Memory stick (optional)

700hPa-1060hPa

2.7 Dimensions and Weight

- Length: 387mm
- Width: 340mm
- Depth: 47mm
- Weight: without accessories approx.

3.8kg **2.1.1**

Svoris be priedų 3,8 kg

2.8 Electrical Power

- Voltage: 100-240V AC
- Frequency: 50/60Hz
- Power: Max.170VA

2.9 Operating Environment

- Ambient temperature:
10-40° C
- Relative humidity: 30-75%
- Atmospheric pressure:
700hPa-1060hPa

2.10 Storage & Transportation

Environment

- Ambient temperature: -5-50° C
- Relative humidity: 10%-80%
(no condensation)
- Atmospheric pressure:

3. Transducers

3.1 Transducer Technology

- Xcen technology for wideband frequency
- Pure wave technology for high resolution imaging
- Unique and high technical Xcen probe connector to adapt all different type of VINNO product models

3.2 Transducer types

- Convex array
- Linear array
- Phase array
- 4D probe
- Endocavity probe
- Micro-convex array

3.3 Transducer selection

- Electronic switching of transducers
- User customizable imaging presets for each transducer and application
- Automatic dynamic receiving focus in all transducers
- Multiple adjustable transmit

focal zone, up to 8 focal zoom

G2-5C Broadband Curved Array

- Field of view: 66 degree
- Convex radius: 50mm
- Application: abdomen, OB/Gyn, urology, pediatric
- B-mode Frequency range: 2.0 -6.0MHz
- Physical Footprint: 68.5mm x 27mm
- Center frequency: 3.2 MHz
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes
- Reusable biopsy guide available

D3-6C broadband curved array volume probe

- Field of view: 75 degree
- Convex radius: 40mm
- Application: abdomen, OB/Gyn, urology
- B-mode Frequency range: 3.0 – 6.0MHz
- Physical Footprint: 82mm × 53mm
- Center frequency: 4.0 MHz
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic, 3D/4D

grayscale and 3D color modes

- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

D3-6CE broadband curved array volume probe

- Field of view: 68 degree
- Convex radius: 40mm
- Application: abdomen, ob/gyn, urology
- Center frequency: 4.5 MHz
- Transducer elements:128
- B-mode Frequency range: 3-5.5MHz
- Physical footprint: 74mm × 51mm
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic, 3D/4D grayscale and 3D color modes

F2-5CE Broadband Curved Array

- Field of view: 75 degree 8.2.2
- Convex radius: 60mm 75 laipsniai
- Application: abdomen, OB/Gyn, urology, pediatric
- B-mode Frequency range: 1.0 -6.5MHz 8.2.1 Dažnio diapazonas 1,0 - 6,5 MHz
- Center frequency: 3.2 MHz
- Physical footprint: 72mm x 27mm Elementai 128
- Transducer element: 128 8.2.3
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes
- Reusable biopsy guide available



G4-9M broadband micro convex array

- Field of view: 138 degree
- Convex radius: 12mm
- Application: pediatric, abdomen, cardiac
- B-mode Frequency range:
5.0 – 11.0MHz
- Physical Footprint:
34mm × 29mm
- Center frequency: 7.0MHz
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

G4-9E broadband micro convex endocavity array

- Field of view: 138 degree
- Convex radius: 11.5mm
- Application: Ob/Gyn, urology
- B-mode Frequency range:
5.0 – 11.0MHz
- Center frequency: 6.9MHz
- Physical footprint:
24.0mm x 18.8mm
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

- Reusable biopsy guide available

F4-9E broadband micro convex endocavity array

- Field of view: 150degree
- Convex radius: 10mm
- Application: Ob/Gyn, urology
- B-mode Frequency range:
5.0 - 11.0MHz
- Center frequency: 3.2MHz
- Physical footprint:
21.0mm (lens) x 19.0mm(lens)
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes
- Reusable biopsy guide available

D4-9E broadband micro convex 4D endocavity array

- Field of view: 146 degree
- Convex radius: 10mm
- Application: Ob/Gyn, urology
- B-mode Frequency range:
5.0 - 11.0MHz
- Center frequency:
6.5MHz
- Physical footprint:
24.7mm x 23.7mm
- Transducer element: 148
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic , 3D/4D grayscale, 3D color



- Multi-imaging frequency setting in 2D, 3D/4D, Harmonic, color Doppler and Wave Doppler modes

X4-12L broadband linear array

- Fine pitch, high resolution
- Applications: vascular, small parts
- Aperture size: 38.4mm
- B-mode Frequency range: 6.0 -16.0MHz
- Physical Footprint: 50.5mm × 17mm
- Center frequency: 7.8MHz
- Transducer element: 192
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes
- Reusable biopsy guide available

F4-12L broadband linear array

- Fine pitch, high resolution
- Applications: vascular, small parts
- Aperture size: 38.4mm
- B-mode Frequency range: 3.0 -16.0MHz **8.3.1**
- Center frequency: 7.5MHz
- Physical footprint: 50mm × 18.5mm **Elementai 128**
- Transducer element: 128 **8.3.2**
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave

Doppler modes

- Reusable biopsy guide available

I7-18L broadband linear array (Hock stick)

- Fine pitch, high resolution
- Applications: vascular, small parts, Hock stick
- Aperture size: 25.6mm
- B-mode Frequency range: 7.3-18.0 Mhz
- Center frequency: 9.5MHz
- Physical footprint: 31.7mm x 15mm
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

X6-16L broadband linear array

- Fine pitch, high resolution
- Applications: vascular, small parts
- Aperture size: 38.4mm
- B-mode Frequency range: 7.3 -18.0MHz
- Physical Footprint: 50.5mm × 20mm
- Center frequency: 10.5MHz
- Transducer element: 192
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave

Dažnio diapazonas
3,0-16,0 MHz

Elementai 128

Doppler modes

- Reusable biopsy guide available

X10-23L broadband linear array

- Fine pitch, high resolution
- Applications: small parts
- Aperture size: 12.8mm
- B-mode Frequency range: 10.0 -23MHz
- Center frequency: 15.0MHz
- Physical footprint: 32mm × 27.1mm
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

I4-11T broadband linear array

- Fine pitch, high resolution
- Aperture size: 25.6mm
- B-mode Frequency range: 6.0-12.0Mhz
- Physical Footprint: 48mm × 15mm
- Center frequency: 7.5MHz
- Transducer element: 128
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

G1-4P phased array

- Applications: cardiac, abdomen

- Field of view 90 degree **8.1.2**

- Aperture size: 17.92mm

- B-mode Frequency range: **8.1.1**
1.0-5.0Mhz

- Physical Footprint:

Dažnio diapazonas 1,0-5,0 MHz

34.5mm × 28.5mm

- Center frequency: 2.5MHz

- Transducer element: 64

- Pulsed wave Doppler, continuous wave Doppler, color Doppler, power Doppler, harmonic

- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

- Reusable biopsy guide available

U5-15LE broadband linear array

- Fine pitch, high resolution

- Applications: small parts, specially for breast, vascular

- Aperture size: 51.2mm

- B-mode Frequency range: 6.0 -14.0Mhz

- Physical Footprint: 65.5mm × 25.5mm

- Center frequency: 7.3MHz

- Transducer element: 256

- Pulsed wave Doppler, color Doppler, power Doppler, harmonic

- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

X9-22L broadband linear array



- Fine pitch, high resolution
- Applications: msk, nerve, small parts
- Aperture size: 28.8mm
- B-mode Frequency range: 9.0-22.0MHz
- Physical Footprint: 49.5mm x 22mm
- Center frequency: 14.0MHz
- Transducer element: 192
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

G3-10PX phased array

- Application: pediatric cardiology, abdomen
- Aperture size: 15.36 mm
- Field of view: 90 degree
- B-mode Frequency range: 3.0-10.0Mhz
- Physical Footprint: 33mmx 33mm
- Center frequency: 4.7MHz
- Transducer element: 96
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

X3-10L Low-frequency linear array

- Applications: Musculoskeletal, Peripheral Vascular

- Aperture size: 44.16mm
- B mode Frequency: 4-10MHz
- Physical Footprint: 60.18mm × 25.2mm
- Center frequency: 5.5MHz
- Transducer elements:192
- Pulsed wave Doppler, color Doppler, power Doppler, harmonic
- Multi-imaging frequency setting in 2D, Harmonic, color Doppler and Wave Doppler modes

4. Advanced Imaging controls

4.1 VFusion

- Available on all transducers and for 2D, 3D/4D
- Operate in conjunction with VSpeckle, harmonic imaging

4.2 VSpeckle

- Available on all transducers and for 2D, 3D/4D
- Virtually eliminate speckle noise artifact and dynamically enhances tissue margins
- Selectable multiple levels of speckle noise reduction and smoothing
- Operates in conjunction with VFusion and harmonic imaging



4.3 VTissue

- Special imaging processing to adapt to the speed of the ultrasound variation in different tissue
- Improved conspicuity of lesions, such as stone and tendon

4.4 Tissue Doppler (TD)

- Present wall motion spectrum by using Doppler principle
- Provide wall motion direction and velocity information

4.5 Tissue Velocity Imaging (TVI)(optional)*

- Color codes the velocities in tissue
- Present tissue color imaging by using Doppler principle
- This color image is overlaid onto the 2D image
- Captures low flow but high amplitude signals associated with wall motion

4.6 3D/4D

- 3D/4D rotation
- Grayscale imaging controls
- Selectable rendering approaches
- Unique high quality rendering algorithm

- Selectable gray maps
- Multi slide cutting
- Cineloop 3D
- Review volume

4.6.1 HQ (optional)*

- Amazing high image quality
- Extreme realistic rendering, images
- Similar operation as normal rendering

4.6.2 Spatio-Temporal Image

Correlation (STIC) (optional)*

- Visualize the fetal heart or an artery
- One complete heart cycle represented
- Using 3D static acquisition

4.6.3 3D Smart Face(optional)*

An intelligent tool for fetal face optimization. This tool detects the fluid/tissue interface and smartly removes noise in front of the baby inside the ROI, to obtain an optimal baby face.

- Use Auto key on the keyboard to remove the obstacle in front of the baby
- Only works on 3D Render
- Can not use this feature together with MagicCut



4.7 Smart operation

4.7.1 Free View(optional)*

- Provide any plane view to visualize the internal tissue information
- Improve the contrast resolution to facilitate the detection of diffuse lesions in organs

4.7.2 Inversion mode(optional)*

- This render mode is used to display anechoic structures such as vessels
 - It invert the gray values of the rendered image, such as black image information become white and vice versa

4.7.3 Magic Cut(optional)*

- Ability to edit images, make possible to cut away structure obstructing the view in the ROI
- Several cutting methods available

4.8 Smart 3D

4.8.1 Volume

Measurement(optional)*

Trace the margin of the irregular circle in different slices of volume data in irregular shape

Automatically report the volume of the irregular object

4.8.2 Auto Follicle(2D/3D)(optional)*

- Just click on the area of follicle in B mode, the area of this follicle
 - will be reported automatically
- Report the area of different follicle in the volume data automatically

4.9 Tissue Velocity M

mode(TVM)(optional) *

- Color codes the velocities in tissue
- Present wall motion spectrum based on tissue moving
- This color image is overlaid onto the 2D image
- Captures low flow but high amplitude signals associated with wall motion

4.10 Multi-angle M mode (optional)*

- Sample on moving tissue from multi-angle
- Present wall motion spectrum based on tissue moving

4.11 Elastic Imaging(EI)

(optional)*

- use the probe to press tissue artificially
- color codes the elasticity in tissue



- This color image is overlaid onto the 2D image

4.12 Next generation RF-based image processing

- Available on all imaging transducers in 2D grayscale modes
- Virtually eliminates speckle noise artifact and dynamically enhance tissue edge
- Operates with other real-time processing algorithms

4.13 Stress Echo(optional)*

- Stress echo is a non-invasive, dynamic evaluation of myocardial structure and its function under an external stress(exercise or pharmacology)
- 12 Ready to use templates (max 8 stages * 6 views) Editable User definable template
- Re-arrange & Select default template
- 8 View names available
- 9 Stage names are available (can add user defined stage name)
- One Touch Shuffle (Stage / View)
- Touch & Compare any view of stage
- Systole only review

4.14 Strain Imaging(optional)*

- Auto-ROI (after selecting Mitral Valve

Plane)

- Adjust Segment-wise (Longitudinal strain)
- Adjust Segment-wise and Rotate whole ROI (Radial & Circumf. Strain)
- ECG to select heart cycle
- View based Bulls Eye view
- Result type (Peak Strain or Peak Time)
- Parameter type (L Strain, R Strain& C Strain)

5. Imaging modes

2D vaizavimas / B režimas

5.1 2D Imaging 3.1

- Pre-defined ATGC (adaptive temporal gain compensation) curves optimized for consistently excellent imaging
- B/M acoustic output: 10%-100%
- Select between 1 to 8 transmit focal zones
- Reverse function: on/off
- 2D optimization: on/off
- Centerline: on/off
- L/R flip and U/D flip: on/off
- Display format: Single, Dual, Quad
- TGC: 6 slides on control panel
- SGC: 8 ponds on touch panel(optional)
- VFusion : ≥ 7 steps
- VSpeckle: ≥ 13 steps
- Harmonic imaging both tissue harmonic and phase inversion



- Cineloop image review
- Selectable 2D line density
- Dual imaging with independent cineloop
- 256(8 bit) gray level
- Up to 8 focus zone adjustable
- Multiple color maps with chroma imaging
- FULL screen imaging to larger image size
- Multi frequency: probe dependent
- Gray filter: ≥ 7 steps
- Persistence: ≥ 8 steps
- Selectable image angles, probe dependent
- Gain: 0-100%
- Dynamic range: 30-280 db
- VSharpen to enhance edge contrast: ≥ 8 steps
- Smooth to improve spatial resolution: ≥ 11 steps
- VNear to enhance SNR of near field, 4 steps
- Gray Map: ≥ 32 types
- Tint Map: ≥ 24 types
- TI heat index: TIB, TIS, TIC
- Rotation: $0^\circ, 90^\circ, 180^\circ, 270^\circ$
- Zoom (up to $10\times$)

5.2 Harmonic Imaging

- Supports both tissue harmonic and phase inversion imaging (transducer and frequency dependence)

- Second harmonic processing to reduce artifacts and improve image clarity
- Maximize detail resolution and enhance contrast
- Available on all imaging transducers
- Extends high performance imaging capabilities to all patient body types

5.3 M mode 3.2

M režimas

- Selectable sweeping rates, ≥ 10 steps
- Time marks: 0.025 - 0.5s
- Selectable display format prospective or retrospective (V1/3, V1/2, V2/3, H1/2, H3/4, full screen)
- Chroma colorization with multiple color maps
- Cineloop review for retrospective analysis of M-mode data
- 256 gray levels
- Acoustic output: 10%-100%
- Gray filter: ≥ 7 steps
- Dynamic range: 108db-128db, 2db/step
- Vsharpen: ≥ 6 steps
- Gray Map: ≥ 32 types
- Tint Map: ≥ 24 types
- Gain: 0-100%
- Color M mode: available
- MultiAngle: available

Spalvinis dopleris

5.4 Color Doppler mode 3.3

- Available on all imaging transducers



- Automatically adapts transmit and receive bandwidth processing based on the color box position
- Cineloop review with full playback control
- Steering on linear array transducers
- Selectable in baseline, line density, flash reduction, persistence, maps, frequency, PRF, wall filter, packet size, color level, sensitivity, focus position, acoustic power, and smooth
- FULL screen imaging to larger image size
- L/R flip and U/D flip: on/off
- Frequency: ≥ 4 steps, depend on probes
- Baseline: 0-100%
- Acoustic power: 0-100%
- Line density: ≥ 7 steps
- Flash reduction: ≥ 6 steps
- Persistence: ≥ 20 steps
- Color Map: ≥ 33 types
- Smooth : ≥ 7 steps
- Sensitivity: ≥ 6 steps
- Transparency: ≥ 6 steps
- Color level: ≥ 16 steps
- Packet size: ≥ 9 steps
- Reverse function: on/off
- Color gain: 0-100%
- Region of interest
- Baseline invert
- Simultaneous mode during PW mode
- Zoom

5.5 Power Doppler mode

- High sensitive mode for small vessel visualization
- Available on all transducers
- Cineloop review
- Display format: Single, Dual, Quad
- Individual controls for gain
- Selectable line density, flash reduction, persistence, maps, frequency, PRF, wall filter, packet size, color level, sensitivity, focus position, acoustic power, and smooth
- Color maps: ≥ 24 types
- Color levels: ≥ 11 steps
- Sensitivity: ≥ 6 steps
- Smooth: ≥ 7 steps
- Persistence: ≥ 20 steps
- Adjustable region of interest

PW pulsusis dopleris

5.6 Pulsed Wave (PW) Doppler 3.6

- Ultra high resolution spectral FFT rate
- Angle correction with automatic velocity scale adjustment
- Normal, invert display around horizontal zero line
- Selectable gray filter, dynamic range, frequency, PRF, wall filter, baseline, angel correct, sample volume
- Selectable sweep speeds: ≥ 10 steps



- PW acoustic output: 5%-100%
- Gray filter: \geq 6steps
- Dynamic range: 108db-128db
- Baseline: 5%-95%
- Sample volume: 0.5mm-10mm
- Angle correct: $-80^{\circ} \sim 80^{\circ}$
- Sensitivity: \geq 21steps
- Audio Volume: 0~20
- Spectrum Optimize: \geq 28steps
- Gray map: \geq 13types
- Tint map: \geq 11types
- Trace direction: above, below, both
- Trace type: max, mean, both
- Cardiac cycle: 1-5
- Selectable low frequency signal filtering with adjustable wall filter settings
- Selectable grayscale curve for optimal display
- Selectable chroma colorization maps
- Selectable display format prospective or retrospective (V1/3, V1/2, V2/3, H1/2, H3/4, full screen)
- Auto function to optimize spectral Doppler displ
- Digitally enhanced stereo output
- 256 gray levels
- Post-processing in frozen mode includes map, baseline, invert and chroma
- Simultaneous or duplex mode of operation
- Simultaneous 2D, color Doppler, pulsed Doppler

- High PRF capability in all modes including duplex and triplex **4.2.1**

Nuolatinės bangos dopleris

5.7 Continuous Wave Doppler (CWD)(optinal)* 3.4

- Cardiac sector array transducer only
- Maximum velocity range: 19m/sec

5.8 Pview(optional)*

- Real time extended field of view composite imaging
- Ability to back up and realign the image during acquisition
- Full zoom, cineloop review and image rotation capabilities
- User can measure distance and area
- Measurement can be made on individual frames during cineloop review
- Available on linear transducers

5.9 Tview

- Expand view of scanning
- Available on linear transducers

5.10 Auto

- Intelligent one button automatic optimization in 2D and Doppler modes
- Automatically adjust PRF and baseline



in Doppler

Audinių dopleris

3.10

5.11 Tissue Doppler Imaging (TD)

- Present wall motion spectrum by using Doppler principle
- Provide wall motion direction and velocity information
- Available on all sector

Transducer for cardiac imaging

- Selectable frequency, PRF, Focus position, wall filter, Gain
- Sweep speed: ≥ 10 steps
- Baseline: 5%-95%
- Angle correct: $-80^\circ \sim 80^\circ$
- Sample volume: 0.5mm-10mm, 0.5mm/step
- Spectrum optimize: ≥ 20 steps
- Acoustic power: 5%-100%
- Dynamic range: 108db-128db
- Trace sensitive: ≥ 21 steps
- Gray filter: ≥ 6 steps
- Audio volume: ≥ 21 steps
- Mode: max, mean, both
- Direction: above, below, both
- Heart cycle: 1-5
- Gray map: ≥ 13 types
- Tint map: ≥ 11 types

5.12 Tissue Velocity Imaging (TVI)

(optional)*

- Color codes the velocities in tissue

- Present tissue color imaging by using Doppler principle
- This color image is overlaid onto the 2D image
- Captures low flow but high amplitude signals associated with wall motion
- Available on all sector transducer for cardiac imaging
- Tissue velocity M mode display for wall motion(optional)
- Gain
- Velocity
- Color level: ≥ 11 steps
- Transparency: ≥ 13 steps
- Smooth: ≥ 7 steps
- Line density: ≥ 3 steps
- Persistence: ≥ 7 steps
- Color map: ≥ 10 types

5.13 Tissue Velocity M mode (TVM)

(optional)*

- Color codes the velocities in tissue
- Present wall motion spectrum based on tissue moving
- This color image is overlaid onto the 2D image
- Captures low flow but high amplitude signals associated with wall motion
- Selectable frequency, PRF, Focus position
- Baseline: 0%-100%
- Color level: ≥ 11 steps

- Transparency: ≥ 13 steps
- Packet size: 3,4,5,6
- Acoustic power: 5%-100%
- Display format: Single, Dual, Quad

5.14 Elastography

imaging(optional)*

- Shows the spatial distribution of tissue elasticity properties in a region of interest to estimate the strain before and after tissue distortion caused by external force
- The strain estimation is scaled by color to have smooth distribution display
- Have quality index to indicate if there is proper external force
- Winsize: 0,1,2,3,4
- Overlap: 0,1,2,3,4
- Dynamic range: 0-10
- Sensitivity: 0,1
- Transparency: ≥ 13 steps
- Smooth: ≥ 7 steps
- Line density: ≥ 7 steps
- Persistence: ≥ 20 steps
- Map: EIO
- Display format: Single, Dual, Quad

5.15 3D/4D

- 3D/4D rotation
- Grayscale imaging controls
- Selectable rendering approaches: HQ Surface, HQ Grad, HQ

Silhouette, Surf Texture, Surf Smooth, Grad Light, Surf HDR, Trans Max, X-ray, Transp Min, Light

- Unique high quality rendering algorithm
- Review volume
- Volume Angle: 15%-75%
- Quality: low, mid, good, high
- Threshold: 256
- Transparency: 0.1-2, 0.1/step
- Category: Face, Spine, Brain, Heart, Hi speed, Lip&plate, Limbs, Custom
- Display format: single, double, triplex, Quad
- Image Reference: A, B, C, 3D
- Flip: 0° , 90° , 180° , 270°
- View: Front/Back, Back/Front; Left/Right, Right/Left; Up/Down, Down/Up
- Rotation Direction: X, Y, Z
- 3D Map: ≥ 8 types
- Tint maps: ≥ 24 Types
- Gray maps: ≥ 32 Types
- 2D VSpeckle: ≥ 3 types
- 3D VSpeckle: ≥ 3 types
- Render Type: Gray, GrayInv
- Inverse Available
- **MCUT(optional)***
- Slice Number: 2×2 , 3×3 , 4×4 , 5×5
- Max Slice Number: 25
- Gray Map: ≥ 32 types
- Tint Map: ≥ 24 types
- Cut plane: A, B, C



- Rotation Direction: X, Y, Z
- Volume Angle: 15° ~ 75°
- Interval: 1mm-20mm, 0.5mm/step
- Quality: low,mid,good,high

- **Free view(optional)***

- Direction: left, right, up, down
- Route: curve, straight line
- Reference image: A,B,C
- Slice thickness: 0mm-20mm
- Active line: 1,2,3
- Mix: 10-90
- Threshold: 256steps
- Transparency: 0.1-2.0, 0.1/step

- **Magic cut(optional)***

- Erase mode: inside lasso, outside lasso, big circle, small circle
- Erase type: trace, rectangle, ellipse
- Rotation direction: X, Y, Z

- **VOCAL(optional)***

- Vocal layers: 8, 12,16,20,24,28,32
- Display format: single, Quad
- Image reference: A, B, C

- **Niche view(optional)***

- Model type: upper, lower
- Display format: single, quad
- Rotation direction: X, Y, Z
- Image reference: A, B, C, N

- **STIC(optional)***

- View: Front/Back, Back/Front; Left/Right, Right/Left; Up/Down, Down/Up;

- Image reference: A, B, C, D
- Flip: 0° 90° 180° 270°

6. Touch Panel Interface

6.1 2D mode

- New patient
- BodyPattern
- Archive
- Probe&APP
- Comments
- End exam
- Sys setting
- Report
- PView
- Fullscreen
- L/R
- U/D
- Center line
- VTissue
- VSpeckle
- VFusion
- Gray Filter
- Persistence
- Display Format
- Image reference
- Maps
- Frequency
- Focus position
- Focus
- Dynamic Range
- Line density
- VSharpen
- Biopsy



- Image angle
- Focus width
- Smooth
- Acoustic power

6.2 M Mode

- New patient
- BodyPattern
- Archive
- Probe&APP
- Comments
- End exam
- Sys setting
- Report
- L/R format
- U/D format
- Maps
- Dynamic range
- Acoustic power
- Sweep speed
- Gray filter
- VSharpen
- ECG

6.3 CF mode

- New patient
- BodyPattern
- Archive
- Probe&APP
- Comments
- End exam
- Sys setting
- Report

- Invert
- Full Screen
- L/R
- U/D
- Baseline
- Flash Reduction
- Line density
- Persistence
- Display format
- Image reference
- Maps
- Frequency
- PRF
- Wall filter
- Packet size
- Colorlevel
- Sensitivity
- Focus position
- Acoustic power
- Smooth

CW režimas

6.4 PW/CW mode 3.11

- New patient
- BodyPattern
- Archive
- Probe&APP
- Comments
- End exam
- Sys setting
- Report
- Invert
- Triplex
- Display format
- Sweep speed



- Gray filter
- Dynamic range
- Trace sensitive
- Auto trace
- Mode/direction
- Maps
- Frequency
- PRF
- Wall filter
- Baseline
- Angle correct
- Sample volume
- Audio Volume
- Spectrum optimize
- Acoustic power

6.5 3D mode

- Comments
- BodyPattern
- Back to B
- Start3D
- ROI shape
- Render
- Display format
- Image reference
- View
- Gray map
- VSpeckle
- Quality
- Threshold
- Transparency
- Volume angle
- Movement step (after data acquisition)

- HQ Light(after data acquisition)
- Rotation angle (after data acquisition)
- Rotation direction (after data acquisition)

6.6 4D mode

- Comments
- Body Pattern
- Back to B
- Start 4D
- ROI shape
- Movement step
- Rotation direction
- Render
- Display format
- Image reference
- View
- Gray map
- VSpeckle
- Quality
- Threshold
- Transparency
- Volume angle

7. System Feature

7.1 Display modes

- Simultaneous capability
 - 2D/PW/CW
 - 2D/CF or PDI **4.3.1**
 - 2D/M
 - Dual, 2D/2D



- Dual, 2D/2D+CF or PDI
- Dual, duplex and triplex
- Duplex and Triplex mode
- Quad display in 3D/4D

application

- 25 slice images display in 3D/4D

application

- Time line display
- Independent dual 2D/PW

or CW

- Timed based sweep update mode

7.2 Display annotation

- Institution/hospital name
- Date: 3 types selectable, Year-Month-Day, Day-Month-Year, Month-Day-Year
- Time: 2 types selectable, 24hours and 12 hours
- Operator identification
- Patient name, first, last
- Patient identification: 30 characters
- Gestational age LMP/BBT/DOC/IVF/GA/Avg.US
- VINNO image symbol: Ginkgo leaf
- Power output index
- MI: mechanical index
- TIS: thermal index soft tissue
- TIC: thermal index cranial (Bone)
- TIB: thermal index bone
- Probe orientation marker:

coincide with a probe orientation marking on the probe

- Gray/color bar
- Measurement result window
- Probe type
- Application name
- Image depth
- Imaging parameters by mode
- 2D/M mode:

acoustic power output, gain, frequency, frame rate, dynamic range

- Color mode: color acoustic power output, color gain, color flow frequency, PRF, wall filter

• PW/CW mode: Doppler acoustic power output, Doppler gain, Doppler frequency, PRF, wall filter, sample depth

- Focus zone marker
- Body pattern
- PW and CW scale markers: time/speed
- M scale markers: time/depth, time
- System measurement display
- System message display
- Biopsy guide line
- Heart rate

7.3 Cineloop

- Acquisition, storage in memory and display of up to 1000 seconds long of 2D, color and PW/CW images for review



7.4 Compare

- Compare live imaging with stored imaging

Paciento duomenų perdavimas į UB kaupiklius su DICOM

Quick save feature

5.3

- The system provides quick save function through USB stick, internal/external HDD during or after exam
- Configurable saving file format, VRD (VINNO Raw Data), DICOM, JPEG, BMP, PNG, MP4 and AVI

7.5 Physio

- One 3-lead ECG input*(optional)
- Gain, sweep rate and display position controls
- Automatic heart rate calculation and display
- Fault condition display

7.6 Archive

- Patient data input which include patient ID, name, birth date, sex, exam physician, quality check, exam operator
- Physical data such as weight, height
- Patient exam management
- Patient exam images storage and management
- Import VRD format data into the system from outside media, such as USB stick, external HDD

- Export patient data into outside medias

7.7 Report

- Automatically pull patient data into the report
- Automatically load measurement worksheet into the report
- Pull related exams' images into the report
- Write comments in the report
- Print report through network or local printer

7.8 Connectivity

- Standard connectivity features
- Local print to on-board or off-board video printers through USB port
- Page report print
- Image export to removable media (external HDD, USB stick)
- Network linkage
- Image export to network

Storage servers

- DICOM export and retrieve *(optional)
- Mobile data transfer solution by
 - Blue tooth*(optional)
 - Email*(optional)
 - Hot point connection
- VCloud * (optional)
- DICOM workstation for remote

diagnostic solution *(optional)

- DICOM, JPEG, BMP, PNG, AVI
 - VRD and DICOM images stored in disc can be recalled on the VINNO system
 - JPEG, BMP, PNG and AVI images can be played on normal computers
- On-board patient exam storage
 - Direct digital storage of static image or cineloop images to internal hard disk drives
- Fully integrated user interface
- Image transfer to an external device with HDMI connection 7. (cable 2 m or more)

Vaizdų perdavimas išoriniams įrenginiams per HDMI jungtį (galimas 2 m arba ilgesnis kabelis)

7.9 Probes/application

- Selectable multiple applications
- Edit exist application preset
- Edit user defined preset
- Rename preset
- Return to factory preset
- Quick save user defined parameters in related application

7.10 Safety Conformance

- Regulatory Notice:
This device is tested to meet all applicable requirements in relevant. According to 93/42 EEC, it is class IIa medical device.
- Conformity to Standards:
 - IEC 60601-1 : 2012 Medical electrical equipment - Part 1: General

requirements for basic safety and essential performance

- IEC 60601-1-2:2007

Electromagnetic compatibility - Requirements and tests

- IEC 60601-1-6:2010 Usability
- IEC 60601-2-37:2007

Medical electrical equipment - Particular requirements for the safety of ultrasonic medical diagnostic and monitoring equipment

- IEC 61157:2007

Declaration of acoustic output parameters

- ISO 10993-1:2009 Biological evaluation of medical devices

• IEC 62304:2006 Medical device software – Software life cycle processes

- IEC 62366:2007 Medical

devices - Application of usability engineering to medical devices

- Council Directive 93/42/EEC on

Medical Device

- WEEE according to 2012/19/EU
- RoHS according to 2011/65/EU

8. Measurement and Analysis

8.1 Measurement in different modes



8.1.1 Generic Measurement in 2D

mode

- Depth
- Distance
- Perimeter
 - Length and width method
 - Ellipse method
 - Polygon method
 - Spline method
 - Tracing method
- Area
 - Length and width method
 - Ellipse method
 - Polygon method
 - Spline method
 - Tracing method
- Volume
 - Single line method
 - Dual line method
- Triple line method
 - Single ellipse method
 - Single ellipse and single line

method

- Angle
- Stenosis
 - Diameter method
 - Square meter method
- A and B ratio
 - Diameter ratio

Generic Measurement in CFM

mode

- CFV
 - point
 - profile

8.1.2 Generic Measurement in M

mode

- Depth
- Distance
- Time
- Slope
- Heart rate
- Stenosis
- A and B ratio
 - Diameter ratio
 - Time ratio
 - Velocity ratio

8.1.3 Generic Measurement in PW

mode

- Speed (include PV (Peak Velocity))
- Time (include AT (Accelerate Time))
- Acceleration
- PS (Peak Speed in systole period)
- ED (The speed in the end of diastole period)
- MD (Minimum speed in diastole period)
- TAMAX (maximum speed in time average)
- TAMEAN (mean speed in time average)
- TAMIN (minmum speed in time average)
- PI (Pulsatility Index)



- RI (Resistance Index)
- PS and ED ratio
- ED and PS ratio
- A and B ratio (A/B ratio)
 - Speed ratio
 - Time ratio
 - Acceleration ratio
- Square meter ratio
- FLOWVOL (Flow Volume)
- MaxPG (maximum pressure gradient)
- MeanPG (Mean pressure gradient)
- SV (Stroke Volume)
- Each volume diameter cardiac
- Time mean speed in each stroke volume
- Cardiac output
- Heart rate
- $SV(LVOT)/SV(RVOT)$

8.2 Measurement in different applications

8.2.1 Abdominal Measurement

- General abdomen
- Difficult abdomen
- Kidney
- Renal vessel
- Abdominal trauma

8.2.2 Small Part Measurement

- Thyroid
- Breast
- Testis

- Musculoskeletal
- Upper and lower extremity joint
- Nerve block

8.2.3 Vessel Measurement

- Carotid artery
- Upper artery
- Upper vein
- Lower artery
- Lower vein
- Vessel puncture
- Transcranial Doppler

8.2.4 Gynecology Measurement

- Uterus and Pelvis
- Follicle

8.2.5 Urology Measurement

- Bladder
- Prostate
- Renal
- Kidney and ureter
- Pelvic Floor dysfunction

8.2.6 Pediatric Measurement

- Neonatal Head
- Neonatal Abdomen
- Pediatric Abdomen
- Pediatric Hip

8.2.7 Obstetrics Measurement

- OB Early

- OB Mid
- OB Late
- Fetal Heart

8.2.8 Cardiac Measurement

- General
- LV
- MV
- Ao
- AV
- LA
- RV
- TV
- PV
- RA
- System

8.2.9 Auto NT (Nuchal Translucency measurement)(Optional)*

- Automatically detect Nuchal Translucency in interest box
 - Automatically report thickness result of NT
- Auto IMT (Intima-Media Thickness) measurement
- Automatically detect intima media thickness in interest box
 - Automatically report the result of IMT
 - Available in linear probe

8.2.10 Smart 3D Volume

Measurement(Optional)*

- Trace the margin of the irregular circle in different slices of volume data in irregular shape
- Automatically report the volume of the irregular object

8.2.11 Live IMT (Intima-Media

Thickness) measurement (Optional) *

- Real-time automatically display IMT items with the different ROI positions
- The IMT items include: max, min, average, SD, points (how many points are used for the result), size of ROI .
- Available in carotid application

8.2.12 Auto IT (Intracranial translucency) measurement(Optional)

*

- Support Auto IT(Intracranial translucency) measurement
- Draw the ROI and the system analyses and displays the result

8.2.13 Auto Follicle(2D/3D)(Optional)

*

- Just click on the area of follicle in B mode, the area of this follicle will be reported automatically
- Report the area of different follicle in the volume data automatically



8.2.14 Smart 3D Volume

Measurement(Optional) *

- Trace the margin of the irregular circle in different slices of volume data in irregular shape
- Automatically report the volume of the irregular object

8.2.15 VAim OB measurement

(Optional) *

- VAim OB is an intelligent tool for fetal growth calculation, just one touch to activate the measurement items (BPD, OFD, HC, AC, FL, HL) and get the results, helps to make clinical decisions quickly and confidently, improving the speed and ease of exams
- The intelligent results will be added into the worksheet and report automatically

8.2.16 VAim Hip measurement

(Optional) *

VAim Hip is an intelligent solution in the assessment of DDH (Developmental Dysplasia of Hip) with one simple touch.

- Based on 'Ped HIP' application

8.2.17 VAim Follicle (2D)

measurement (Optional) *

An intelligent tool for follicle calculation, one touch to get the follicle status, dedicated for women's reproductive health.

- Choose left or right follicle
- Automatically identify all the follicles with different colors and calculate follicle volume and diameter



VINNO Technology (Suzhou) Co., Ltd

VINNO is focusing on producing premium diagnostic ultrasound development to provide customer clinical value through Continuous Innovation, Excellent Performance and Accessible Solutions.



Thanks you for your interest in VINNO.

You can contact us by phone and email (below) or contact our local representatives.

📍 5F, A Building, NO.27 Xinfra Rd, Suzhou Industrial Park, 215123, China(215.6123)

☎ Tel: +86 512 62873806

📠 Fax: +86 512 62873801

✉ email address: vinno@vinno.com

🌐 website: www.vinno.com