



8403XSI

C-MAC® S IMAGER, with C-MAC® system interface, for manual and machine disinfection up to 65 °C and High Level Disinfection (HLD) acc. to US standards, for use with C-MAC® Connecting Cable 8403X, C-MAC® Monitor 8403ZXX/8404ZXX, C-MAC® PM 8403XD and single-use C-MAC® S Video Laryngoscopes 051113-10, 051114-10, 051116-10



051114-10*

C-MAC® S Video Laryngoscope MAC #4, with MACINTOSH laryngoscope blade, size 4, for single use, package of 10, for use with C-MAC® S IMAGER 8402XS, 8403XS or 8403XSI



051113-10*

C-MAC® S Video Laryngoscope MAC #3, with MACINTOSH laryngoscope blade, size 3, for single use, package of 10, for use with C-MAC® S IMAGER 8402XS, 8403XS or 8403XSI



3. Video laringoskopo mentelė:
-vienkartinė (pažymėta simboliu);
-skirta apsunkintai intubacijai;
-pakuotėje 10 vnt.;
-skirta naudojimui su Karl Storz C-Mac monitoriais ir C-MAC® S Imager 8403 XSI;



■ 051116-10*

C-MAC® S Video Laryngoscope D-BLADE, laryngoscope blade for difficult intubation, for single use, package of 10, for use with C-MAC® S IMAGER 8402XS, 8403XS or 8403XSI



8401DS

C-MAC® GUIDE, guide rod made of stainless steel with atraumatic tip, distal region is adapted to the blade shape of the C-MAC® video laryngoscope D-BLADE, fixation of endotracheal tube with the integrated tube holder possible, package of 10, for use with C-MAC® video laryngoscopes



-su numatyta pakuotės atidarymo vieta.

 **KARL STORZ SE & Co. KG**
Dr.-Karl-Storz-Straße 34
78532 Tuttlingen/Germany

STORZ
KARL STORZ - ENDOSKOPE

REF 051116-01 **LOT** 240123
QTY 1

en C-MAC S Video Laryngoscope D-BLADE

D-BLADE

CE

mtip Single-Use Products

    

Rx only **MD**  2024-01-23

  www.karlstorz.com/ifu

LV020002

UDI (01,04048438004531(11)240123(10)240123(422)756

-ant pakuotės pažymėta produkto pagaminimo data;
PASTABA: laringoskopo mentelės yra tiekiamos nesterilios.