

## EU DECLARATION OF CONFORMITY

We, **Leica Biosystems Nussloch GmbH, Heidelberger Str. 17-19, 69226 Nussloch, Germany**

declare on our own responsibility, that the medical device

### **HistoCore MULTICUT HistoCore Arcadia C**

complies with the essential requirements of:

- Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in vitro diagnostic medical devices (OJ L 331, 7.12.1998, p. 1–37), Annex III

EN 61010-2-101:2017

EN ISO 14971:2012

EN 61326-2-6:2013

Device classification: IVD general/other


- Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (OJ L 174, 1.7.2011, p. 88–110)

EN 50581:2012

Quality Management System: Certified according to EN ISO 13485:2016 and ISO 9001:2015

Manufacturing site: Leica Microsystems Ltd. Shanghai, Building 1, 258 JinZang Road, Jinqiao Export&Processing Zone, Pudong, Shanghai 201206, PRC

Nussloch, 27.03.2018

  
\_\_\_\_\_  
Martin Laszak  
Director Customer Service Histology

  
\_\_\_\_\_  
Robert Gropp  
RA/QA Manager