

Quote Q-3300

| Approved



7.

Scan-RAM Sources (only one option required)		<i>Unit Price(€)</i>	<i>Qty</i>	<i>Amount(€)</i>
P1F-FXX-54	Calibration Source of Cs-137 x1 1x Calibration Source; - Activity 0.262 μ Ci (9,694 Bq) - Source type 0.5" laminated disk. - Delivery charges not included - Supplied with calibration certificate This source has been calibrated using a HPGe detector		1	
	whose efficiency was established with the NIST traceable mixed nuclide standard SRS: 80899-854. The stated activity is the weighted average of the measured gamma lines with an estimated uncertainty of 5%. The source is of sealed construction and wipe tested for surface contamination prior to shipping.			