

**Tina-quant Apolipoprotein A-1 ver.2****Order information**

REF	CONTENT	Analyzer(s) on which <b>cobas c</b> pack(s) can be used
03032566 122	Tina-quant Apolipoprotein A-1 ver.2 100 tests	System-ID 07 6568 6 Roche/Hitachi <b>cobas c</b> 311, <b>cobas c</b> 501/502
12172623 122	Calibrator f.a.s. Lipids (3 x 1 mL)	Code 424
12172623 160	Calibrator f.a.s. Lipids (3 x 1 mL, for USA)	Code 424
10781827 122	Precinorm L (4 x 3 mL)	Code 304
11285874 122	Precipath L (4 x 3 mL)	Code 305
05117003 190	PreciControl ClinChem Multi 1 (20 x 5 mL)	Code 391
05947626 190	PreciControl ClinChem Multi 1 (4 x 5 mL)	Code 391
05947626 160	PreciControl ClinChem Multi 1 (4 x 5 mL, for USA)	Code 391
05117216 190	PreciControl ClinChem Multi 2 (20 x 5 mL)	Code 392
05947774 190	PreciControl ClinChem Multi 2 (4 x 5 mL)	Code 392
05947774 160	PreciControl ClinChem Multi 2 (4 x 5 mL, for USA)	Code 392
04489357 190	Diluent NaCl 9 % (50 mL)	System-ID 07 6869 3

**English****System information**

For **cobas c** 311/501 analyzers:

**APOAT**: ACN 168

For **cobas c** 502 analyzer:

**APOAT**: ACN 8168

**Intended use**

In vitro test for the quantitative determination of Apolipoprotein A-1 in human serum and plasma on Roche/Hitachi **cobas c** systems.

**Summary**<sup>1,2</sup>

Apolipoproteins are the protein constituents of the lipoproteins. The lipoproteins are classified according to their ultracentrifugal flotation density. Apolipoprotein A-1 is the major protein constituent of high-density lipoproteins (HDL). HDL are synthesized by the intestines and the liver. They transport excess cellular cholesterol from extrahepatic tissue and peripheral cells to the liver. Additionally, apolipoprotein A-1 activates the enzyme lecithin-cholesterol-acyltransferase (LCAT), which catalyzes the esterification of cholesterol, thereby enhancing the lipid-carrying capacity of the lipoproteins.

Apolipoprotein A-1 levels increase in liver disease, pregnancy and as a result of estrogen administration (e.g. oral contraceptives).

Apolipoprotein A-1 levels decrease in inherited hypo- $\alpha$ -lipoproteinemia (e.g. Tangier disease), cholestasis, sepsis and atherosclerosis. The liver also synthesizes very low density lipoproteins (VLDL) which mainly contain triglycerides and cholesterol. In the presence of lipoprotein lipase the triglycerides are hydrolyzed and LDL-particles with a high proportion of cholesterol are formed. Apolipoprotein B is the main constituent of LDL.

The combined determination of apolipoprotein A-1/apolipoprotein B and the calculation of the apolipoprotein B : apolipoprotein A-1 ratio can reflect a lipid metabolism disorder and the risk of developing atherosclerosis or coronary heart disease particularly well, thus providing an excellent addition to the classical HDL/LDL-cholesterol determination. A high level of apolipoprotein A-1 (HDL) and a low level of apolipoprotein B (LDL) correlate best with a low risk for these diseases.

**Test principle**<sup>3,4,5,6</sup>

Immunoturbidimetric assay.

Anti-apolipoprotein A-1 antibodies react with the antigen in the sample to form antigen/antibody complexes which, following agglutination, can be measured turbidimetrically.

**Reagents - working solutions**

**R1** TRIS buffer: 50 mmol/L, pH 8.0; PEG: 3.8 %; detergent; preservative

**R2** Anti-human apolipoprotein A-1 antibodies (sheep): dependent on titer; TRIS buffer: 100 mmol/L, pH 8.0; preservative

R1 is in position B, and R2 is in position C.

**Precautions and warnings**

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

**Reagent handling**

Ready for use

**Storage and stability****APOAT**

Shelf life at 2-8 °C:

See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer:

12 weeks

**Diluent NaCl 9 %**

Shelf life at 2-8 °C:

See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer:

12 weeks

**Specimen collection and preparation**

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable. Serum.

Plasma: Li-heparin and K<sub>2</sub>-EDTA plasma

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Stability:

1 day at 15-25 °C<sup>7</sup>

8 days at 2-8 °C<sup>7</sup>

2 months at (-15)-(-25) °C<sup>8</sup> (only freeze once)

**Materials provided**

See "Reagents – working solutions" section for reagents.

**Materials required (but not provided)**

See "Order information" section

General laboratory equipment

**Assay**

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

**Application for serum and plasma****cobas c 311 test definition**

Assay type	2-Point End	
Reaction time / Assay points	10 / 6-19	
Wavelength (sub/main)	700/340 nm	
Reaction direction	Increase	
Units	g/L (µmol/L, mg/dL)	
Reagent pipetting	Diluent (H <sub>2</sub> O)	
R1	100 µL	–
R2	25 µL	30 µL

Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	3 µL	9 µL	180 µL
Decreased	3 µL	9 µL	180 µL
Increased	6 µL	9 µL	180 µL

**cobas c 501/502 test definition**

Assay type	2-Point End	
Reaction time / Assay points	10 / 10-28	
Wavelength (sub/main)	700/340 nm	
Reaction direction	Increase	
Units	g/L (µmol/L, mg/dL)	
Reagent pipetting	Diluent (H <sub>2</sub> O)	
R1	100 µL	–
R2	25 µL	30 µL

Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	3 µL	9 µL	180 µL
Decreased	3 µL	9 µL	180 µL
Increased	6 µL	9 µL	180 µL

**Calibration**

Calibrators	S1: H <sub>2</sub> O S2-S6: C.f.a.s. Lipids
	Multiply the lot-specific C.f.a.s. Lipids calibrator values by the factors below to determine the standard concentrations for the 6-point calibration curve:
	S2: 0.208                      S5: 1.313
	S3: 0.412                      S6: 2.006
	S4: 1.000
Calibration mode	RCM
Calibration frequency	Full calibration <ul style="list-style-type: none"> <li>• after reagent lot change</li> <li>• as required following quality control procedures</li> </ul>

Traceability: This method has been standardized against the IFCC SP1-01 reference standard (WHO-IRP October 1992).<sup>9,10,11,12</sup>

**Quality control**

For quality control, use control materials as listed in the "Order information" section.

In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

**Calculation**

Roche/Hitachi **cobas c** systems automatically calculate the analyte concentration of each sample.

Conversion factors:	g/L x 35.7 = µmol/L <sup>13</sup>
	g/L x 100 = mg/dL
	mg/dL x 0.01 = g/L

**Limitations - interference**

Criterion: Recovery within ± 10 % of initial values at apolipoprotein A-1 levels of 1.00 g/L (35.7 µmol/L, 100 mg/dL).

Icterus:<sup>14</sup> No significant interference up to an I index of 60 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 60 mg/dL or 1026 µmol/L).

Hemolysis:<sup>14</sup> No significant interference up to an H index of 1000 (approximate hemoglobin concentration: 621 µmol/L or 1000 mg/dL).

Lipemia (Intralipid):<sup>14</sup> No significant interference up to an L index of 1000. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Rheumatoid factors ≤ 1200 IU/mL do not interfere.

High dose hook-effect: No false result occurs up to an apolipoprotein A-1 concentration of 11 g/L (392 µmol/L, 1100 mg/dL).

Drugs: No interference was found at therapeutic concentrations using common drug panels.<sup>15,16</sup>

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.<sup>17</sup>

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

**ACTION REQUIRED**

**Special Wash Programming:** The use of special wash steps is mandatory when certain test combinations are run together on Roche/Hitachi **cobas c** systems. The latest version of the carry-over evasion list can be found with the NaOHD/SMS/Multiclean/SCCS or the NaOHD/SMS/SmpCln1+2/SCCS Method Sheets. For further instructions refer to the operator's manual. **cobas c 502** analyzer: All special wash programming necessary for avoiding carry-over is available via the **cobas** link, manual input is not required.

**Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.**

**Limits and ranges****Measuring range**

0.2-4.0 g/L (7.14-143 µmol/L, 20-400 mg/dL)

Determine samples having lower concentrations via the rerun function. For samples with lower concentrations, the rerun function increases the sample volume by a factor of 2. The results are automatically divided by this factor.

**Lower limits of measurement****Lower detection limit of the test**

0.03 g/L (1.07 µmol/L, 3 mg/dL)

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying 3 standard deviations above that of the lowest standard (standard 1 + 3 SD, repeatability, n = 21).

**Expected values<sup>18</sup>**

The following values were obtained using serum from healthy subjects:

Men 1.04-2.02 g/L (37.1-72.1 µmol/L, 104-202 mg/dL)

Women 1.08-2.25 g/L (38.6-80.3 µmol/L, 108-225 mg/dL)

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

**Specific performance data**

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

**Precision**

Precision was determined using human samples and controls in an internal protocol with repeatability (n = 21) and intermediate precision (3 aliquots per run, 1 run per day, 21 days). The following results were obtained:

	<i>Repeatability</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>
		<i>g/L (µmol/L, mg/dL)</i>	<i>g/L (µmol/L, mg/dL)</i>	<i>%</i>
Precinorm L		1.60 (57.1, 160)	0.02 (0.7, 2)	1.1
Precipath L		1.00 (35.7, 100)	0.01 (0.4, 1)	1.5
Human serum 1		0.99 (35.3, 99.0)	0.02 (0.7, 2)	1.5
Human serum 2		2.59 (92.5, 259)	0.02 (0.7, 2)	1.0
	<i>Intermediate precision</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>
		<i>g/L (µmol/L, mg/dL)</i>	<i>g/L (µmol/L, mg/dL)</i>	<i>%</i>
Precinorm L		1.74 (62.1, 174)	0.08 (2.9, 8)	4.7
Precipath L		1.06 (37.8, 106)	0.04 (1.4, 4)	3.9
Human serum 3		1.17 (41.8, 117)	0.04 (1.4, 4)	3.6
Human serum 4		2.40 (85.7, 240)	0.03 (1.1, 3)	1.4

**Method comparison**

Apolipoprotein A-1 values for human serum and plasma samples obtained on a Roche/Hitachi **cobas c** 501 analyzer (y) were compared with those determined using the same reagent on a Roche/Hitachi 917 analyzer (x).

Sample size (n) = 126

Passing/Bablok <sup>19</sup>	Linear regression
$y = 1.014x + 0.073$ g/L	$y = 1.006x + 0.087$ g/L
$r = 0.949$	$r = 0.996$

The sample concentrations were between 0.380 and 3.34 g/L (13.6 and 119 µmol/L, 38.0 and 334 mg/dL).

**References**

- Riesen WF. Apolipoproteine. In: Thomas L, ed. Labor und Diagnose, 5th ed. Frankfurt 1998;171-190.
- Brewer HB Jr, Gregg RE, Hoeg JM, et al. Apolipoproteins and lipoproteins in human plasma: an overview. Clin Chem. 1988;34:B4-B8.
- Becker W, Rapp W, Schwick HG, et al. Methoden zur quantitativen Bestimmung von Plasmaproteinen durch Immunpräzipitation. Z Klin Chem Klin Biochem. 1968;6:113-122.
- Siedel J, Schiefer S, Rosseneu M, et al. Immunoturbidimetric Method for Routine Determinations of Apolipoproteins A-I, A-II and B in Normo- and Hyperlipemic Sera compared with Immunonephelometry. Clin Chem. 1988;34(9):1821-1825.
- Rifai N, King ME. Immunoturbidimetric Assays of Apolipoproteins A, AI, AII and B in serum. Clin Chem. 1986;32:957-961.
- Naito HK. Reliability of Lipid, Lipoprotein and Apolipoprotein Measurements. Clin Chem. 1988;34:B84-B94.
- Use of Anticoagulants in Diagnostic Laboratory Investigations. WHO Publication WHO/DIL/LAB/99.1 Rev. 2. Jan. 2002.
- Evans K, Mitcheson J, Laker M. Effect of Storage at 4 °C and -20 °C on Lipid, Lipoprotein, and Apolipoprotein Concentrations. Clin Chem. 1995;41:392-396.
- Marcovina SM, Albers JJ, Dati F, et al. International Federation of Clinical Chemistry Standardization Project for Measurements of Apolipoproteins A-I and B. Clin Chem. 1991;37:1676-1682.
- Albers JJ, Marcovina SM, Kennedy H. International Federation of Clinical Chemistry Standardization Project for Measurements of Apolipoproteins A-I and B. II. Evaluation and Selection of Candidate Reference Materials. Clin Chem. 1992;38:658-662.
- Marcovina SM, Albers JJ, Henderson LO, et al. International Federation of Clinical Chemistry Standardization Project for Measurements of Apolipoproteins A-I and B. III. Comparability of Apolipoprotein A-I Values by Use of International Reference Material. Clin Chem. 1993;39:773-781.
- Marcovina SM, Albers JJ, Kennedy H, et al. International Federation of Clinical Chemistry Standardization Project for Measurements of Apolipoproteins A-I and B. IV. Comparability of Apolipoprotein B Values by Use of International Reference Material. Clin Chem. 1994;40:586-592.
- Young DS, et al. Effects of drugs on clinical laboratory tests. American College of Physicians 1998.
- Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. Clin Chem 1986;32:470-475.
- Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". Eur J Clin Chem Clin Biochem 1996;34:385-386.
- Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem 2001;38:376-385.
- Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med 2007;45(9):1240-1243.
- Data on file at Roche Diagnostics.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

**Symbols**

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT

Contents of kit



Volume after reconstitution or mixing

**FOR US CUSTOMERS ONLY: LIMITED WARRANTY**

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C, PRECICONTROL, PRECINORM, PRECIPATH and TINA-QUANT are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics

000303256612c501V9.0

# APOAT

Tina-quant Apolipoprotein A-1 ver.2

**cobas**<sup>®</sup>



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)



Distribution in USA by:  
Roche Diagnostics, Indianapolis, IN  
US Customer Technical Support 1-800-428-2336