

## MANUFACTURER DECLARATION

Date: 2022 10 19

Customer: LITHUANIAN UNIVERSITY OF HEALTH SCIENCES HOSPITAL  
KAUNAS CLINIC

Tender name: Disinfection machine for washing urinals and basins

Tender no.: 629608

Herewith below we, as manufacturer (PROHS-Equipamento Hospitalar e Serviços Associados, SA) declare and confirm our true parameters of our manufactured (Prohs PLUS):

1. Automatinė plovimo - dezinfekavimo mašina, pritaikyta bazonų ir šlapimo indų (su visu turiniu) išplovimui bei terminei dezinfekcijai, pagaminta iš nerūdijančio plieno

No.	Technical requirements	Parameter	Manufacturer declared exact parameter
1.	Purpose	Automatic washing-disinfecting machine, adapted for washing and thermal disinfection of basins and urine containers (with all contents), made of stainless steel (or equivalent material)	<b>Automatic washing-disinfecting machine, adapted for washing and thermal disinfection of basins and urine containers (with all contents), made of stainless steel</b>
7.	Complete disinfection of the machine pipeline after each washing - disinfection cycle	Complete disinfection of the machine pipeline is necessary after each washing-disinfection cycle	<b>Complete disinfection of the machine pipeline is necessary after each washing-disinfection cycle</b>
8.	Automatic emptying of basons when the machine door is closed	Automatic emptying of the basons after closing the machine door is necessary.	<b>Automatic emptying of the basons after closing the machine door</b>
10.	The shortest cycle time of bason washing-disinfecting programs	No more than 10 minutes	<b>10 minutes</b>

7. Pilna mašinos vamzdyno dezinfekcija po kiekvieno plovimo – dezinfekavimo ciklo.

8. Automatinis bazonų ištuštinimas užsidarius mašinos durims


10. 10 min

11.	Machine control	Microprocessor or equivalent	Microprocessor
-----	-----------------	------------------------------	----------------

11. Mikroprocesorinis

Maia, 20th October 2022

Yours Sincerely,



Milton RODRIGUES  
Chief Operations Officer