

Attestation

We, Devea, hereby declare that our device Phileas ONE with accessories fully meets the requirements of Santaros klinikos tender no 658723 for the decontamination of premises and biological safety cabinets.

Operating principle: The device operates on the principle of centrifugation and reaches a speed of 48,000 rpm.

Diffusion time:

- 30 minutes for a room of 50 m³ using 12 ml/m³.
- 60 minutes for a room of 100 m³ using 12 ml/m³.
- 90 minutes for a room of 150 m³ using 12 ml/m³.

Accessories included with the device:

- Trolley for transporting the device in premises.
- Hydrogen peroxide solution (with a validity of at least 24 months), sufficient for decontamination of premises and three biological safety cabinets for two years - at all 32 liters.
- Tablet with the manufacturer's software for remote device control.

The tablet included with the device comes with pre-installed manufacturer's software for the following actions:


- Remote monitoring
- Automatic programming system
- Password-protected control
- Access to all disinfection protocols.

St Etienne de Montluc, le 19/04/2023

Devea SAS,

Export Sales Manager

Xavier Blanc



devea-environnement.com



puminski@devea-environnement.com



+33 (0)2 40 57 07 40



PA du Bois de la Noue
44360 St-Étienne-de-Montluc

