



EXCELLENT
EXCELLENT MEDICAL PRODUCTS CO., LTD.

Adhesive aperture drape 100x125cm

REF. EA100125C

-  Apklotas 100x125cm su lipnia anga
-  Pārklājs 100x125cm ar pašlīpošu atveri
-  Kleepavaga lina 100x125cm



100x125cm



Size	100x125cm, adhesive aperture 9x11cm
Raw materials	2ply: PP (polypropylene), PE (polyethylene)
Sterile	EO
Traceability labels quantity	4
Package type	3 level [1pc/bag, 20pcs/box, 40pcs/ctn]
Applicable standards	MDD 93/42/EEC Class I, (EU) 2017/745, EN ISO 13485:2016, EN ISO 14971:2012, EN 10993, EN1041, EN ISO 14155-1:2011, EN ISO11135-1:2007, EN 15233-1:2012, EN ISO 11607, EN ISO11737-1:2006/AC:2009, EN ISO 13795:2011+A1:2013, EN ISO 22610:2006, EN ISO 11948
CE marking	CE0123
EAN code for inner bags	6972823182765
EAN code for boxes	6972823182772
EAN code for outer carton	6972823182789



UAB "Barameda"
 Perkūnkiemis g. 5, 12129, Vilnius, Lithuania
 Tel. +370 650 96966 e-mail. info@barameda.com
 www.barameda.com

EXCELLENT MEDICAL PRODUCTS CO., LIMITED
 Room2308,Decheng Jiayuan Plaza, No.1539 Xiangcheng Road,
 Xiangcheng District, 215131 Suzhou, Jiangsu,
 PEOPLE'S REPUBLIC OF CHINA

Shanghai International Holding Corp. GmbH (Europe)
 Eiltestrasse 80, 20537 Hamburg Germany



STERILE EO   **CE** 0123

   40°C **ISO13485:2016**