

## ***Letter of explanation***

***Date:12/9/2025***

***Concerning: Request N350320 /11.09.2025/ from***

*STATE PROCUREMENT COMMISSION (Purchase No. 3608733) KAUNAS UNIVERSITY CLINIC (KMUK)*

*Open tender 'DISINFECTANTS, STERILISATION AGENTS AND CONTAINERS FOR USED NEEDLES'*

*(Procurement No. 3608733)*

Dear Members of the State Procurement Commission,

We Zhivas Ltd, here under state that we are the manufacturer of the product OXISEPT – Medical device Class II B.

We would like to present to the Commission the following clarifications:

Concerning Point 3

We confirm that the product Oxisept has been tested as per standard EN 14562 Fungicidal activity (*Aspergillus brasiliensis* (niger)) – test report № 1/15,01,2025. The test is conducted by the National Center for infectious and parasitic diseases (Bulgaria). We are sending the test report and the certificate of the laboratory as reference.

Concerning Point 4

We confirm that the product Oxisept has been tested as per standard EN 17126 Sporicidal activity (*B. subtilis*, *B. cereus*, *C.difficile*) – test report S172/2023-1. The test report confirms the Sporicidal efficacy at concentration of 3% for 10 min.. The test is conducted by laboratory Chemila (Czech Republic). We are sending the test report and the certificate of the laboratory as reference.



Damian Kamburov  
(Managing director)