

1.01424.0100
 1.01424.0103
 1.01424.0500
 1.01424.0503
 1.01424.1000
 1.01424.1022
 1.01424.2500
 1.01424.9025
 1.01424.9026

Microscopy

May-Grünwald's eosine-methylene blue solution modified

for microscopy



In Vitro Diagnostic Medical Device



for staining of blood and bone marrow smears and clinical-cytological specimens

This "May-Grünwald's eosine-methylene blue solution modified - for microscopy" is used for human-medical cell diagnosis and serves the purpose of the investigation of sample material of human origin. It is a staining solution that when used together with other in vitro diagnostic products from our portfolio makes target structures (by fixing, embedding, staining, where necessary counterstaining, mounting) in hematological and clinico-cytological specimen materials, for example whole blood or bone-marrow smears, evaluable for diagnostic purposes.

Principle

May-Grünwald's stain is frequently used for diagnostic purposes in the areas of hematology and histology.

When used in hematological applications, May-Grünwald's stain is frequently used in combination with other dye solutions, e.g. with Giemsa's solution for Pappenheim (MGG) overview staining. This staining solution generally stains the nuclei red to violet, based on the molecular interaction between the eosin Y dye and an Azure B-DNA complex. Both dyes assemble to an Eosin Y - Azure B-DNA complex and the intensity of the resulting stain depends on the content of Azure B and the ratio of Azure B : Eosin Y.

Furthermore, the resulting stain can vary depending on the influence of fixation, staining times, pH-value of the solutions or buffer substances.

Sample material

air-dried blood or bone-marrow smears just as clinical cytological material like urine sediment, sputum, smears from fine needle aspiration biopsies (FNAB), rinses, imprints

Reagents

Cat. No. 101424
 May-Grünwald's eosine-methylene blue solution modified for microscopy 100 ml, 500 ml, 1 l, 2.5 l, 25 l

Also required:

for the staining of air-dried smears:

Cat. No. 109468 Buffer tablets pH 7.2 100 tabs
 for preparing buffer solution acc. to WEISE for staining of blood smears

or

Cat. No. 111373 Buffer tablets pH 6.4 100 tabs
 for preparing buffer solution acc. to WEISE for the staining of blood smears

or

Cat. No. 111374 Buffer tablets pH 6.8 100 tabs
 for preparing buffer solution acc. to WEISE for the staining of blood smears

for the Pappenheim staining:

Cat. No. 109204 Giemsa's azur eosin methylene blue solution for microscopy 100 ml, 500 ml, 1 l, 2.5 l

Sample preparation

The sampling must be performed by qualified personnel.

All samples must be treated using state-of-the-art technology. All samples must be clearly labeled.

Suitable instruments must be used for taking samples and their preparation. Follow the manufacturer's instructions for application / use.

Reagent preparation

May-Grünwald's eosine-methylene blue solution modified

The solution is supplied as a concentrated staining solution and before use must be diluted with a buffer solution as described below. The diluted staining solution should be filtered before use.

Buffer solution

For preparation of approx. 100 ml of solution, add and dissolve:

Buffer tablet, Cat. No. 111373 (pH 6.4), Cat. No. 111374 (pH 6.8), or Cat. No. 109468 (pH 7.2) depending on the required reaction color	1 tablet
Distilled water	1000 ml

Diluted May-Grünwald's staining solution for manual staining

For preparation of approx. 200 ml solution mix:

May-Grünwald's eosine-methylene blue solution modified	30 ml
Buffer solution	20 ml
Distilled water	150 ml
mix and filter	

Dilute May-Grünwald's staining solution for staining with staining automate

For preparation of approx. 300 ml solution mix **slowly**:

May-Grünwald's eosine-methylene blue solution modified	50 ml
Buffer solution	30 ml
Distilled water	220 ml
Mix, leave to stand for 10 min, and filter	

Dilute Giemsa's staining solution for manual staining

For preparation of approx. 200 ml solution mix:

Giemsa's azur eosin methylene blue solution	10 ml
Buffer solution	190 ml
Mix well, leave to stand for 10 min, and filter if necessary	

May-Grünwald's stain

Procedure

Air-dried smears

Staining in the staining cell

The slides should be allowed to drip off well after the individual staining steps, as a measure to avoid any unnecessary cross-contamination of solutions.

The stated times should be adhered to to guarantee an optimal staining result.

Slide with air-dried smear	
May-Grünwald's eosine-methylene blue solution modified	3 min
Dilute May-Grünwald's staining solution for manual staining	6 min
Buffer solution	1 min
Buffer solution	1 min
Air-dry (e.g. over night or at 50 °C in the drying cabinet)	

Staining on the staining rack

The stated times should be adhered to to guarantee an optimal staining result.

Slide with air-dried smear			
May-Grünwald's eosine-methylene blue solution modified	1 ml	cover completely	3 min
Buffer solution	1 ml	mix	6 min
Buffer solution		rinse	
Air-dry (e.g. over night or at 50 °C in the drying cabinet)			

Staining in the staining automate

The stated times should be adhered to to guarantee an optimal staining result.

	Time	Station	Dip
Slide with air-dried smear			
May-Grünwald's eosine-methylene blue solution modified	3 min	2	on
Dilute May-Grünwald's staining solution for staining with staining automate	6 min	3	on
Buffer solution	1 min	4	on
Running tap water	2 min	5	on
Dry	3 min	6	-

Covering with non-aqueous mounting media (e.g. Neo-Mount®, DPX new, or Entellan® new) and a cover glass is recommended for the storage of hematological specimens for several months. When left unmounted, the stain remains stable for approx. 3 days, covered with immersion oil for just a few hours.

After dehydration (ascending alcohol series) and clarification with xylene or Neo-Clear®, cytological samples can be mounted with water-free mounting agents (e.g. Entellan® new, DPX new, or Neo-Mount®) and a cover glass and can then be stored.

The use of immersion oil is recommended for the analysis of stained slides with a microscopic magnification >40x.

Result

	Buffer solution pH 6.4	Buffer solution pH 6.8	Buffer solution pH 7.2
Cell nuclei	red-violet	red-violet	violet
Lymphocytes	plasma blue	plasma blue	plasma blue
Monocytes	plasma grey-blue	plasma grey-blue	plasma grey-blue
neutrophilic granulocytes	granule light violet	granule light violet	granule light violet
eosinophilic granulocytes	granule brick-red to red-brown	granule brick-red	granule red-brown
basophilic granulocytes	granule dark violet to black	granule dark violet to black	granule dark violet to black
Thrombocytes	violet	violet	violet
Erythrocytes	reddish	reddish	pink to brownish

Pappenheim's stain

with May-Grünwald's solution and Giemsa's solution

Procedure

Air-dried smears

Staining in the staining cell

The slides must be immersed and moved briefly in the solutions, simple immersion alone yields inadequate staining results.

The slides should be allowed to drip off well after the individual staining steps, as a measure to avoid any unnecessary cross-contamination of solutions.

The stated times should be adhered to to guarantee an optimal staining result.

Slide with air-dried smear		
May-Grünwald's eosin-methylene blue solution	3 min	
Dilute Giemsa's staining solution for manual staining	20 min	
Buffer solution	1 min	
Buffer solution	1 min	
Air-dry (e.g. over night or at 50 °C in the drying cabinet)		

Staining on the staining rack

The stated times should be adhered to to guarantee an optimal staining result.

Slide with air-dried smear			
May-Grünwald's eosin-methylene blue solution	1 ml	cover completely	3 min
Buffer solution	1 ml	mix	
Dilute Giemsa's staining solution for manual staining		cover completely	20 min
Buffer solution		rinse	
Air-dry (e.g. over night or at 50 °C in the drying cabinet)			

Covering with non-aqueous mounting media (e.g. Neo-Mount®, Entellan®, DPX new, or Entellan® new) and a cover glass is recommended for the storage of hematological specimens for several months. When left unmounted, the stain remains stable for approx. 3 days, covered with immersion oil for just a few hours.

After dehydration (ascending alcohol series) and clarification with xylene or Neo-Clear®, cytological samples can be mounted with water-free mounting agents (e.g. Entellan® new, DPX new, or Neo-Mount®) and a cover glass and can then be stored.

The use of immersion oil is recommended for the analysis of stained slides with a microscopic magnification >40x.

Result

	Buffer solution pH 6.8	Buffer solution pH 7.2
Cell nuclei	purple to violet	violet
Lymphocytes	plasma blue	plasma blue
Monocytes	plasma grey-blue	plasma grey-blue
neutrophilic granulocytes	granule light violet	granule light violet
eosinophilic granulocytes	granule brick-red	granule red-brown
basophilic granulocytes	granule dark violet to black	granule dark violet to black
Thrombocytes	violet	violet
Erythrocytes	reddish	reddish-grey

Technical notes

The microscope used should meet the requirements of a medical diagnostic laboratory.

When using histoprocessors and automatic staining systems, please follow the instructions for use supplied by the supplier of the system and software.

The freshly prepared staining solutions should be filtered before use.

Remove surplus immersion oil before filing.

Diagnostics

Diagnoses are to be made only by authorized and trained personnel.

Valid nomenclatures must be used.

Further tests must be selected and implemented according to recognized methods.

Suitable controls should be conducted with each application in order to avoid an incorrect result.

Storage

Store the May-Grünwald's eosine-methylene blue solution modified - for microscopy at +15 °C to +25 °C.

Shelf-life

The May-Grünwald's eosine-methylene blue solution modified - for microscopy can be used until the stated expiry date.

After first opening of the bottle, the contents can be used up to the stated expiry date when stored at +15 °C to +25 °C.

The bottles must be kept tightly closed at all times.

Capacity

approx. 1500 stainings / 500 ml

Additional instructions

For professional use only.

In order to avoid errors, the application must be carried out by qualified personnel only.

National guidelines for work safety and quality assurance must be followed. Microscopes equipped according to the standard must be used.

Protection against infection

Effective measures must be taken to protect against infection in line with laboratory guidelines.

Instructions for disposal

The package must be disposed of in accordance with the current disposal guidelines.

Used solutions and solutions that are past their shelf-life must be disposed of as special waste in accordance with local guidelines. Information on disposal can be obtained under the Quick Link "Hints for Disposal of Microscopy Products" at www.microscopy-products.com. Within the EU the currently applicable REGULATION (EC) No 1272/2008 on classification, labelling and packaging of substances and mixtures, amending and repealing. Directives 67/548/EEC and 1999/45/EC, and amending Regulation (EC) No 1907/2006 applies.

Auxiliary reagents

Cat. No. 100496	Formaldehyde solution 4%, buffered, pH 6.9 (approx. 10% Formalin solution) for histology	350 ml and 700 ml (in bottle with wide neck), 5 l, 10 l, 10 l Titripac®
Cat. No. 100579	DPX new non-aqueous mounting medium for microscopy	500 ml
Cat. No. 100974	Ethanol denatured with about 1 % methyl ethyl ketone for analysis EMSURE®	1 l, 2.5 l
Cat. No. 101352	May-Grünwald's eosin methylene blue for microscopy	25 g, 100 g
Cat. No. 101728	OSTEOSOFT® mild decalcifier-solution for histology	1 l, 10 l Titripac®
Cat. No. 104699	Immersion oil for microscopy	100-ml dropping bottle, 100 ml, 500 ml
Cat. No. 106009	Methanol for analysis EMSURE® ACS, ISO, Reag. Ph Eur	1 l, 2.5 l, 5 l
Cat. No. 107961	Entellan® new rapid mounting medium for microscopy	100 ml, 500 ml, 1 l
Cat. No. 108298	Xylene (isomeric mixture) for histology	4 l
Cat. No. 109016	Neo-Mount® anhydrous mounting medium for microscopy	100-ml dropping bottle, 500 ml
Cat. No. 109204	Giemsa's azur eosin methylene blue solution for microscopy	100 ml, 500 ml, 1 l, 2.5 l
Cat. No. 109468	Buffer tablets pH 7.2 for preparing buffer solution acc. to WEISE for staining of blood smears	100 tabs
Cat. No. 109843	Neo-Clear® (xylene substitute) for microscopy	5 l
Cat. No. 111373	Buffer tablets pH 6.4 for preparing buffer solution acc. to WEISE for the staining of blood smears	100 tabs
Cat. No. 111374	Buffer tablets pH 6.8 for preparing buffer solution acc. to WEISE for the staining of blood smears	100 tabs
Cat. No. 111609	Histosec® pastilles solidification point 56-58°C embedding agent for histology	1 kg, 10 kg (4x 2.5 kg), 25 kg
Cat. No. 115161	Histosec® pastilles (without DMSO) solidification point 56-58°C embedding agent for histology	10 kg (4x 2.5 kg), 25 kg

Hazard classification

Cat. No. 101424

Please observe the hazard classification printed on the label and the information given in the safety data sheet.

The safety data sheet is available on the website and on request.

Main components of the product

Cat. No. 101424

C.I. 52015	0.7 g/l
C.I. 45380	0.5 g/l

contains CH₃OH
1 l = 0.79 kg

Other IVD products

Cat. No. 100869	Entellan® new for cover slipper for microscopy	500 ml
Cat. No. 101383	Wright's eosin methylene blue solution for microscopy	100 ml, 500 ml, 2.5 l
Cat. No. 102439	Eosin Y-solution 0.5%, alcoholic for microscopy	500 ml, 2.5 l
Cat. No. 103999	Formaldehyde solution min. 37% free from acid stabilized with about 10% methanol and calcium carbonate for histology	1 l, 2.5 l, 25 l
Cat. No. 105174	Hematoxylin solution modified acc. to Gill III for microscopy	500 ml, 1 l, 2.5 l
Cat. No. 105175	Hematoxylin solution modified acc. to Gill II for microscopy	500 ml, 2.5 l
Cat. No. 105387	Leishman's eosin methylene blue solution modified for microscopy	500 ml
Cat. No. 109844	Eosin Y-solution 0.5% aqueous for microscopy	1 l, 2.5 l
Cat. No. 111661	Hemacolor® Rapid staining of blood smear staining set for microscopy	1 set
Cat. No. 117081	Eosin Y solution 1%, alcoholic for microscopy	1 l

Literature

1. Löffler, H., Rastetter, J., Haferlach, T, Atlas der klinischen Hämatologie, 2004, Springer-Verlag Berlin Heidelberg
2. Routine Cytological Staining Techniques: Theoretical Background and Practice, Mathilde E. Boon, Johanna S. Drijver, 1986, Elsevier Science Publishing Company
3. Conn's Biological Stains: A Handbook of Dyes, Stains and Fluorochromes for Use in Biology and Medicine, 10th Edition, (ed. Horobin, R.W. and Kiernan, J.A). Bios, 2002



Consult instructions for use



Manufacturer



Catalog number



Batch code



Caution, consult accompanying documents



Use by YYYY-MM-DD



Temperature limitation

Status: 2017-08-29

Merck KGaA, 64271 Darmstadt, Germany
Tel. +49(0)6151 72-2440
www.microscopy-products.com

EMD Millipore Corporation, 290 Concord Road, Billerica, MA 01821, USA, Tel. +1-978-715-4321

