

Declaration of Conformity

Manufacturer's Name : SHIMADZU CORPORATION
Address : Analytical & Measuring Instruments Division
1, Nishinokyo-Kuwabara-cho, Nakagyo-ku,
Kyoto 604-8511, Japan

declares in sole responsibility that the following product

Product Name : Table-Top Type Universal Testing Instruments
Model Name : EZ Test series (EZ-SX, EX-LX, EZ-LX HS)
P/N : Depend on configuration. See Appendix 1

referred to in this declaration conforms to the following directives and standards

Machinery directive 2006/42/EC
EN ISO 12100:2010, EN60204-1:2006+A1:2009
EMC directive 2014/30/EU
EN61326-1:2013
RoHS directive 2011/65/EU

- Note 1) This declaration becomes invalid if technical or operational modifications are introduced without manufacturer's consent.
- Note 2) This declaration is valid if this product is used alone or in combination with the accessories of this product which are mentioned in attached Appendix 1 or other instruments which fulfill with the requirement of mentioned directive.
- Note3) Importer/Distributor and Authorised Representative in EU is as follows:
SHIMADZU EUROPA GmbH
Address :Albert-Hahn-Strasse 6-10, 47269 Duisburg, F.R. Germany

Kyoto, JAPAN 17 Oct. 2017

Place and date of issue

N Ohsaka

Naoki Ohsaka

General Manager of Quality Assurance Department
Analytical & Measuring Instruments Division



SHIMADZU

Appendix 1

Product Name : Table-Top Type Universal Testing Instruments

Final assembly

Model No.	P/N
EZ-LX SET ASSY EU	337-01311-30, 346-57100-21 (EZ-LX)
EZ-LX HS SET ASSY EU	337-01312-30, 346-57100-25 (EZ-LX HS)
EZ-SX SET ASSY EU	337-01313-30, 346-57110-21 (EZ-SX)

Product Options

Model No.	P/N
-----------	-----

新規・改訂内容一覧表

担当部署： 品証部

改訂記号	改訂理由	改訂内容	作成者 日付	検討者 日付	承認者 日付
制定			清水 '15.01.07	遠矢 2015.01.07	谷水 2015.01.07
A	最新規格に対応	EN61326-1:2003、EN ISO 12100 :2010、2014/30/EU 対応 低電圧指令記載削除	清水 2017.07.24	分析品証 17.7.27 遠矢	分析品証 17.7.27 逢坂
B	情報追記	Final Assy o P/W 追記	分析品証 17.10.17 清水	分析品証 17.10.19 遠矢	分析品証 17.10.19 逢坂

配付先