



宁波海曙昊润医疗器械有限公司
NINGBO HAISHU HAORUN MEDICAL DRESSING CO.,LTD

NANZHAN ROAD NO.66 HAISHU NINGBO ZHEJIANG CHINA 315000
www.haorunmedical.com sales@haorunmed.com TEL:+8657487706193

QUALITY TEST REPORT

Commodity	Non-Sterile cotton wool 250g		Size	250G
Batch No	202204		Sample Quantity	50ROLLS
Mfg.Date	2022-04-30		Test Date	2022-04-30
Report Date	2022-04-30			
Testing Items		Quality standard		Results
Loss on drying		Not more than 8.0 per cent, determined on 5.000 g by drying in an oven at 105 ° C.		6.7%
Foreign Fibres		Exam from microscope, it is seen to consist exclusively of typical cotton fibres, except that occasionally a few isolated foreign fibres may be present.		Comply
Absorbency	Sinking	10s		3.6s
	Waterholdi	23.00gg-1		25.8g
Acidity & Alkalinity		To 25 mL of solution S add 0.1 mL of phenolphthalein solution R and to another 25 mL add 0.05 mL of methyl orange solution R. Neither solution is pink.		Comply
Easily oxidized substances		Add the KMnO4 into 40ml testing water,the red color will be not disappeared within 5min.		Comply
Surface-active Substances		After 5 min, any foam present must not cover the entire surface of the liquid.		Comply
Ether-soluble substances		Not more than 0.5% for 100ml testing water.		0.23%
Fluorescence		No intense blue fluorescence.		Comply

The conclusion is OK

Signature and stamp

宁波海曙昊润医疗器械有限公司
NINGBO HAISHU HAORUN MEDICAL DRESSING CO.,LTD