

CK-MB STAT CalSet

REF 05957656 190

→ 4 x 1.0 mL

For USA: Elecsys CK-MB STAT CalSet

English

Intended use

CK-MB STAT CalSet is used for calibrating the quantitative Elecsys CK-MB STAT assay on the Elecsys and **cobas e** immunoassay analyzers.

Summary

CK-MB STAT CalSet is a lyophilized human serum with added CK-MB in two concentration ranges.

The CalSet can be used with all reagent lots.

Reagents - working solutions

- CKMBSTAT Cal1: 2 bottles, each for 1.0 mL of calibrator 1
- CKMBSTAT Cal2: 2 bottles, each for 1.0 mL of calibrator 2

CK-MB (human) in two concentration ranges (approximately 1.5 ng/mL and approximately 25 ng/mL) in a human serum matrix.

The exact lot-specific calibrator values are encoded in the barcode.

Calibrator values

Traceability: The Elecsys CK-MB STAT assay has been standardized against the previous Elecsys CK-MB STAT assay (for details see Method Sheet of the Elecsys CK-MB STAT assay).

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

For USA: For prescription use only.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{1,2}

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Handling

Carefully dissolve the contents of one bottle by adding exactly 1.0 mL of distilled or deionized water and allow to stand closed for 15 minutes to reconstitute. Mix carefully, avoiding foam formation.

Transfer the reconstituted calibrators into the supplied empty labeled snap-cap bottles.

Elecsys 2010 and **cobas e 411** analyzers: The reconstituted calibrators should only be left on the analyzers during calibration at 20-25 °C. After use, close the bottles as soon as possible and store upright at 2-8 °C.

Due to possible evaporation effects, not more than 5 calibration procedures per bottle set should be performed.

If necessary, freeze in aliquots; see section on **cobas e 601** and **cobas e 602** analyzers.

cobas e 601 and **cobas e 602** analyzers: Unless the entire volume is necessary for calibration on the analyzers, transfer aliquots of the reconstituted calibrators into empty snap-cap bottles (CalSet Vials). Attach the supplied labels to these additional bottles. Store the aliquots at -20 °C for later use. Allow the thawed bottles to stand closed for 15 minutes at 20-25 °C and mix carefully before use, avoiding foam formation.

Perform **only one** calibration procedure per aliquot.

Please note: Both the vial labels, and the additional labels (if available) contain 2 different barcodes. The barcode between the yellow markers is for **cobas 8000** systems only. If using a **cobas 8000** system, please turn the vial cap 180° into the correct position so the barcode can be read by the system. Place the vial on the instrument as usual.

Storage and stability

Store at 2-8 °C.

The lyophilized calibrators are stable up to the stated expiration date.

| Stability of the reconstituted calibrators: | |
|--|-----------------------------|
| either at -20 °C | 3 months (freeze only once) |
| or at 2-8 °C | 2 weeks |
| on Elecsys 2010 and cobas e 411 analyzers at 20-25 °C | up to 5 hours |
| on cobas e 601 and cobas e 602 analyzers at 20-25 °C | use only once |

Store calibrators **upright** in order to prevent the calibrator solution from adhering to the snap-cap.

Materials provided

- CK-MB STAT CalSet, barcode card, 4 empty labeled snap-cap bottles, 2 x 6 bottle labels

Materials required (but not provided)

- REF 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
- Elecsys 2010 or **cobas e** immunoassay analyzers and Elecsys CK-MB STAT assay reagents
- Distilled or deionized water

See the assay Method Sheet and the operator's manual for additionally required materials.

Assay

Place the reconstituted calibrators (in the system-compatible bottles with barcoded labels) in the sample zone.

Read in all the information necessary for calibrating the assay.

Ensure the calibrators are at 20-25 °C prior to measurement.

References

- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

| | |
|---|---|
|  | Contents of kit |
|  | Analyzers/Instruments on which reagents can be used |
|  | Reagent |
|  | Calibrator |
|  | Volume after reconstitution or mixing |
|  | Global Trade Item Number |

CK-MB STAT CalSet



FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS E and ELECSYS are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics



 Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com



Distribution in USA by:
Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336