

# SURESCAN® MRI CARDIAC DEVICES

Full Body 1.5T and 3T MRI ACCESS\*

Suderinamas su 1,5 Teslos magnetiniu lauku

## SIMPLE CONDITIONS IN ANY COMBINATION



SureScan Dual and Single Chamber Pacemakers

Product Name	Model Number
Advisa DR MRI®	A3DR01
Advisa SR MRI®	A3SR01
Ensura DR MRI®	EN1DR01
Ensura SR MRI®	EN1SR01
EnRhythm MRI®**	EMDR01**



SureScan Dual and Single Chamber ICDs

Product Name	Model Number
Evera MRI® XT DR	DDMB2D4
Evera MRI® S DR	DDMC3D4
Evera MRI® XT VR	DVMB2D4
Evera MRI® S VR	DVMC3D4
Visia AF MRI™ XT	DVFB2D4
Visia AF MRI™ S	DVFC3D4



SureScan CRT-Ds\*\*\*

Product Name	Model Number
Claria MRI® Quad CRT-D	DTMA2QQ
Claria MRI® CRT-D	DTMA2D4
Amplia MRI® Quad CRT-D	DTMB2QQ
Amplia MRI® CRT-D	DTMB2D4
Compia MRI® Quad CRT-D	DTMC2QQ
Compia MRI® CRT-D	DTMC2D4



SureScan Pacing Leads

Product Name	Model Number (length)
CapSureFix MRI®	5086MRI (45, 52, 58 cm)
CapSureFix Novus MRI®	4076 and 5076 (35, 45, 52, 58, 65, 85 cm)
CapSure Sense MRI®	4074 (52, 58 cm) 4574 (45, 53 cm)
CapSure Z Novus MRI SureScan™	5054 (52, 58 cm) 5554 (45, 53 cm)



SureScan Defibrillation Leads

Product Name	Model Number (length)
Sprint Quattro Secure S MRI™	6935M (55, 62 cm)
Sprint Quattro Secure MRI™	6947M (55, 62 cm)



SureScan Left-Heart Leads

Product Name	Model Number (length)
Attain Stability® MRI	4796 or 20066 (88 cm)
Attain® Performa® MRI	4298, 4398, 4598 (78, 88 cm)
Attain Ability® MRI	4196, 4296, 4396 (78, 88 cm)



Insertable Cardiac Monitors

Product Name	Model Number
Reveal® XT	9529
Reveal LINQ®	LNQ11



Transcatheter Pacing System

Product Name	Model Number
Micra®	MC1VR01

Update: July 2016

Check for MR conditions for use and latest MR conditional products on:

[mrisurescan.com/productsearch](http://mrisurescan.com/productsearch)



Find product manuals on:  
[medtronic.com/manuals](http://medtronic.com/manuals)

\*MR-conditional.

\*\* EnRhythm MRI is not labelled for 3T, 1.5T conditions apply.

\*\*\* The Model 6725 pin plug accessory may be used as a part of the MRI CRT-D SureScan systems in place of a right atrial lead. The pin plug is MR-Conditional only when used with these systems.

**MORE INFORMATION VISIT: [mrisurescan.com](http://mrisurescan.com)**

**Brief Statement**

See the MRI SureScan® Technical Manual before performing an MRI Scan and Device Manual for detailed information regarding the implant procedure, indications, contraindications, warnings, precautions, and potential adverse events.



[www.medtronic.com/manuals](http://www.medtronic.com/manuals)

Consult instructions for use on this website. Manuals can be viewed using a current version of any major Internet browser. For best results, use Adobe Acrobat® Reader with the browser.

**Medtronic**

**Europe**  
Medtronic International  
Trading Sàrl.  
Route du Molliau 31  
Case postale  
CH-1131 Tolochenaz  
[medtronic.eu](http://medtronic.eu)  
Tel: +41 (0)21 802 70 00  
Fax: +41 (0)21 802 79 00

**United Kingdom/Ireland**  
Medtronic Limited  
Building 9  
Croxley Green Business Park  
Hatters Lane  
Watford  
Herts WD18 8WW  
[medtronic.co.uk](http://medtronic.co.uk)  
Tel: +44 (0)1923 212213  
Fax: +44 (0)1923 241004

UC201407136cEE © 2016 Medtronic.  
All Rights Reserved. Printed in Europe.