

Tina-quant D-Dimer**Order information**

REF	CONTENT	Analyzer(s) on which cobas c pack(s) can be used
03001245 322	Tina-quant D-Dimer (100 tests)	System-ID 07 6458 2 Roche/Hitachi cobas c 311, cobas c 501/502
11556495 216	D-Dimer Calibrator (1 × 0.5 mL) and Zero Standard (1 × 2.5 mL)	Code 680-685
11556509 216	D-Dimer Control I/II (2 × 0.5 mL)	Code 200/Code 201

English**System information**

For **cobas c** 311/501 analyzers:

D-DI: ACN 123 (Citratated plasma)

D-DIH: ACN 125 (Li-heparin plasma)

For **cobas c** 502 analyzer:

D-DI: ACN 8123 (Citratated plasma)

D-DIH: ACN 8125 (Li-heparin plasma)

Intended use

In vitro test for the quantitative immunological determination of fibrin degradation products (D-Dimer and X-oligomers) in human plasma on Roche/Hitachi **cobas c** systems.^{1,2}

In conjunction with a non-high clinical probability assessment, a normal (< 0.5 µg FEU/mL) result excludes deep vein thrombosis (DVT) and pulmonary embolism (PE) with high sensitivity.

Summary

Thrombin converts fibrinogen to soluble fibrin by cleaving the fibrinopeptides A and B. The fibrin monomers polymerize spontaneously. Active factor XIII links two D-domains and generates a solid fibrin clot. A new plasmin-resistant antigenic determinant ("D-Dimer") is produced. Fragments containing D-Dimer are accordingly formed during the degradation of a fibrin clot by plasmin.

A large proportion of the fibrin degradation products consist of high molecular weight X-oligomers. The Tina-quant D-Dimer assay has a strong affinity for these high molecular weight degradation products. Only in vitro or during lysis therapy does complete degradation to D-Dimer molecules take place.

D-Dimer is a very sensitive marker for the activation of coagulation. When D-Dimer values below the cutoff are obtained, deep venous thrombosis (DVT) of the lower limb and pulmonary embolism (PE) can be excluded with high sensitivity.^{3,4,5,6}

The evidence for the use of Tina-quant D-Dimer in exclusion diagnosis comes from prospective management studies.^{7,8,9}

In one such study of 812 outpatients with symptoms of DVT, Schutgens et al. found that the combination of a non-high clinical probability score and a normal Tina-quant D-Dimer concentration allowed rule-out of DVT with a sensitivity of 99.3 % and a Negative Predictive Value (NPV) of 99.4 %.⁷ This rule-out strategy was found to be very safe, with a failure rate of only 0.6 %. Only 1 of 176 patients with a non-high pretest probability and a normal D-Dimer developed thrombosis during the three month follow-up. In a study involving 202 patients with suspected PE, Leclercq et al. found that PE could be ruled out by a normal Tina-quant D-Dimer result combined with a non-high clinical probability score, with a sensitivity of 100 %, an NPV of 100 % and a failure rate of 0 %.⁹ Further supporting evidence comes from numerous other clinical studies.^{10,11,12,13,14,15,16,17,18,19,20}

The D-Dimer result should not be used in isolation but in combination with a clinical probability assessment like the Wells score. DVT/PE should only be excluded on the basis of a low or moderate (non-high) clinical probability and a normal (< 0.5 µg FEU/mL) Tina-quant D-Dimer result.

It has been reported that patients with a distal DVT or a subsegmental/peripheral PE may have a normal Tina-quant D-Dimer result.²¹ The clinical relevance of such small(er) thrombi is unclear. The good results obtained in the management studies where patients were treated based on the Tina-quant D-Dimer result and then followed-up for 3 months suggest that these smaller thrombi do not result in adverse patient outcomes.

In disseminated intravascular coagulation (DIC)/consumptive coagulopathy, fibrin degradation products are a sensitive marker. Monitoring the fibrin-specific degradation products can be used to

- confirm or refute a tentative diagnosis
- estimate the potential risk for patients with existing DIC
- monitor an initiated therapy

Apart from DVT, PE, and DIC, D-Dimer may reflect other causes associated with fibrin formation such as trauma, pregnancy complications, malignant disease or vascular abnormalities. Elevated D-Dimer levels therefore have to be interpreted in the context of possible underlying diseases and clinical symptoms.^{22,23,24}

Test principle

Particle enhanced immunoturbidimetric assay.

Latex particles of uniform size are coated with monoclonal antibodies (F(ab')₂ fragments) to the D-Dimer epitope. The antigen/antibody complexes produced by the addition of samples containing D-Dimer lead to an increase in the turbidity of the test reactants. The change of absorbance with time is dependent on the concentration of D-Dimer epitopes in the sample. The precipitate is determined turbidimetrically.

Reagents - working solutions

R1 TRIS/HCl buffer: 370 mmol/L, pH 8.2; NaCl: 267 mmol/L

R2 Latex particles coated with monoclonal anti-human D-Dimer antibodies (mouse): 0.15 %

R1 is in position A and R2 is in position B.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

For USA: For prescription use only.

Reagent handling

Ready for use

Mix **cobas c** pack well before placing on the analyzer.

Carefully invert reagent container several times prior to use to ensure that the reagent components are mixed.

Storage and stability

D-DI

Shelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 12 weeks

Specimen collection and preparation

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable. Citratated plasma.

Collect venous blood using standard sampling tubes for clotting tests; employ sterile 0.11 molar sodium citrate solution. Maintain a precise mixture of 1 + 9 for sodium citrate and blood.

If necessary, pipette off the supernatant and store in a stoppered plastic tube.

Li-heparin plasma may also be used.²⁵ Unlike when using citratated tubes, there is no sample dilution with heparin tubes. Therefore D-Dimer values in heparin plasma are on average 19 % higher over the entire measuring range. However, by using adjusted calibrator and control values, identical values are measured in patient specimens with both sample materials.

CAUTION. To avoid erroneous patient values, we recommend that all D-Dimer measurements are performed uniformly in the laboratory from either citratated plasma or heparin plasma.

Tina-quant D-Dimer

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Thaw frozen samples completely at 37 °C and then mix thoroughly. Leave to stand for 15 minutes at room temperature before use; then assay immediately. Once thawed, a sample may not be refrozen for coagulation analysis.

Use the samples undiluted.

Centrifuge samples containing precipitates before performing the assay.

Stability: ²⁶	8 hours at 15-25 °C
	4 days at 2-8 °C
	6 months at (-15)-(-25) °C

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- See "Order information" section
- General laboratory equipment

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

Application for plasma**cobas c 311 test definition**

Assay type	2-Point Rate		
Reaction time / Assay points	10 / 7-33		
Wavelength (sub/main)	800/660 nm		
Reaction direction	Increase		
Units	µg FEU/mL (mg FEU/L, ng FEU/mL)		
Reagent pipetting	Diluent (H ₂ O)		
R1	90 µL	20 µL	
R2	90 µL	–	
<i>Sample volumes</i>	<i>Sample</i>	<i>Sample dilution</i>	
		Sample	Diluent (H ₂ O)
Normal	5.0 µL	–	–
Decreased	2.1 µL	–	–
Increased	5.0 µL	–	–

cobas c 501 test definition

Assay type	2-Point Rate		
Reaction time / Assay points	10 / 12-51		
Wavelength (sub/main)	800/660 nm		
Reaction direction	Increase		
Units	µg FEU/mL (mg FEU/L, ng FEU/mL)		
Reagent pipetting	Diluent (H ₂ O)		
R1	90 µL	20 µL	
R2	90 µL	–	
<i>Sample volumes</i>	<i>Sample</i>	<i>Sample dilution</i>	

		Sample	Diluent (H ₂ O)
Normal	5.0 µL	–	–
Decreased	2.1 µL	–	–
Increased	5.0 µL	–	–

cobas c 502 test definition

Assay type	2-Point Rate	
Reaction time / Assay points	10 / 12-51	
Wavelength (sub/main)	800/660 nm	
Reaction direction	Increase	
Units	µg FEU/mL (mg FEU/L, ng FEU/mL)	
Reagent pipetting	Diluent (H ₂ O)	
R1	90 µL	20 µL
R2	90 µL	–

<i>Sample volumes</i>	<i>Sample</i>	<i>Sample dilution</i>	
		Sample	Diluent (H ₂ O)
Normal	5.0 µL	–	–
Decreased	2.1 µL	–	–
Increased	10.0 µL	–	–

Calibration

Calibrators	S1-S6: D-Dimer Calibrator
	Refer to the calibrator package insert for instructions on dilution of standards.
Calibration mode	RCM2
Calibration frequency	Full calibration
	<ul style="list-style-type: none"> • after reagent lot change • every 6 months when using a single reagent lot • as required following quality control procedures

Traceability: This method has been standardized against the Asserachrom D-Dimer method.²⁷

Quality control

For quality control, use control materials as listed in the "Order information" section.

In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

Roche/Hitachi **cobas c** systems automatically calculate the analyte concentration of each sample.

Conversion factors:	µg FEU/mL = mg FEU/L
	µg FEU/mL x 1000 = ng FEU/mL

Limitations - interference

Criterion: Recovery within ± 10 % of initial values at a D-Dimer concentration of 0.5 µg FEU/mL.

Icterus:²⁸ No significant interference up to an I index of 20 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 342 µmol/L or 20 mg/dL).

Hemolysis:²⁸ No significant interference up to an H index of 500 (approximate hemoglobin concentration: 310 µmol/L or 500 mg/dL).

Lipemia:²⁸ No significant interference up to an L index of 750. There is a poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Rheumatoid factors up to 100 IU/mL do not interfere.

High dose hook-effect: No false result occurs up to a D-Dimer concentration of 220 µg FEU/mL.

Drugs: No interference was found at therapeutic concentrations using common drug panels.^{29,30}

Heparin concentrations up to 1.0 IU/mL do not interfere.

Others: High concentrations of D-fragments, as can occur during lysis therapy, lead to depressed measurements.

In rare cases, IgM, particularly in samples from patients with myeloma, can give falsely high results.

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.³¹

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

ACTION REQUIRED

Special Wash Programming: The use of special wash steps is mandatory when certain test combinations are run together on Roche/Hitachi **cobas c** systems. The latest version of the carry-over evasion list can be found with the NaOHD-SMS-SmpCln1+2-SCCS Method Sheets. For further instructions refer to the operator's manual. **cobas c** 502 analyzer: All special wash programming necessary for avoiding carry-over is available via the **cobas** link, manual input is not required.

Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.

Limits and ranges

Measuring range

0.10-9.00 µg FEU/mL

Determine samples having higher concentrations via the rerun function. Dilution of samples via the rerun function is a 1:2.4 dilution. Results from samples diluted using the rerun function are automatically multiplied by a factor of 2.4.

Lower limits of measurement

Lower detection limit of the test

0.10 µg FEU/mL

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying 3 standard deviations above that of the lowest standard (standard 1 + 3 SD, repeatability, n = 21).

Expected values³²

< 0.5 µg fibrinogen equivalent/mL (µg FEU/mL).

The stated fibrinogen equivalent is based on the quantity of fibrinogen used in the preparation of the original Asserachrom standard.

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using human samples and controls in an internal protocol with repeatability (n = 21) and intermediate precision (3 aliquots per run, 1 run per day, 21 days). The following results were obtained:

<i>Repeatability</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>
	<i>µg FEU/mL</i>	<i>µg FEU/mL</i>	<i>%</i>
Control 1	0.88	0.02	1.9

Control 2	3.57	0.03	0.8
Human plasma 1	0.39	0.03	8.2
Human plasma 2	2.09	0.03	1.7
<i>Intermediate precision</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>
	<i>µg FEU/mL</i>	<i>µg FEU/mL</i>	<i>%</i>
Control 1	0.87	0.02	2.8
Control 2	3.55	0.05	1.3
Human plasma 3	0.33	0.05	13.7
Human plasma 4	1.97	0.06	3.1

Method comparison

D-Dimer values for human citrated plasma samples obtained on a Roche/Hitachi **cobas c** 501 analyzer (y) were compared with those determined using the corresponding reagent on a Roche/Hitachi 917 analyzer (x).

Sample size (n) = 72

Passing/Bablok³³

Linear regression

y = 0.984x + 0.132 µg FEU/mL

y = 0.926x + 0.223 µg FEU/mL

τ = 0.894

r = 0.995

The sample concentrations were between 0.305 and 7.86 µg FEU/mL.

Clinical performance in the exclusion of DVT

Tina-quant D-Dimer was used in a multicenter management study involving 812 outpatients with suspected DVT.⁷ Using the Wells probability assessment score, patients were classified as having a high (> 3) or non-high (< 3) pretest probability of DVT. The Tina-quant D-Dimer test was then performed using a cutoff of 0.5 µg FEU/mL. Those patients having a normal (negative) D-Dimer test result and a non-high pretest probability had no further diagnostic testing and were followed up for 3 months for development of DVT. Only one of 176 such patients developed DVT during the follow-up period. The performance characteristics of the Tina-quant D-Dimer assay in conjunction with a non-high pretest probability is summarized below:

Sensitivity:	99.3 %	(95 % CI: 96.4-100 %)
Negative Predictive Value:	99.4 %	(95 % CI: 96.9-100 %)
Specificity:	45.8 %	(95 % CI: 40.7-51.0 %)
Positive Predictive Value:	42.0 %	(95 % CI: 36.8-47.3 %)
Failure Rate:	0.6 %	(95 % CI: 0.02-3.1 %)

Clinical performance in the exclusion of PE

Tina-quant D-Dimer was used in a management study involving 202 patients with suspected PE.⁹ Using the Wells clinical model for PE probability,³⁴ patients were classified as having a low, moderate, or high pretest probability of PE. The Tina-quant D-Dimer test was then performed using a cutoff of 0.5 µg FEU/mL. Those patients having a normal (negative) D-Dimer test result and a non-high (low or moderate) pretest probability had no further diagnostic testing and were followed up for 3 months for development of PE. No patients developed PE during the follow-up period. The performance characteristics of the Tina-quant D-Dimer in conjunction with a non-high pretest probability is summarized below:

Sensitivity:	100.0 %	(95 % CI: 91.8-100.0 %)
Negative Predictive Value:	100.0 %	(95 % CI: 94.4-100.0 %)
Specificity:	50.4 %	(95 % CI: 41.4-59.4 %)
Positive Predictive Value:	40.5 %	(95 % CI: 31.1-50.5 %)
Failure Rate:	0 %	(95 % CI: 0.0-5.6 %)

References

- Gaffney PJ. Fibrinolysis Supplement 2 1993;7:2-8.
- Matsuda M, Yoshida N, Terukina S, et al. Fibrinogen 4. Current basic and clinical aspects. Amsterdam/New York/Oxford: Elsevier Science Publishers 1990;43-48.

- 3 Agency for Healthcare Research and Quality, Evidence Report /Technology Assessment Number 68: Diagnosis and Treatment of Deep Venous Thrombosis and Pulmonary Embolism: Summary. AHRQ Pub No. 03-E012, January, Full report available online at www.ahrq.com 2003.
 - 4 American College of Emergency Physicians Board of Directors. Clinical Policy: Critical Issues in the Evaluation and Management of Adult Patients Presenting with Suspected Lower-Extremity Deep Venous Thrombosis Ann Em Med 2003;42(1):124.
 - 5 Ramzi DW, Leeper KV. DVT and Pulmonary Embolism: Part 1, Diagnosis. Am. Fam. Phys 2004;69(12):2829.
 - 6 American College of Emergency Physicians Board of Directors. Clinical Policy: Critical Issues in the Evaluation and Management of Adult Patients Presenting with Suspected Pulmonary Embolism. Ann Em Med 2003;41:257.
 - 7 Schutgens REG, Ackermack P, Haas FJLM, et al. Combination of a Normal D-Dimer Concentration and a Non-High Pretest Clinical Probability Score is a Safe Strategy to Exclude Deep Venous Thrombosis. Circulation 2003;107:593-597.
 - 8 Schutgens RE, Haas FJ, Biesma DH. Reduced efficacy of clinical probability score and D-Dimer assay in elderly subjects suspected of having deep vein thrombosis. Br J Haemat 2005;129:653-657.
 - 9 LeClerq LGL, Lusitan JG, Kooy MvM, et al. Ruling out clinically suspected pulmonary embolism by assessment of clinical probability and D-Dimer levels: a management study. Thromb Haemost 2003;89:97-103.
 - 10 Van Belle A, Büller HR, Huisman MV, et al. for the Christopher Study Investigators. Effectiveness of managing suspected pulmonary embolism using an algorithm combining clinical probability, D-dimer testing, and computed tomography. JAMA 2006. 295(2), 172-179.
 - 11 Knecht MF, Heinrich F. Clinical Evaluation of an Immunoturbidimetric D-Dimer Assay in the Diagnostic Procedure of Deep Vein Thrombosis and Pulmonary Embolism. Thromb Res 1997;88:413-417.
 - 12 Janssen MCH, Heebles AE, deMetz M, et al. Reliability of Five Rapid D-Dimer Assays Compared to ELISA in the Exclusion of Deep Venous Thrombosis. Thromb Haemost 1997;77(2):262-266.
 - 13 Lindahl TL, Lundahl TH, Fransson SG. Evaluation of an automated micro-latex D-Dimer assay (Tina-quant on Hitachi 911) in symptomatic outpatients. Thromb Haemost 1999;82(6):1772-1773.
 - 14 Van der Graaf F, van den Borne H, van der Kolk M, et al. Exclusion of deep venous thrombosis with D-dimer testing--comparison of 13 D-dimer methods in 99 outpatients suspected of deep venous thrombosis using venography as reference standard. Thromb Haemost 2000;83(2):191-198.
 - 15 Fünfsinn N, Caliezi F, Biasutti FD, et al. Rapid D-Dimer testing and pre-test clinical probability in the exclusion of deep venous thrombosis in symptomatic outpatients. Blood Coagul Fibrinolysis 2001;12:165-170.
 - 16 Diamond S, Goldwbeber R, Katz S. Use of D-Dimer to aid in excluding deep venous thrombosis in ambulatory patients. Am J Surg 2005;189:23-26.
 - 17 Schutgens RE, Haas FJ, Gerritsen WB, et al. The usefulness of five D-Dimer assays in the exclusion of deep venous thrombosis. J Thromb Haemost 2003;1:976-981.
 - 18 Stolba R, Lenglinger FX, Rezanka E, et al. Diagnostic Value of a new, quantitative D-Dimer assay for the exclusion of pulmonary embolism in symptomatic patients. J Lab Med 2000;24(3):153-157.
 - 19 De Monyé W, Sanson B-J, Büller HR, et al. ANTELOPE study group. The performance of two rapid quantitative D-Dimer assays in 287 patients with clinically suspected pulmonary embolism. Thromb Res 2002;107:283-286.
 - 20 Söhne M, Kamphuisen PW, van Mierlo PJWB, et al. Diagnostic strategy using a modified clinical decision rule and D-Dimer test to rule out pulmonary embolism in elderly in- and outpatients. Thromb Haemost 2005;94(1):206-210.
 - 21 Jennersjö C, Fagerberg I, Karlander S, et al. Normal D-Dimer concentration is a common finding in symptomatic outpatients with distal deep vein thrombosis. Blood Coagul Fibrinolysis 2005;16:517-523.
 - 22 Angstwurm MW, Reiningner AJ, Spannagl M. D-Dimer as marker for microcirculatory failure: correlation with LOD and APACHE II scores. Thromb Res 2004;113(6):353-359.
 - 23 Wakai A, Gleeson A, Winter D. Role of fibrin D-Dimer testing in emergency medicine. Emerg Med J 2003;20:319-325.
 - 24 Dempfle CE. Bestimmung des D-Dimer-Antigens in der klinischen Routine. 102, Ausgabe 24 vom 17.06.2005.
 - 25 Schutgens REG, Haas FJML, Ruven HJT, et al. No Influence of Heparin Plasma and Other (Pre)analytic variables on D-Dimer Determinations. Clin Chem 2002;48(9):1611-1613.
 - 26 Guder WG, Narayanan S, Wisser H, et al. List of Analytes; Preanalytical Variables. Brochure in: Samples: From the Patient to the Laboratory. Darmstadt: GIT-Verlag 1996.
 - 27 Adema E, Gebert U. Pooled patient samples as reference material for D-Dimer. Thromb Res 1995;80(1):85-88.
 - 28 Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. Clin Chem 1986;32:470-475.
 - 29 Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". Eur J Clin Chem Clin Biochem 1996;34:385-386.
 - 30 Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem 2001;38:376-385.
 - 31 Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med 2007;45(9):1240-1243.
 - 32 Dempfle CE, Hafner G, Lestin HG, et al. Multizentrische Evaluierung von Tina-quant D-Dimer. J Lab Med 1996;20:31-37.
 - 33 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.
 - 34 Wells PS, Ginseberg JF, Anderson DR, et al. Use of a clinical model for safe management of patients with suspected pulmonary embolism. Ann Intern Med 1998;129:997-1005.
- A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard:

CONTENT

Contents of kit



Volume after reconstitution or mixing

GTIN

Global Trade Item Number

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C and TINA-QUANT are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2016, Roche Diagnostics

0003001245322c501V9.0

D-DI

Tina-quant D-Dimer

cobas®



 Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

Distribution in USA by:
Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336

