

## TPHA TEST KIT

### For the detection of antibodies to *T.pallidum* in human Serum using micro haemagglutination.

**IVD** For In-Vitro diagnostic and professional use only

 Store at 2° to 8° C

#### INTENDED USE

For the detection of antibodies to *T. pallidum* in human serum based on haemagglutination.

#### INTRODUCTION

Syphilis is a venereal disease caused by the spirochaete micro-organism *Treponema pallidum*. As this organism cannot be cultured on artificial media the diagnosis of syphilis depends on the correlation of clinical data with the specific antibody demonstrated by serological tests. Serological screening tests for syphilis using cardiolipin and lecithin as antigens are simple to perform but biological false positive (BFP) reactions occur frequently because the tests use non-treponemal antigens.

The TPI and FTA-ABS tests utilize pathogenic *Treponema pallidum* as the antigen but these tests present some difficulties for routine serodiagnosis. The TPI test requires living pathogenic *T.Pallidum* and the FTA-ABS test requires a fluorescence microscope. Both tests require a high level of expertise.

TPHA test kit has been shown to be a convenient and specific test for the diagnosis of treponemal infection, having specificity similar to that of the TPI test and sensitivity comparable to that of the FTA-ABS test. It requires minimum laboratory equipment and is very simple to perform.

TPHA reagents are used to detect human serum antibody to *T.pallidum* by means of an indirect haemagglutination (IHA) method. Preserved avian erythrocytes are coated with antigenic components of pathogenic *T.pallidum* (Nichol's strain). These Test Cells agglutinate in the presence of specific antibodies to *T.pallidum*, and show characteristic patterns in microtitration plates.

Any non-specific reactions occurring are detected using the Control Cells, which are avian erythrocytes not coated with *T.pallidum* antigens. Non-specific reactions may also be absorbed out using these Control Cells.

Antibodies to non-pathogenic treponemes are absorbed by an extract of Reiter's treponemes, included in the cell suspension. Test results are obtained in 45-60 minutes and the cell agglutination patterns are both easily read and long lasting

#### MATERIALS

##### MATERIALS PROVIDED

- Test cells; preserved avian erythrocytes sensitised with *T.pallidum* antigen.
- Control cells; preserved avian erythrocyte.
- Diluent.
- Positive control serum; (prediluted 1:20), Use neat. This will give an equivalent titer of 1/640:/2560 in the quantitative test.
- Negative control serum; (prediluted 1:20), Use neat.
- Package Insert.

##### MATERIALS NEEDED BUT NOT PROVIDED

- Accurate pipettes for delivering 10:25:75 and 190 microlitres.
- U-Well microtitration plates.

#### PRECAUTIONS

The reagents and controls contain 0.1% sodium azide as a preservative. Avoid ingestion and contact with skin or mucus membrane. Normal laboratory precautions should be maintained while handling test reagents.

#### REAGENTS HANDLING

- All the reagents must be allowed to reach room temperature before use.
- Do not freeze any of the reagents.

#### REAGENTS STORAGE

- The kit should be stored at 2-8° C in an upright position at all times.
- Under these conditions, kit performance characteristics will be maintained for at least 15 or 18 months from date of manufacture. See expiry date on kit label.

- Reagents should be discarded if they become contaminated or do not demonstrate correct activity with the controls.
- The reagents in each kit have been standardized to produce the proper reaction and reagents should not be interchanged with those from other batches.

#### SAMPLE PREPARATION

- The test is designed for use with serum only.
- Plasma samples should not be used.
- The samples should be free from haemolysis and contamination.
- Serum samples may be stored at 2-8° C if a preservative is added prior to storage.
- For long term storage sera should be stored at -20° C Strictly avoid contaminating any of the reagents or serum dilutions with saliva. This will cause confusing patterns similar to positive results with specimens which should be negative.

#### PROCEDURES

##### QUALITATIVE METHOD

Each sample requires 3 wells of a microtitration plate.

1. Add 190  $\mu$ l of diluent to Well 1.
2. Add 10  $\mu$ l serum to Well 1. (Sample dilution 1:20).
3. Using a micropipette, mix contents of Well 1 and transfer 25  $\mu$ l to Wells 2 & 3.
4. Ensure that the Test and Control Cells are thoroughly resuspended. Add 75  $\mu$ l of control cells to Well 2. Add 75  $\mu$ l of Test Cells to Well 3.
5. Tap the plate gently to mix the contents thoroughly.
6. Incubate 45-60 minutes at room temperature.
7. Caution! Keep the plate away from heat, direct sunlight and any source of vibration.
8. Read results. Results are stable for 24hrs if the plate is covered and the above precautions are observed.

#### NOTE

Kit controls can be run in parallel and are diluted and ready for use.

##### QUANTITATIVE TEST

Each sample requires 8 Wells of a microtitration plate, Labeled A through to H.

1. Add 25  $\mu$ l of diluent to Wells B to H inclusive.

- Transfer 25 l of 1:20 serum dilution from screening test to Wells A and B.
- Take 25 l of diluted serum from Well B and serially dilute from Wells B to H inclusive in 25 l aliquots, discarding 25 l of diluted serum from Well H.
- Ensure that the Test Cells are thoroughly resuspended. Add 75 l of Test cells to wells A to H inclusive. This will give a dilution of serum of 1/ 80 in well A through 1/ 10240 Well H.
- Shake the plate gently to mix the contents thoroughly.
- Incubate for 45-60 minutes at room temperature.
- Caution! Keep the plate away from heat, direct sunlight and any source of vibration.
- Read results. Results are stable for 24hrs. if the plate is covered and the above precautions are observed.

## RESULTS

RESULTS	TEST CELLS	CONTROL CELLS
Strong Positive	Full cell pattern covering the bottom of the well.	No agglutination tight button
Weak Positive	Cell pattern covers approx. 1/3 of well bottom	No agglutination tight button
Indeterminate	Cell pattern shows a distinctly open center	No agglutination tight button
Negative	Cells settled to a compact bottom, typically with a small clear center.	No agglutination tight button
Non-specific *	Positive reaction	Positive reaction

### Non-specific absorption \*

- Add 10µ to a small tube then add 190 l of Control Cells. Mix well and stand for 30 minutes.
- Centrifuge for 15 minutes at 1000 rpm and test the supernatant by the qualitative method.

### Note:

If the result is repeatedly non-specific the sample should be tested by another method eg. Reagin or FTA-ABS.

### INTERPRETATION OF RESULTS.

#### INVALID

A serum that gives this result may be absorbed using the Control Cells as detailed under Non-specific absorption. A doubtful reaction with Test Cells should be reported as

#### INDETERMINATE

This result may indicate a low level of antibody in early primary syphilis or yaws. This sample should be first retested in the qualitative test then a further sample should be tested at a later date to determine whether or not there is a rising titer. It is also advisable to perform a regain test and/or another confirmation test (FTA-ABS) to complete the profile of the test serum.

### PERFORMANCE

#### SENSITIVITY

With clinical samples when compared to FTA-ABS and/or clinical diagnosis was 99.7% (298/299)

#### SPECIFICITY

With clinical samples was 99.3% (301/303).

### CROSS REACTIVITY

The following have all been shown not to interfere with the test results (10 clinical samples of each)

- Rheumatoid Factor.
- Post Hepatitis B vaccination.
- Genital Herpes.
- Leptospirosis.
- EBV Infection.
- SLE.
- Lyme's Disease.

### REFERENCES:

- Rathlev T. - Haemagglutination tests utilizing antigens from pathogenic and apathogenic Treponema pallidum WHO/VDT/RES 1965 ; 77 : 65.
- Tomizawa T, Kasamatsu S. - Haemagglutination tests for diagnosis of syphilis. A preliminary report. Japan. J. Med. Sci. Biol. 19, 305-308, 1966.
- Rathlev T. - Haemagglutination test utilizing pathogenic Treponema pallidum for the serodiagnosis of syphilis. Br J Vener Dis 1967 ; 43 : 181-5.

- Tomizawa T. Kasamatsu S. Yamaya S. - Usefulness of the haemagglutination test using Treponema pallidum antigen (TPHA) for the serodiagnosis of syphilis. Jap J Med Sci Biol 1969; 22: 341-50.
- Sequeira P, J, L. Eldridge A, E. – Treponemal Haemagglutination test. Br J Vener Dis 1973; 49: 242-8.
- Larsen S.A., Hambie E.A., et coll., Specificity, sensitivity and reproducibility among the fluorescent treponemal antibody absorption test, the microhemagglutination assay for Treponema pallidum antibodies, and the hemagglutination treponemal test for syphilis. J. Clin. Microbiol., 1981 ; 14 : 441 – 445.
- Houng H. - Syphilis: new diagnostic directions. Intern. J. STD and AIDS 1992; 3: 391-413.
- Sluis J.J. Van Der. - Laboratory Techniques in the diagnosis of syphilis: a review. Genitourin Med. 1992; 68 : 413-9.



**ATLAS Medical**

**William James House,  
Cowley Road, Cambridge, CB4 4WX, UK**

**Tel: ++44 (0) 1223 858 910**

**Fax: ++44 (0) 1223 858 524**

**PPI080A01**

**Rev E (18.06.2015)**

REF	Catalogue Number		Store at
IVD	For In-Vitro Diagnostic use		Caution
	Number of tests in the pack		Read product insert before use
LOT	Lot (batch) number		Manufacturer
	Fragile, handle with care		Expiry date
	Manufacturer fax number		Do not use if package is <b>damaged</b>
	Manufacturer telephone number		