

Polyflux L

DESIGNED FOR:

LFHD [Low flux]

MEMBRANE:

POLYAMIX [PAES/PVP/PA, BPA-free]

THE PROVEN BALANCE OF QUALITY AND PERFORMANCE IN LOW-FLUX

The **Polyflux L** dialyzer series is specialized for low-flux hemodialysis treatments, featuring a distinctive membrane acting as an effective barrier to potential fluid contaminants¹, while still delivering high performance². **Polyflux L** dialyzers are a good choice for proven biocompatible yet effective low-flux therapies, designed with safety in mind.

DESIGNED TO PROMOTE BIOCOMPATIBILITY²

The **Polyflux L** dialyzers are designed to deliver high-quality low-flux hemodialysis treatments.

- Since 1988, over 300 million **Polyflux** dialyzers have been used globally³
- The **Polyflux L** dialyzers are designed to prevent endotoxins from crossing the dialyzer membrane^{1,2}
- The **Polyflux L** dialyzers are steam sterilized inside-out, to promote biocompatibility, avoiding exposure to chemicals such as ethylene oxide and manufacturing residues^{4,5}

WITH HIGH PERFORMANCE IN MIND

The **Polyflux L** dialyzers feature an exclusive 3-layered membrane structure, designed to support a stable high performance over time.

- Effective clearance of standard dialysis markers, such as urea or phosphates⁶
- A clinical case series study suggests the **Polyflux L** dialyzers may reduce the signs and symptoms of hemodialysis-associated eosinophilia⁷



Polyflux L Specifications

	Pozicija 1	Pozicija 3	Pozicija 4		Pozicija 1	Pozicija 3	Pozicija 4
MATERIALS	POLYFLUX 14 L	POLYFLUX 17 L	POLYFLUX 21 L	CLEARANCES IN VITRO (mL/min)*	POLYFLUX 14 L	POLYFLUX 17 L	POLYFLUX 21 L
Membrane	Polyamix Polyarylethersulfone, Polyvinylpyrrolidone and Polyamide blend BPA-free			Urea (60 Da) (Q _B -Q _D , mL/min)			
Potting	Polyurethane (PUR)			200/500	190	194	
Housing	Polycarbonate (PC)			300/500	252	264	275
Gaskets	Silicone rubber (SIR)			400/500	293	310	328
Protection caps	Polypropylene (PP)			500/500		342	364
Sterilization	Steam (inside-out)			200/700	194	197	
Sterile barrier	Medical Grade Paper			300/700	267	276	285
				400/700	319	336	353
				500/700		380	403
SPECIFICATIONS				Creatinine (113 Da)			
UF-Coefficient (mL/(h*mmHg))*	10	12.5	15	200/500	171	179	
KoA urea*	851	1026	1268	300/500	214	230	246
Blood Compartment volume (mL)	81	104	123	400/500	241	262	283
Minimum recommended priming volume (mL)		500		500/500		284	310
Maximum TMP (mmHg)		600		200/700	178	185	
Recommended Q _B (mL/min)	200-400	200-500	300-500	300/700	229	244	258
Storage conditions	<30°C (or <86°F)			400/700	264	284	306
Units per box		24		500/700		313	341
Gross/net weight (g)	254/225	274/245	294/265	Phosphate (142 Da)			
				200/500	152	163	
MEMBRANE				300/500	183	200	218
Effective Membrane Area (m²)	1.4	1.7	2.1	400/500	203	224	247
Fiber inner diameter (µm)		215		500/500		240	267
Fiber wall thickness (µm)		50		200/700	160	170	
				300/700	197	213	231
				400/700	221	242	266
				500/700		264	292
				Vitamin B12 (1.4 kDa)			
				200/500	90	101	
				300/500	100	114	131
				400/500	106	122	142
				500/500		128	149
				200/700	96	107	
				300/700	107	121	138
				400/700	114	130	150
				500/700		137	159

* According to ISO 8637-1
– UF-Coefficient: measured with bovine blood, Hct 32%, Pct 60g/L, 37°C
– KoA urea: calculated at Q_B=300 mL/min, Q_D=500mL/min, UF=0 mL/min
– Clearances In-Vitro: measured at UF=0 mL/min, ±10%

1. Schepers E, Glorieux G, Eloot S, et al. *Assessment of the association between increasing membrane pore size and endotoxin permeability using a novel experimental dialysis simulation set up.* BMC Nephrology. 2018; 19:1.
2. Ronco C, et al. *Evolution of synthetic membranes for blood purification: the case of the Polyflux family.* Nephrol Dial Transplant 2003;18[Suppl 7]:vii10-20.
3. Baxter. Data on file. *Dialyzers Sales Report.* 2018.
4. Golli-Bennour EE, et al. *Cytotoxic effects exerted by polyarylsulfone dialyser membranes depend on different sterilization processes.* Int Urol Nephrol. 2011; 43:483-490.
5. D'Ambrosio FP, et al. *Ethylene oxide allergy in dialysis patients.* Nephrol Dial 1997;12:1461-1463.
6. Krause B, et al. *Polymeric Membranes for Medical Applications.* Chemie Ingenieur Technik. 2003; 75:1725-1732.
7. Tielemans C, et al. *Clinical assessment of Performance and Blood Compatibility Profile of a New Synthetic Low Flux Hemodialyzer.* Blood Purif. 2002; 20:214-215.

The capillary dialyzer/filter is intended for use in hemodialysis for the treatment of chronic or acute renal failure.
The products meet the applicable provisions of Annex I (Essential Requirements) and Annex II (Full quality assurance system of the Council Directive 93/42/EEC of 14 June 1993, amended by Directive 2007/47/EC).

For safe and proper use of the device, please refer to the Instructions for Use

