



05199549001V4.0

# PreciControl Troponin

cobas®

REF 05095107 190

→ 4 x 2.0 mL

REF 05095107 922 (QCS)

## English

### Intended use

PreciControl Troponin is used for quality control of the Elecsys Troponin T hs, Elecsys Troponin T hs STAT, Elecsys Troponin I and Elecsys Troponin I STAT immunoassays on the Elecsys and cobas e immunoassay analyzers.

### Summary

PreciControl Troponin is a lyophilized human serum with added troponin T and troponin I in two concentration ranges each. The controls are used for monitoring the accuracy and precision of the Elecsys Troponin T hs, Elecsys Troponin T hs STAT, Elecsys Troponin I and Elecsys Troponin I STAT immunoassays.

### Reagents - working solutions

- PC TN1: 2 bottles, each for 2.0 mL of control serum
- PC TN2: 2 bottles, each for 2.0 mL of control serum

Troponin T (recombinant, human) in two concentration ranges (approx. 30 ng/L or pg/mL and approx. 2500 ng/L or pg/mL) and troponin I (recombinant, human) in two concentration ranges (approx. 0.75 µg/L or ng/mL and approx. 18 µg/L or ng/mL) in a human serum matrix.

The exact target values and ranges are encoded in the barcodes as well as printed on the enclosed (or electronically available) value sheet.

### Target values and ranges

The target values and ranges were determined and evaluated by Roche. They were obtained using the Elecsys Troponin T hs, Elecsys Troponin T hs STAT, Elecsys Troponin I and Elecsys Troponin I STAT assay reagents and analyzers available at the time of testing.

If the target values and control ranges are updated, this information is conveyed either via the reagent barcodes, or control barcodes (or provided electronically) and in an additional value sheet included in the reagent kit. This value sheet lists all control lots to which the new values apply. If some of the values remain unchanged, the original values conveyed via the CBC (Control Barcode), and in the value sheet included in the control kit (or provided electronically), remain valid.

Results must be within the specified ranges. In the event that increasing or decreasing trends, or any other suddenly occurring deviations beyond the range limits are observed, all test steps must be checked.

Traceability information is given in the Method Sheet of the relevant Elecsys assay.

Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

### Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.<sup>1,2</sup>

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

### Handling

Carefully dissolve the contents of one bottle by adding exactly 2.0 mL of distilled or deionized water and allow to stand closed for **60 minutes** at 20-25 °C to reconstitute. Mix carefully, avoiding foam formation. Immediately transfer the reconstituted control into the empty labeled snap-cap bottles supplied. Where required, freeze aliquots immediately in

additional snap-cap bottles at -20 °C (ControlSet Vials). Attach the supplied labels to these additional bottles.

Allow the thawed/2-8 °C stored bottles to stand closed for **60 minutes** at 20-25 °C and mix carefully before use, avoiding foam formation.

### Storage and stability

Store at 2-8 °C.

The lyophilized control serum is stable up to the stated expiration date.

Stability of the reconstituted/thawed control serum:	
at 2-8 °C	4 days
at -20 °C	3 months (freeze only once)
on the analyzers at 20-25 °C	up to 5 hours
after thawing	use only once

Store controls **upright** in order to prevent the control solution from adhering to the snap-cap.

### Materials provided

- PreciControl Troponin, 2 barcode cards, control barcode sheet, 2 x 2 empty labeled snap-cap bottles, 2 x 6 bottle labels

### Materials required (but not provided)

- REF 03142949122, ControlSet Vials, 2 x 56 empty snap-cap bottles
- Elecsys 2010, MODULAR ANALYTICS E170 or cobas e immunoassay analyzers and assay reagents.
- Distilled or deionized water

See the assay Method Sheet and the operator's manual for additionally required material.

### Assay

Treat the reconstituted control serum in the system-compatible labeled bottles for analysis in the same way as patient samples.

Read the data into the analyzer.

Ensure the controls are at 20-25 °C prior to measurement.

Run controls daily in parallel with patient samples, once per reagent kit, and whenever a calibration is performed. The control intervals and limits should be adapted to each laboratory's individual requirements.

Follow the applicable government regulations and local guidelines for quality control.

### References

- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

### Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
→	Volume after reconstitution or mixing



# PreciControl Troponin

**cobas<sup>®</sup>**

COBAS, COBAS E, ELECSYS, MODULAR and PRECICONTROL are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)

