

8/14/2024

To Whom It May Concern

RE: Device noise level summary

The table below is a summary contain the estimate noise level of each pump device during therapy operation.

Unit	Noise level from 1.0m	Noise level from 0.3m
UNO	43.0 dB	49.8 dB
UNO+	46.5 dB	55.4 dB
XLR8+	47.0 dB	56.4 dB
DUO	50.5 dB	58.0 dB

1.1.3 Garso lygis 47 dB

Xiao-Ci Lin

QA/RA

Geandyne Biotechnologies, Inc

Xiao-Ci Lin

Digitally signed
by Xiao-Ci Lin
Date: 2024.08.15
16:10:18 -04'00'