

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>TSI/Basal max delay Alias: Basal seg. Max diff</p>	<p>(GEU-106-0009, 99GEMS, “Basal Segments Maximum Difference”)</p>	<p>(G-0373, SRT, “Image Mode”) = (GEU-106- 0024, 99GEMS, “Tissue Doppler Imaging”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0020, 99GEMS, “Tissue Synchronization Imaging”)</p>
<p>TSI/Basal stdev Alias: Basal stdev</p>	<p>(GEU-106-0010, 99GEMS, “Basal Standard Deviation”)</p>	<p>(G-0373, SRT, “Image Mode”) = (GEU-106- 0024, 99GEMS, “Tissue Doppler Imaging”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0020, 99GEMS, “Tissue Synchronization Imaging”)</p>
<p>TSI/All segments max delay Alias: All seg. Max diff.</p>	<p>(GEU-106-0012, 99GEMS, “All Segments Maximum Difference”)</p>	<p>(G-0373, SRT, “Image Mode”) = (GEU-106- 0024, 99GEMS, “Tissue Doppler Imaging”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0020, 99GEMS, “Tissue Synchronization Imaging”)</p>
<p>TSI/All segments stdev Alias: All segments stdev</p>	<p>(GEU-106-0012, 99GEMS, “All Segments Standard Deviation”)</p>	<p>(G-0373, SRT, “Image Mode”) = (GEU-106- 0024, 99GEMS, “Tissue Doppler Imaging”) (G-C036, SRT, “Measurement Method”) = (GEU-106-</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		0020, 99GEMS, “Tissue Synchronization Imaging”)
CO(A-L) Alias: CO(A-L)	(F-32100, SRT, “Cardiac Output”)	(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
CO(A-L A4C)/AutoHR Alias: CO A-L A4C	(F-32100, SRT, “Cardiac Output”)	(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
CI(A-L A4C)/AutoHR Alias: CI A-L A4C	(F-32110, SRT, “Cardiac Index”)	(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
CO(MOD A4C)/AutoHR Alias: CO MOD A4C	(F-32100, SRT, “Cardiac Output”)	(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)
<p>CI(MOD A4C)/AutoHR Alias: CI MOD A4C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CO(A-L A2C)/AutoHR Alias: CO A-L A2C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CI(A-L A2C)/AutoHR Alias: CI A-L A2C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT,</p>

		<p>“Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CO(MOD A2C)/AutoHR Alias: CO MOD A2C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CI(MOD A2C)/AutoHR Alias: CI MOD A2C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CO(A-L LAX)/AutoHR Alias: CO A-L LAX</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>CI(A-L LAX)/AutoHR Alias: CI A-L LAX</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CO(MOD LAX)/AutoHR Alias: CO MOD LAX</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CI(MOD LAX)/AutoHR Alias: CI MOD LAX</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVEDV(MOD BP)_03 Alias: LVEDV MOD BP</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVESV(MOD BP)_03 Alias: LVESV MOD BP</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>EF(Biplane)_03 Alias: EF Biplane</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>SV(Biplane)_03 Alias: SV Biplane</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>SI(Biplane)_03 Alias: SI Biplane</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>CO(Biplane)_03 Alias: CO Biplane</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>CI(Biplane)_03 Alias: CI Biplane</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>ECG/HeartRate/Auto Alias: HR</p>	<p>(8867-4, LN, “Heart rate”)</p>	
<p>2D/LV Major Alias: LV Major</p>	<p>(G-A193, SRT, “Major Axis”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/LV Minor Alias: LV Minor</p>	<p>(G-A194, SRT, "Minor Axis")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LV Minor D1 diastole Alias: LV Minor D1 diastole</p>	<p>(GEU-106-0183, 99GEMS, "Left ventricular minor axis dimension perpendicular to the septum")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32011, SRT, "End Diastole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0181, 99Gems, "LV mid level") (111031, DCM, "Image View") = (G-039B, SRT, "Parasternal short axis at the Papillary Muscle level")</p>
<p>2D/LV Minor D2 diastole Alias: LV Minor D2 diastole</p>	<p>(GEU-106-0184, 99GEMS, "Left ventricular minor axis dimension parallel to the septum")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32011, SRT, "End Diastole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0181, 99Gems, "LV mid level") (111031, DCM, "Image View") = (G-039B, SRT, "Parasternal short axis at the Papillary Muscle level")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/LVD Minor Base Alias: LVD Minor Base</p>	<p>(G-A194, SRT, "Minor Axis")</p>	<p>(G-0373, SRT, "Image Mode")= (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32011, SRT, "End Diastole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0182, 99Gems, "LV base level") (111031, DCM, "Image View") = (G-A19C, SRT, "Apical four chamber")</p>
<p>2D/LV Eccentricity Index diastole Alias: LV Eccentricity Index diastole</p>	<p>(GEU-106-0178, 99Gems, "LV Eccentricity Index in diastole")</p>	<p>(G-0373, SRT, "Image Mode")= (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32011, SRT, "End Diastole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0181, 99Gems, "LV mid level") (111031, DCM, "Image View") = (G-039B, SRT, "Parasternal short axis at the Papillary Muscle level")</p>
<p>2D/LV Minor D1 systole Alias: LV Minor D1 systole</p>	<p>(GEU-106-0183, 99GEMS, "Left ventricular minor axis dimension perpendicular to the septum")</p>	<p>(G-0373, SRT, "Image Mode")= (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (109070, DCM, "End Systole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0181,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>99Gems, "LV mid level") (111031, DCM, "Image View") = (G-039B, SRT, "Parasternal short axis at the Papillary Muscle level")</p>
<p>2D/LV Minor D2 systole Alias: LV Minor D2 systole</p>	<p>(GEU-106-0184, 99GEMS, "Left ventricular minor axis dimension parallel to the septum")</p>	<p>(G-0373, SRT, "Image Mode")= (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (109070, DCM, "End Systole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0181, 99Gems, "LV mid level") (111031, DCM, "Image View") = (G-039B, SRT, "Parasternal short axis at the Papillary Muscle level")</p>
<p>2D/LV Eccentricity Index systole Alias: LV Eccentricity Index systole</p>	<p>(GEU-106-0179, 99Gems, "LV Eccentricity Index in systole")</p>	<p>(G-0373, SRT, "Image Mode")= (G-03A2, SRT, "2D mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32011, SRT, "End Diastole") (G-C0E3, SRT, "Finding Site") = (GEU-106-0181, 99Gems, "LV mid level") (111031, DCM, "Image View") = (G-039B, SRT, "Parasternal short axis at the Papillary Muscle level")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/IVSd Alias: IVSd</p>	<p>(18154-5, LN, "Interventricular Septum Diastolic Thickness")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/IVSd (sax) Alias : IVSd (sax)</p>	<p>(18154-5, LN, "Interventricular Septum Diastolic Thickness")</p>	<p>(111031, DCM, "Image View") = (SRT, G-0397, "Parasternal short axis")</p>
<p>2D/LVIDd Alias: LVIDd</p>	<p>(29436-3, LN, "Left Ventricle Internal End Diastolic Dimension")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LVIDd (sax) Alias : LVIDd (sax)</p>	<p>(29436-3, LN, "Left Ventricle Internal End Diastolic Dimension")</p>	<p>(111031, DCM, "Image View") = (SRT, G-0397, "Parasternal short axis")</p>
<p>2D/LVIDs Alias: LVIDs</p>	<p>(29438-9, LN, "Left Ventricle Internal Systolic Dimension")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LVIDs Index Alias: LVIDs Index</p>	<p>(GEU-106-0029, 99GEMS, "Left Ventricle Internal Systolic Dimension Index")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LVPWd Alias: LVPWd</p>	<p>(18152-9, LN, "Left Ventricle Posterior Wall Diastolic Thickness")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/LVPWd (sax) Alias : LVPWd (sax)</p>	<p>(18152-9, LN, "Left Ventricle Posterior Wall Diastolic Thickness")</p>	<p>(111031, DCM, "Image View") = (SRT, G-0397, "Parasternal short axis")</p>
<p>2D/LVPWs Alias: LVPWs</p>	<p>(18156-0, LN, "Left Ventricle Posterior Wall Systolic Thickness")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LV Sphericity Index Alias : LV Sphericity Index</p>	<p>(GEU-106-0154, 99GEMS, "LV Sphericity Index")</p>	<p>(R-4089A, SRT, "Cardiac Cycle Point") = (SRT, F-32010, "Diastole")</p>
<p>2D/LVTTD Alias : LVTTD</p>	<p>(GEU-106-0155, 99GEMS, "LV thickness/dimension ratio")</p>	<p>(R-4089A, SRT, "Cardiac Cycle Point") = (DCM, 109070, "End Systole")</p>
<p>2D/IVSs Alias: IVSs</p>	<p>(18158-6, LN, "Interventricular Septum Systolic Thickness")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LVOT Diam Alias: LVOT Diam</p>	<p>(G-038F, SRT, "Cardiovascular Orifice Diameter")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract") (G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LV FAC Alias: LV FAC</p>	<p>(G-0376, SRT, "Left Ventricular Fractional Area Change")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/EDV(Teich) Alias: EDV(Teich)</p>	<p>(18026-5, LN, "Left Ventricular End Diastolic Volume")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (G-C036, SRT, "Measurement Method") = (125209, DCM, "Teichholz")</p>
<p>2D/ESV(Teich) Alias: ESV(Teich)</p>	<p>(18148-7, LN, "Left Ventricular End Systolic Volume")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (G-C036, SRT, "Measurement Method") = (125209, DCM, "Teichholz")</p>
<p>2D/EF(Teich) Alias: EF(Teich)</p>	<p>(18043-0, LN, "Left Ventricular Ejection Fraction")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (G-C036, SRT, "Measurement Method") = (125209, DCM, "Teichholz")</p>
<p>2D/EDV(Cube) Alias: EDV(Cube)</p>	<p>(18026-5, LN, "Left Ventricular End Diastolic Volume")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (G-C036, SRT, "Measurement Method") = (125206, DCM, "Cube Method")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/ESV(Cube) Alias: ESV(Cube)</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)</p>
<p>2D/EF(Cube) Alias: EF(Cube)</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)</p>
<p>2D/%FS Alias: %FS</p>	<p>(18051-3, LN, “Left Ventricular Fractional Shortening”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode</p>
<p>2D/%IVS Thck Alias: %IVS Thck</p>	<p>(18054-7, LN, “Interventricular Septum % Thickening”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode</p>
<p>2D/%LVPW Thck Alias: %LVPW Thck</p>	<p>(18053-9, LN, “Left Ventricle Posterior Wall % Thickening”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/SV(Teich) Alias: SV(Teich)</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>2D/SI(Teich) Alias: SI(Teich)</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>2D/CO(Teich) Alias: CO(Teich)</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		Method”) = (125209, DCM, “Teichholz”)
<p>2D/CI(Teich) Alias: CI(Teich)</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>2D/SV(Cube) Alias: SV(Cube)</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)</p>
<p>2D/SI(Cube) Alias: SI(Cube)</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT,</p>

		<p>“Measurement Method”) = (125206, DCM, “Cube Method”)</p>
<p>2D/CO(Cube) Alias: CO(Cube)</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)</p>
<p>2D/CI(Cube) Alias: CI(Cube)</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)</p>
<p>2D/LVd Mass Alias: LVd Mass</p>	<p>(18087-7, LN, “Left Ventricle Mass”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32010, SRT, “Diastole”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/LVs Mass Alias: LVs Mass</p>	<p>(18087-7, LN, "Left Ventricle Mass")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole")</p>
<p>2D/LVd Mass/ASE Alias: LVd Mass (ASE)</p>	<p>(18087-7, LN, "Left Ventricle Mass")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (R-4089A, SRT, "Cardiac Cycle Point") = (F-32010, SRT, "Diastole") (G-C036, SRT, "Measurement Method") = (125221, DCM, "Left Ventricle Mass by M-mode")</p>
<p>LVd Mass/Vol(bullet) Alias : LVd Mass/Vol(bullet)</p>	<p>(GEU-106-0156, 99GEMS, "LV mass/volume ratio")</p>	<p>(G-C036, SRT, "Measurement Method") = (DCM, 125228, "Bullet Method") (R-4089A, SRT, "Cardiac Cycle Point") = (SRT, F-32010, "Diastole")</p>
<p>LVd Mass(bullet) Alias : LVd Mass(bullet)</p>	<p>(18087-7, LN, "Left Ventricle Mass")</p>	<p>(G-C036, SRT, "Measurement Method") = (DCM, 125228, "Bullet Method") (R-4089A, SRT, "Cardiac Cycle Point")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		= (SRT, F-32010, "Diastole")
<p>2D/LVs Mass/ASE Alias: LVs Mass (ASE)</p>	<p>(18087-7, LN, "Left Ventricle Mass")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (125221, DCM, "Left Ventricle Mass by M-mode")</p>
<p>2D/LVA diastole Alias: LVA (d)</p>	<p>(G-0375, SRT, "Left Ventricular Diastolic Area")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/LVA systole Alias: LVA (s)</p>	<p>(G-0374, SRT, "Left Ventricular Systolic Area")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode</p>
<p>2D/SAX/LVA diastole Alias: LVA (d)</p>	<p>(G-0375, SRT, "Left Ventricular Diastolic Area")</p>	<p>(G-0373, SRT, "Image Mode") = (G-03A2, SRT, "2D mode") or (R-409E2, SRT, "Doppler Color Flow") depending on scan mode (111031, DCM, "Image View") = (G-0397, SRT, "Parasternal short axis")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/SAX/LVA systole Alias: LVA (s)</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”)</p>
<p>2D/SAX/LVAepi diastole Alias: LVAepi (d)</p>	<p>(59093-5, LN, “Epicardial Area”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32010, SRT, “Diastole”)</p>
<p>2D/SAX/LVAepi systole Alias: LVAepi (s)</p>	<p>(59093-5, LN, “Epicardial Area”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)</p>
<p>2D/SAX/LVAend diastole Alias: LVAend (d)</p>	<p>(59094-3, LN, “Endocardial Area”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or</p>

		(R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32010, SRT, “Diastole”)
2D/SAX/LVAend systole Alias: LVAend (s)	(59094-3, LN, “Endocardial Area”)	(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)
2D/LVOT Area Alias: LVOT Area	(G-038E, SRT, “Cardiovascular Orifice Area”)	(G-C0E3, SRT, “Finding Site”) = (T-32650, SRT, “Left Ventricle Outflow Tract”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode
2D/EDV(A-L) Alias: EDV(A-L)	(18026-5, LN, “Left Ventricular End Diastolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		depending on scan mode (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
2D/EDV(MOD) Alias: EDV(MOD)	(18026-5, LN, “Left Ventricular End Diastolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)
2D/ESV(A-L) Alias: ESV(A-L)	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
2D/ESV(MOD) Alias: ESV(MOD)	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/EF(A-L) Alias: EF(A-L)</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>2D/SV(A-L) Alias: SV(A-L)</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>2D/SI(A-L) Alias: SI(A-L)</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>2D/EF(MOD) Alias: EF(MOD)</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>2D/SV(MOD) Alias: SV(MOD)</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>2D/SI(MOD) Alias: SI(MOD)</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVLd(A4C) Alias: LVLd A4C</p>	<p>(18074-5, LN, “Left Ventricular Major Axis Diastolic Dimension, 4-chamber view”)</p>	
<p>LVAd(A4C) Alias: LVAd A4C</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”)</p>
<p>LVEDV(A-L A4C) Alias: LVEDV A-L A4C</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVEDV Index(A-L A4C) Alias: LVEDV Index A-L A4C</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVEDV(MOD A4C) Alias: LVEDV MOD A4C</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVEDV Index(MOD A4C) Alias: LVEDV Index MOD A4C</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		DCM, “Method of Disks, Single Plane”)
LVLs(A4C) Alias: LVLs A4C	(18075-2, LN, “Left Ventricular Major Axis Systolic Dimension, 4-chamber view”)	
LVAs(A4C) Alias: LVAs A4C	(G-0374, SRT, “Left Ventricular Systolic Area”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”)
LVESV(A-L A4C) Alias: LVESV A-L A4C	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
LVESV Index(A-L A4C) Alias: LVESV Index A-L A4C	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
LVESV(MOD A4C) Alias: LVESV MOD A4C	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)
LVESV Index(MOD A4C) Alias: LVESV Index MOD A4C	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT,

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		“Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)
EF(A-L A4C) Alias: EF A-L A4C	(18043-0, LN, “Left Ventricular Ejection Fraction”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
SV(A-L A4C) Alias: SV A-L A4C	(F-32120, SRT, “Stroke Volume”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
SI(A-L A4C) Alias: SI A-L A4C	(F-00078, SRT, “Stroke Index”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
CO(A-L A4C) Alias: CO A-L A4C	(F-32100, SRT, “Cardiac Output”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
<p>CI(A-L A4C) Alias: CI A-L A4C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>EF(MOD A4C) Alias: LVEF MOD A4C</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>SV(MOD A4C) Alias: SV MOD A4C</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>SI(MOD A4C) Alias: SI MOD A4C</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CO(MOD A4C) Alias: CO MOD A4C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CI(MOD A4C) Alias: CI MOD A4C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVLd(A2C) Alias: LVLd A2C</p>	<p>(18072-9, LN, “Left Ventricular Major Axis Diastolic Dimension, 2-chamber view”)</p>	

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVAd(A2C) Alias: LVAd A2C</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”)</p>
<p>LVEDV(A-L A2C) Alias: LVEDV A-L A2C</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVEDV(MOD A2C) Alias: LVEDV MOD A2C</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVLs(A2C) Alias: LVLs A2C</p>	<p>(18073-7, LN, “Left Ventricular Major Axis Systolic Dimension, 2-chamber view”)</p>	
<p>LVAs(A2C) Alias: LVAs A2C</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”)</p>
<p>LVESV(A-L A2C) Alias: LVESV A-L A2C</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVESV(MOD A2C) Alias: LVESV MOD A2C</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>EF(A-L A2C) Alias: EF A-L A2C</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>SV(A-L A2C) Alias: SV A-L A2C</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>SI(A-L A2C) Alias: SI A-L A2C</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>CO(A-L A2C) Alias: CO A-L A2C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CI(A-L A2C) Alias: CI A-L A2C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>EF(MOD A2C) Alias: LVEF MOD A2C</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>SV(MOD A2C) Alias: SV MOD A2C</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>“Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>SI(MOD A2C) Alias: SI MOD A2C</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CO(MOD A2C) Alias: CO MOD A2C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CI(MOD A2C) Alias: CI MOD A2C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVAd(LAX) Alias: LVAd LAX</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>
<p>LVLd(LAX) Alias: LVLd LAX</p>	<p>(18077-8, LN, “Left Ventricle diastolic major axis”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32011, SRT, “End Diastole”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>
<p>LVLs(LAX) Alias: LVLd LAX</p>	<p>(18076-0, LN, “Left Ventricle systolic major axis”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (109070, DCM, “End Systole”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>
<p>LVEDV(A-L LAX) Alias: LVEDV A-L LAX</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVEDV(MOD LAX) Alias: LVEDV MOD LAX</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVAAs(LAX) Alias: LVAAs LAX</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVESV(A-L LAX) Alias: LVESV A-L LAX</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVESV(MOD LAX) Alias: LVESV MOD LAX</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>EF(A-L LAX) Alias: EF A-L LAX</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>SV(A-L LAX) Alias: SV A-L LAX</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>SI(A-L LAX) Alias: SI A-L LAX</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CO(A-L LAX) Alias: CO A-L LAX</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CI(A-L LAX) Alias: CI A-L LAX</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>EF(MOD LAX) Alias: LVEF MOD LAX</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		DCM, "Method of Disks, Single Plane")
SV(MOD LAX) Alias: SV MOD LAX	(F-32120, SRT, "Stroke Volume")	(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208, DCM, "Method of Disks, Single Plane")
SI(MOD LAX) Alias: SI MOD LAX	(F-00078, SRT, "Stroke Index")	(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208, DCM, "Method of Disks, Single Plane")
CO(MOD LAX) Alias: CO MOD LAX	(F-32100, SRT, "Cardiac Output")	(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208, DCM, "Method of Disks, Single Plane")

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>CI(MOD LAX) Alias: CI MOD LAX</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVEDV(MOD BP) Alias: LVEDV MOD BP</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVEDV Index(MOD BP) Alias: LVEDV Index MOD BP</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVESV(MOD BP) Alias: LVESV MOD BP</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVESV Index(MOD BP) Alias: LVESV Index MOD BP</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>EF(Biplane) Alias: EF Biplane</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>SV(Biplane) Alias: SV Biplane</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>SI(Biplane) Alias: SI Biplane</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>CO(Biplane) Alias: CO Biplane</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>CI(Biplane) Alias: CI Biplane</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVLd(apical) Alias: LVLd apical</p>	<p>(18077-8, LN, “Left Ventricle diastolic major axis”)</p>	
<p>LVLd(apical epi) Alias : LVLd(apical epi)</p>	<p>(18077-8, LN, "Left Ventricle diastolic major axis")</p>	<p>(G-C0E3, SRT, "Finding Site") = (GEU, GEU-106-0157, "Epicardium")</p>
<p>LVLs(apical) Alias: LVLs apical</p>	<p>(18076-0, LN, “Left Ventricle systolic major axis”)</p>	
<p>LVAd(sax MV) Alias: LVAd sax MV</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039A, SRT, “Parasternal short axis at the Mitral Valve level”)</p>
<p>LVA(sax MV) Alias: LVA(sax MV)</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039A, SRT, “Parasternal short axis at the Mitral Valve level”)</p>
<p>LVAd(sax PM) Alias: LVAd sax PM</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039B, SRT, “Parasternal short axis at the Papillary Muscle level”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVAs(sax PM) Alias: LVAs sax PM</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039B, SRT, “Parasternal short axis at the Papillary Muscle level”)</p>
<p>LVAd(sax epi) Alias: LVAd sax EPI</p>	<p>(G-0379, SRT, “Left Ventricle Epicardial Diastolic Area, psax pap view”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode</p>
<p>LVAs(sax epi) Alias: LVAs sax EPI</p>	<p>(59093-5, LN, “Epicardial Area”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-039B, SRT, “Parasternal short axis at the Papillary Muscle level”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)</p>
<p>LVAd(sax) Alias: LVAd sax</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”)</p>
<p>LVAs(sax) Alias: LVAs sax</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”)</p>
<p>EDV(mod sim) Alias: EDV mod sim</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>ESV(mod sim) Alias: ESV mod sim</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	
<p>EF(mod sim) Alias: EF mod sim</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	
<p>SV(mod sim) Alias: SV mod sim</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>SI(mod sim) Alias: SI mod sim</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>CO(mod sim) Alias: CO mod sim</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>CI(mod sim) Alias: CI mod sim</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>EDV(bullet) Alias: EDV bullet</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	
<p>EDV(bullet epi) Alias : EDV(bullet epi)</p>	<p>(18026-5, LN, "Left Ventricular End Diastolic Volume")</p>	<p>(G-C0E3, SRT, "Finding Site") = (GEU, GEU-106-0157, "Epicardium") (G-C036, SRT, "Measurement Method") = (DCM, 125228, "Bullet Method") (R-4089A, SRT, "Cardiac Cycle Point") = (SRT, F-32011, "End Diastole")</p>
<p>ESV(bullet) Alias: ESV bullet</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	