

CERTIFICATE OF CONFORMANCE

PRODUCT NAME: ORTHO CONFIDENCE™ WB

ALBA BIOSCIENCE
PRODUCT CODE: Z484

OCD
PRODUCT CODE: 6842785

LOT / BATCH No: V105193

MANUFACTURING
DATE: 07 January 2011

EXPIRATION DATE: 01 March 2011

DONOR ACCREDITATION: All blood units have been tested by an approved method and found non-reactive (negative) for HBsAg, Anti-HIV 1/2, Anti-HCV and syphilis.

TEST RESULTS:

Vial 1:	A ₁ B R ₁ R ₂ , Kell negative. No antibodies present
Vial 2:	O R ₁ R ₁ , Kell positive containing anti-A, anti-B and anti-c
Vial 3:	A R ₂ R ₂ , Kell negative, Fy ^a negative containing anti-B and anti-Fy ^a
Vial 4:	B rr, Kell negative containing anti-A and anti-D

Hematocrit: 15% ± 2% in all vials

Microbiological
Testing: < 1 cfu/mL in all vials

ORTHO CONFIDENCE™ WB has passed internal quality control tests and inspections according to SOP No. TEST.0356.

ORTHO CONFIDENCE™ WB has passed all internal procedures and meets specifications for Product Code Z484 (OCD Code – 6842785).

Signed: 
Scott Gavin
Quality Assurance Officer

Date: 11 January 2011