



Centrum für Hygiene
und medizinische Produktsicherheit

Certificate

for the following products manufactured by

Getinge Disinfection AB
Ljungadalsgatan 11
P O Box 1505
S- 35115 Växjö

FD1600, FD1605, FD1610, FD1615, Ninjo

bedpan washer-disinfectors for the cleaning/disinfection of
containers for human waste.

After type-testing it is hereby confirmed that the appliance
complies with the requirements of

EN ISO 15883-1 (2009 version)

"General requirements, terms and definitions and tests"


and EN ISO 15883-3 (2009 version)

**"Requirements and tests for washer-disinfectors employing
thermal disinfection for human waste containers"**

Identification number: 99264

Schwerin, November 13th, 2012

Dipl.-Ing (FH) Margrit Werner
General Manager



Prof. Dr. med. H.-P. Werner
Scientific and Technical Director