

## Water ChromAR® HPLC



Suitable for Liquid Chromatography and UV-Spectrophotometry.

**Material No: 6795-25**  
**Batch No: 1501509869**  
**Manufactured Date: 2015/01/15**  
**Expiration Date: 2017/01/20**

### Certificate of Analysis

Test	Specification	Result
Residue after Evaporation	Less, Equal 0.0005 %	0.0003 %
Specific Conductance (25°C)	Less, Equal 2 uS/cm	1.25 uS/cm
Gradient Elution Test		
254 nm	Less, Equal 0.002 AU	Less than 0.002 AU
Ultraviolet Absorbance (1.00-cm cell vs. water):		
210 nm	Less, Equal 0.004	Less than 0.004
254 nm	Less, Equal 0.002	Less than 0.002

For Laboratory, Research or Manufacturing Use.

Storage conditions: Preserve in tight and cool conditions.  
Country of Origin: POLAND  
Packaging Site: Gliwice Mfg & DC  
Manufacturer: GLI

# ISO

Phillipsburg, NJ 9001.2008, 14001.2004  
Paris, KY 9001.2008  
Mexico city, Mexico 9001.2008  
Deventer, The Netherlands 9001.2008, 14001.2004  
Selangor, Malaysia 9001.2008  
Panoli, India 9001.2008  
Gliwice, Poland 9001.2008, 17025.2005

Richard M Siberski  
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600

Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610