



## ~~Flat~~ epidural filter

Box 50/1000

**0.22 µm (anti-bacterial) filter for epidural anaesthesia, for retaining microbial particles and contaminants contained in solutions. Luer-lock fittings. Code 809.22 is radiosterilized.**



|                  | Filter                                    |                         |                            |           |                   |                  |                         |                            |                                     |
|------------------|---|-------------------------|----------------------------|-----------|-------------------|------------------|-------------------------|----------------------------|-------------------------------------|
| Code             | Total filter surface area cm <sup>2</sup> | Dimensions with Luer mm | Dimensions without Luer mm | Weight mm | Priming volume ml | Flow rate ml/min | Pressure resistance bar | Flow rate (gravity) ml/min | Flow rate (pressure 3.6 bar) ml/min |
| <u>808.22</u>    | 4.3                                       | 49                      | 30                         | 4.9       | 0.3               | 12               | 5.2                     | -                          | -                                   |
| 9. <u>809.22</u> | 2.8                                       | -                       | -                          | -         | 0.1               | -                | 5                       | 5                          | 145                                 |

- **Contains Latex:** No
- **Contains DEHP:** No
- **Contains products of animal or biological origin:** No
- **Pyrogen-free:** Yes