

	MESH FOR HERNIA AND INCONTINANCE TECHNICAL FILE	DOCUMENT NO : TF1 REVISION NO : 00 PUB. DATE : 15.08.2010 PAGE NO : 1/1
---	--	--

18 PIRKIMO DALIS

TECHNICAL SPECIFICATIONS

PRODUCT	Cat.No.	TAPE SIZE
 <p style="font-size: small;">Vaginal tape tension free Treatment of Stress Urinary Incontinence</p>	SUPRAPUBIC KIT	10X450mm
	TOT KIT	10X450mm
	HELICAL KIT	10X450mm
	VAGINAL TAPE	10X450mm
	CYSTOCELE KIT	CB80X80X380 mm
	CYSTOCELE FOUR CORNER	CB80X80X380 mm
	RECTOCELE KIT	RB36X280mm



MESH FOR INCONTINANCE:

Raw Material: Polypropylene Monofilament 150 micron thread

Structure : Knitted

Weight : 85 g/m²

Thickness : 420 micron

	MESH FOR HERNIA AND INCONTINANCE TECHNICAL FILE	DOCUMENT NO : TF1 REVISION NO : 00 PUB. DATE : 15.08.2010 PAGE NO : 1/1
---	--	--

TECHNINĖ SPECIFIKACIJA

SUPRO SUI ŠLAPIMO NELAIKymo JUOSTELĖ/VAGINAL TAPE

JUOSTELĖS IŠMATAVIMAI: 10X450 MM (1X45 CM.)

MEDŽIAGA –POLIPROPILENAS

SVORIS – 85 G/M²

STORIS – 420 MIKRONŲ



Implantas įvedamas su:

Daugkartinio naudojimo 2 Transobturatoriniais įvedėjais, S formos, Kodas ATOT-RL:



Daugkartinio naudojimo Retrogaktiniu įvedėju, ¼ apskritimo formos, Kodas AVB:

