

Rabbit anti-PLAP (Placental Alkaline Phosphatase)

Cat. No.: RBK056 (1 ml Concentrate); RBK056-05 (0.5 ml Concentrate);
RBG056 (6 ml Ready-to-use)

Instructions for use

Intended use

This antibody is designed for the specific localisation of Placental Alkaline Phosphatase (PLAP) protein in formalin-fixed, paraffin-embedded tissue sections.

Anti-PLAP antibody is intended for research use only.

Specifications

Specificity:	Placental Alkaline Phosphatase (PLAP)
Immunogen:	Recombinant protein corresponding to human PLAP
Clone:	SP15
Isotype:	Rabbit IgG
Species reactivity:	Human +, others not tested

Summary and Description

PLAP (Placental Alkaline Phosphatase) is an oncofetal antigen with a molecular weight of 60-70 kDa and belongs to a family of membrane-bound phosphatases.

PLAP is detectable in syncytiotrophoblasts of the placenta and in germ cell tumours. Labelling of this enzyme is often used for identification of germ cell tumours and their discrimination from other neoplasias.

Reagent provided

Rabbit monoclonal antibody from tissue culture supernatant in buffer including carrier protein and preservative for stabilisation in the following formats:

Concentrate:	1 ml	(Cat. No. RBK056)
Concentrate:	0.5 ml	(Cat. No. RBK056-05)
Ready-to-use:	6 ml	(Cat. No. RBG056)

Dilution of primary antibody

Dilution of Zytomed Systems' concentrated antibody depends on the detection system used. The final working dilution must always be determined by the user. The elaboration of staining protocol should be done by an experienced specialist. For Zytomed Systems' recommendations see chapter 'Staining procedure'.

Storage and handling

The antibody should be stored at 2-8°C without further dilution.

Dilutions of the concentrated antibody should be done in a suitable antibody dilution buffer (e.g. ZUC025 from Zytomed Systems). The diluted antibody should be stored at 2-8°C after use. The stability of this working solution depends on various parameters and has to be confirmed by appropriate controls.

The antibody provided is suitable for use until the expiry date indicated on the label, if stored at 2-8°C. Do not use product after the expiry date. Positive and negative controls should be run simultaneously with all specimens. If unexpected staining is observed which cannot be explained by variations in laboratory procedures and a problem with the antibody is suspected, contact Zytomed Systems' technical support or your local distributor.

Precautions

Use through qualified personnel only.

Wear protective clothing to avoid contact of reagents and specimens with eye, skin and mucous membranes.

If reagents or specimens come in contact with sensitive area, wash with large amounts of water.

Microbial contamination of the reagent must be avoided, since otherwise non-specific staining may occur.

Sodium azide (NaN_3), used for stabilisation, is not considered hazardous material in the concentration used.

Reaction of sodium azide with lead or copper in drainage pipes can result in the formation of highly explosive metallic azides. Sodium azide should be discarded in a large volume of running water to avoid formation of deposits. A material safety data sheet (MSDS) for the pure substance is available upon request.

Staining procedure for formalin-fixed paraffin sections

Refer to the following table for conditions specifically recommended for this antibody. Also refer to detection system data sheets for guidance on specific staining protocols or other requirements.

Parameters	Zytomed Systems recommendations
*Pre-treatment	Heat Induced Epitope Retrieval (for example in Citrate Buffer pH 6.0 (ZUC028))
*Control tissue	Placenta or seminoma
*Working dilution	1:50-1:100 (for concentrates)
*Incubation time	30 - 60 minutes

Quality control

The recommended positive control tissues for this antibody are placenta or seminoma. We recommend carrying out a positive and a negative control with every staining run. Please refer to the instructions of the detection system for guidance on general quality control procedures.

Troubleshooting

If you observe unusual staining or other deviations from the expected results please read these instructions carefully, refer to the instructions of the detection system for relevant information or contact your local distributor.

Expected results

This antibody stains positive in the cytoplasm and the cytoplasmic membrane of PLAP-positive cells in formalin-fixed, paraffin-embedded tissue sections. Further details about the expression pattern of PLAP can be found in the chapter 'Summary and Description'. Interpretation of the staining results is solely the responsibility of the user. Any experimental result should be confirmed by a medically established diagnostic procedure.

Limitations of the Procedure

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983). Endogenous peroxidase, alkaline phosphatase or biotin may cause non-specific staining depending on the detection system used. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive results with HRP (horse radish peroxidase) detection systems (Omata *et al*, 1980). Inadequate counterstaining and mounting can influence the interpretation of the results.

Zytomed Systems warrants that the product will meet all requirements described from its shipping date until the expiry date is reached, if the product is stored and utilised as recommended. No additional guarantees can be given. Under no circumstances shall Zytomed System be liable for any damages arising out of the use of the reagent provided.

Performance characteristics

Zytomed Systems has conducted studies to evaluate the performance of the antibody for use with a standard detection system. The product has been found to be sensitive and specific to the antigen of interest with minimal or no cross-reactivity.

Bibliography

Paiva J *et al*. Am J Pathol 111:156-165, 1984
 Wick MR *et al*. Hum Pathol 18:946-954, 1987
 Manivel JC *et al*. Am J Surg Pathol 11:21-29, 1987
 Burke AP *et al*. Hum Pathol 19:663-670, 1988

Nadji M and Morales AR Ann N.Y. Acad Sci 420:134-9, 1983
 Omata M *et al*. Am J Clin Pathol 73(5): 626-32, 1980
 Goldsmith JD *et al*. Am J Surg Pathol 26:1627-1633, 2002
 Saad RS *et al*. Appl Immunohistochem Mol Morphol 11:107-112, 2003

November 11, 2013

Rev: A0513

Doc: DBE_RBK056_RBK056-05_RBG056

Explanations of the symbols on the product label:

REF	Bestellnummer Catalog Number Reference du catalogue		Verwendbar bis Use By Utiliser jusque		Gebrauchsanweisung beachten Consult Instructions for use Consulter les instructions d'utilisation
LOT	Chargenbezeichnung Batch Code Code du lot		Lagerungstemperatur Temperature Limitation Limites de température	RUO	Nur für Forschungszwecke For Research Use Only Pour la recherche uniquement
IVD	In vitro Diagnostikum In Vitro Diagnostic Medical Device Dispositif médical de diagnostic in vitro		Achtung Warning Attention		Hersteller / Manufacturer / Fabricant Zytomed Systems GmbH • Anhalterstraße 16 14163 Berlin, Germany • Tel: (+49) 30-804 984 990 www.zytomed-systems.com