

Elecsys Toxo IgM

REF			SYSTEM
09755861190	09755861500	100	cobas e 402 cobas e 801

English

System information

Short name	ACN (application code number)
TOXOIGM	10016

Intended use

Immunoassay for the in vitro qualitative determination of IgM antibodies to *Toxoplasma gondii* in human serum and plasma.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.

Summary

The Elecsys Toxo IgM assay is an immunoassay for the qualitative detection of IgM antibodies to *Toxoplasma gondii* in human serum and plasma. The assay is intended to be used as an aid in the diagnosis of an acute or recent *Toxoplasma gondii* infection in suspected patients and pregnant women.

Toxoplasmosis is a relatively common infection caused by the protozoan parasite *Toxoplasma gondii*.

The infection is mainly acquired by ingestion of food or water contaminated by mature oocysts shed by cats or by undercooked meat containing tissue cysts.^{1,2,3,4} Infection can also be transmitted congenitally if a woman is newly infected during, or just prior to pregnancy, and also via organ transplant or blood transfusion from an infected donor.⁴

Primary, acute infection in healthy individuals is mostly mild or asymptomatic and is followed by life-long latency.^{3,4} Reactivation of a latent *Toxoplasma* infection can occur as a result of immunosuppression (e.g. in organ transplant recipients, patients with cancer receiving chemotherapy or with advanced HIV infection) and can be associated with high morbidity and mortality.^{3,4} Reactivated disease in immunocompromised hosts frequently presents with brain lesions, especially in patients with advanced HIV-related immunosuppression.^{3,4,5}

Primary maternal *Toxoplasma* infection occurring during pregnancy may have significant implications for the fetus as the parasite can be transmitted across the placenta.^{3,6} The majority of infants with congenital infection do not present with clinical symptoms at birth but may develop severe sequelae later in life such as chorioretinitis, intellectual and psychomotor disabilities, visual and hearing impairment.^{3,6,7,8} The fetal infection rate increases with gestational age but, the risk of severe clinical manifestations is higher in the case of early maternal infection.^{3,6,7,8}

Early identification of infection and initiation of appropriate drug therapy in acute infection during pregnancy can prevent congenital damage or ameliorate the severity of clinical manifestations.^{6,7}

The diagnosis of *Toxoplasma* infection is most commonly made by the detection of anti-*Toxoplasma*-specific IgG and IgM antibodies.^{3,4,9}

The presence of IgG antibodies to *Toxoplasma gondii* indicates that infection has occurred but does not distinguish between latent and acute infection.⁹ IgM is typically a marker of acute infection, but residual, long-lasting IgM can be detected months or even years after the primary infection.⁹ To differentiate between a recently acquired and past infection, specimens that are positive for IgM may be tested for IgG avidity. A high avidity index for IgG antibodies indicates that the infection occurred at least 4 months ago.⁹ No clinical interpretation can be deduced from a low avidity result.⁹

The diagnosis of the acute acquired infection during pregnancy is established by a seroconversion or a significant rise in antibody titers (IgG and/or IgM) in serial samples.^{8,9}

Test principle

μ-Capture test principle. Total duration of assay: 18 minutes.

- 1st incubation: 6 μL of sample are automatically prediluted 1:20 with Diluent Universal. *T. gondii*-specific recombinant antigen labeled with a ruthenium complex^{a)} is added. Anti-Toxo IgM antibodies present in the sample react with the ruthenium-labeled *T. gondii*-specific recombinant antigen.

- 2nd incubation: Biotinylated monoclonal h-IgM-specific antibodies and streptavidin-coated microparticles are added. The complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell II M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined automatically by the software by comparing the electrochemiluminescence signal obtained from the reaction product of the sample with the signal of the cutoff value previously obtained by calibration.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Reagents - working solutions

The **cobas e** pack (M, R1, R2) is labeled as TOXOIGM.

- M Streptavidin-coated microparticles, 1 bottle, 12.4 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 *Toxoplasma*-Ag-Ru(bpy)₃²⁺, 1 bottle, 8.2 mL:
Toxoplasma-antigen labeled with ruthenium complex > 1 mg/L; MES^{b)} buffer 50 mmol/L, pH 6.0; preservative.
- R2 Anti-h-IgM-Ab-biotin, 1 bottle, 8.2 mL:
Biotinylated monoclonal anti-h-IgM antibody (mouse) > 500 μg/L; HEPES^{c)} buffer 50 mmol/L, pH 7.2; preservative.

b) MES = 2-morpholino-ethane sulfonic acid

c) HEPES = [4-(2-hydroxyethyl)-piperazine]-ethane sulfonic acid

- TOXOIGM Cal1 Negative calibrator 1, 2 bottles of 0.67 mL:
Human serum, negative for anti-Toxo IgM; preservative.
- TOXOIGM Cal2 Positive calibrator 2, 2 bottles of 0.67 mL:
Human serum, reactive for anti-Toxo IgM; preservative.

Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Warning

- H317 May cause an allergic skin reaction.

Prevention:

- P261 Avoid breathing mist or vapours.
- P272 Contaminated work clothing should not be allowed out of the workplace.
- P280 Wear protective gloves.

Response:

Elecsys Toxo IgM

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods use assays that have been approved or cleared by the FDA or that are in compliance with the legal rules of the European Union (IVDR 2017/746/EU, IVDD 98/79/EC, Annex II, List A). However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{10,11}

All products derived from human blood (TOXOIGM Cal1, TOXOIGM Cal2) are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The serum containing anti-Toxo IgM (TOXOIGM Cal2) was 0.2 micron filtrated.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents (M, R1, R2) in the kit are ready-for-use and are supplied in **cobas e** packs.

Calibrators:

The calibrators are supplied ready-for-use in bottles compatible with the system.

Unless the entire volume is necessary for calibration on the analyzer, transfer aliquots of the ready-for-use calibrators into empty snap-cap bottles (CalSet Vials). Attach the supplied labels to these additional bottles. Store the aliquots at 2-8 °C for later use.

Perform **only one** calibration procedure per aliquot.

All information required for correct operation is available via the **cobas** link.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the **cobas e** pack **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability of the cobas e pack:	
unopened at 2-8 °C	up to the stated expiration date
on the analyzers	16 weeks

Stability of the calibrators:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	16 weeks
on the analyzers at 20-25 °C	use only once

Store calibrators **upright** in order to prevent the calibrator solution from adhering to the snap-cap.

Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K₂-EDTA, K₃-EDTA and Na-citrate plasma.

Plasma tubes containing separating gel can be used.

Criterion: Correct assignment of positive and negative samples. Samples with a COI (cutoff index) > 0.5: $\pm 20\%$ recovery; samples with a COI ≤ 0.5 : ± 0.25 COI recovery.

Stable for 3 days at 20-25 °C, 3 weeks at 2-8 °C, 3 months at -20 °C (± 5 °C). The samples may be frozen 6 times.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Specimens should not be subsequently altered with additives (e.g. biocides, anti-oxidants or substances that could possibly change the pH or ionic strength of the sample) in order to avoid erroneous findings.

Pooled samples and other artificial material may have different effects on different assays and thus may lead to discrepant findings.

Centrifuge samples containing precipitates and thawed samples before performing the assay.

Lyophilized samples, heat-inactivated samples and samples and controls stabilized with azide (up to 1 %) can be used.

Ensure the samples and calibrators are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples and calibrators on the analyzers should be analyzed/measured within 2 hours.

Materials provided

See "Reagents – working solutions" section for reagents.

- 2 x 2 bottle labels

Materials required (but not provided)

- [REF] 04618866190, PreciControl Toxo IgM, 16 x 0.67 mL
 - [REF] 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
 - [REF] 07299001190, Diluent Universal, 36 mL sample diluent
 - General laboratory equipment
 - cobas e** analyzer
- Additional materials for **cobas e** 402 and **cobas e** 801 analyzers:
- [REF] 06908799190, ProCell II M, 2 x 2 L system solution
 - [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
 - [REF] 07485409001, Reservoir Cup, 8 cups to supply ProCell II M and CleanCell M
 - [REF] 06908853190, PreClean II M, 2 x 2 L wash solution
 - [REF] 05694302001, Assay Tip/Assay Cup tray, 6 magazines x 6 magazine stacks x 105 assay tips and 105 assay cups, 3 wasteliners
 - [REF] 07485425001, Liquid Flow Cleaning Cup, 2 adaptor cups to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning Detection Unit
 - [REF] 07485433001, PreWash Liquid Flow Cleaning Cup, 1 adaptor cup to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning PreWash Unit
 - [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use.

Place the cooled (stored at 2-8 °C) **cobas e** pack on the reagent manager. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the **cobas e** pack.

Calibrators:

Place the calibrators in the sample zone.

Read in all the information necessary for calibrating the assay.

Elecsys Toxo IgM

Calibration

Traceability: This method has been standardized against a Roche internal standard. The units have been selected arbitrarily. No internationally accepted standard for Toxo IgM exists.

Calibration frequency: Calibration must be performed once per reagent lot using TOXOIGM Cal1, TOXOIGM Cal2 and fresh reagent (i.e. not more than 24 hours since the **cobas e** pack was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 12 weeks when using the same reagent lot
- after 28 days when using the same **cobas e** pack on the analyzer
- as required: e.g. quality control findings outside the defined limits

Range for electrochemiluminescence signals (counts) for the calibrators:

Negative calibrator (TOXOIGM Cal1): 400-2000

Positive calibrator (TOXOIGM Cal2): 4000-30000

Quality control

Use PreciControl Toxo IgM or other suitable controls for routine quality control procedures.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per **cobas e** pack, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

The analyzer automatically calculates the cutoff based on the measurement of TOXOIGM Cal1 and TOXOIGM Cal2. The result of a sample is given either as reactive or non-reactive as well as in the form of a cutoff index (signal sample/cutoff).

Interpretation of the results

Numeric result	Result message	Interpretation/further steps
COI < 0.8	Non-reactive	Negative for IgM antibodies to <i>T. gondii</i> .
COI ≥ 0.8 to < 1.0	Borderline	Sample should be retested. In case the result is still borderline, a second sample should be collected (e.g. within 2-3 weeks) and testing should be repeated.
COI ≥ 1.0	Reactive	Positive for IgM antibodies to <i>T. gondii</i> .

The magnitude of the measured result above the cutoff is not indicative of the total amount of antibody present in the sample.

The anti-toxoplasma IgM results in a given specimen, as determined by assays from different manufacturers, can vary due to differences in reagents and assay methods.

Limitations - interference

A negative Toxo IgM test result, also in combination with a positive Toxo IgG result, does not completely rule out the possibility of an acute infection with *Toxoplasma gondii*.

- Individuals at the early stage of acute infection may not exhibit detectable amounts of Toxo IgM antibodies. In some of these individuals an indeterminate or low positive result with the Elecsys Toxo IgM assay may be found and indicate an early acute infection. A second sample should be tested e.g. within 2 weeks. The detection of Toxo IgM and/or a significant increase of the Elecsys Toxo IgG antibody titer in the second sample supports the diagnosis of acute Toxoplasma infection.

- In some individuals Toxoplasma IgM-specific antibodies may revert to non-reactive levels within few weeks after infection with *T. gondii*.

A positive Toxo IgM test result in a single sample, also in combination with a negative Toxo IgG result, is not sufficient to prove an acute *Toxoplasma gondii* infection:

- Elevated IgM antibody levels may persist even for years after initial infection.^{12,13} Further tests or a combination of test methods should be done for clarification.^{1,13,14,15}
- In very rare cases false reactive Toxo IgM results may occur. If Toxo IgM results are inconsistent with clinical evidence, additional testing is suggested to confirm the result. For diagnostic purposes, results should be used in conjunction with other data (e.g. results of other tests such as Toxo IgG and Toxo IgG Avidity), and clinical impressions, that might however be unspecific.
- In diagnosis of Toxoplasma infection, particularly during pregnancy if therapeutic decisions depend on the diagnosis, the current local or if not available global medical guidelines provided by professional medical societies are to be followed.

Early treatment may prevent an increase in antibody production. IgG and IgM levels may remain low and can coexist for years.

The results in HIV patients, in patients undergoing immunosuppressive therapy, or in patients with other disorders leading to immune suppression, should be interpreted with caution.

Specimens from neonates, cord blood, pretransplant patients or body fluids other than serum and plasma, such as urine, saliva or amniotic fluid have not been tested.

The effect of the following endogenous substances and pharmaceutical compounds on assay performance was tested. Interferences were tested up to the listed concentrations and no impact on results was observed.

Endogenous substances

Compound	Concentration tested
Bilirubin	≤ 1129 μmol/L or ≤ 66 mg/dL
Hemoglobin	≤ 1.24 mmol/L or ≤ 2000 mg/dL
Intralipid	≤ 2000 mg/dL
Biotin	≤ 287 nmol/L or ≤ 70 ng/mL
Rheumatoid factors ^{d)}	≤ 3720 IU/mL
Albumin	≤ 7.0 g/dL
IgG	≤ 7.0 g/dL
IgA	≤ 1.6 g/dL

d) negative samples only

Criterion: Samples with a COI > 0.5: ≤ ± 20 % recovery; samples with a COI ≤ 0.5: ± 0.25 COI recovery.

As with many μ-capture assays an interference with unspecific IgM is observed. Increasing amounts of unspecific IgM may lead to a decrease in the recovery of positive samples with the Elecsys Toxo IgM assay.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. > 5 mg/day) until at least 8 hours following the last biotin administration.

The high-dose hook effect does not lead to false-negative results in the Elecsys Toxo IgM assay.

Pharmaceutical substances

In vitro tests were performed on 16 commonly used pharmaceuticals. No interference with the assay was found.

In addition, the following special drugs used in toxoplasmosis therapy during pregnancy were tested. No interference with the assay was found.

Special drugs

Drug	Concentration tested mg/L
Spiramycine	≤ 3000
Sulfadiazine	≤ 2500
Pyrimethamine	≤ 500

Elecsys Toxo IgM

Drug	Concentration tested mg/L
Folic acid	≤ 3

In rare cases, interference due to extremely high titers of antibodies to immunological components, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents, samples and controls in a protocol (EP05-A3) of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplicate each for 21 days (n = 84). The following results were obtained:

cobas e 402 and cobas e 801 analyzers					
		Repeatability		Intermediate precision	
Sample	Mean COI	SD COI	CV %	SD COI	CV %
HS ^{e)} , negative	0.158	0.002	1.4	0.003	1.9
HS, weakly positive	1.18	0.011	0.9	0.017	1.5
HS, positive	3.87	0.034	0.9	0.052	1.3
PC ^{f)} Toxo IgM 1	0.169	0.003	1.5	0.003	2.0
PC Toxo IgM 2	1.45	0.014	1.0	0.033	2.3

e) HS = human serum

f) PC = PreciControl

Method comparison

In study 1 the performance of the Elecsys Toxo IgM assay was determined by testing a total of 826 fresh and frozen samples at two sites in comparison to a commercially available Toxoplasma IgM test.

In study 2 the Elecsys Toxo IgM assay was compared to another commercially available Toxoplasma IgM assay by testing 400 fresh and frozen samples. In both studies all specimens with initially discordant results were re-tested. Resolution of repeatedly discordant samples was done by avidity testing. 51 specimens with indeterminate results in one of the assays were excluded from the final calculation of relative sensitivity and specificity.

Relative sensitivity and specificity after resolution

Study	N	Relative sensitivity %	Lower confidence limit %	Relative specificity %	Lower confidence limit %
1	785	95.3 (162/170)	91.7	98.8 (595/602)	97.8
2	390	98.8 (83/84)	94.5	99.7 (294/295)	98.4

Study 1: Of 21 samples which were initially discordant negative with the Elecsys Toxo IgM assay, 11 samples revealed a high avidity test result, 2 samples were found negative with Toxo ISAGA IgM. 7 discordant negative samples revealed a low avidity test result, 1 sample was found positive with Toxo ISAGA IgM. 5 samples which were discordant positive with the Elecsys Toxo IgM assay revealed a high avidity test result, 2 samples were from individuals without Toxoplasma infection.

Study 2: Of 12 samples which were initially discordant negative with the Elecsys Toxo IgM assay, 11 samples revealed a high avidity test result. 1 sample revealed a low avidity test result. 1 sample which was discordant positive with the Elecsys Toxo IgM assay was from an individual without Toxoplasma infection.

Analytical specificity

455 potentially cross reacting samples were tested with the Elecsys Toxo IgM assay and a comparison Toxo IgM assay comprising specimens:

- containing antibodies against HAV, HBV*, HCV, HIV, CMV, EBV*, HSV, VZV, Rubella, Treponema pallidum, Malaria**, Chlamydia and Gonorrhoea
- containing autoantibodies (AMA*, ANA) and elevated titers of rheumatoid factors
- after vaccination against HBV and Influenza

An overall agreement of 98.9 % (446/451) was found in these specimens using the Elecsys Toxo IgM assay and the comparison test. 444 samples were found concordantly negative and 2 samples were found positive. 4 samples were found indeterminate either by the Elecsys Toxo IgM assay or the comparison test and were excluded.

* 1 discordant sample in each of these groups

** 2 discordant samples

Seroconversion panels

In two studies seroconversion samples obtained during pregnancy screening were tested with the Elecsys Toxo IgM assay in comparison to two different commercially available Toxo IgM assays.

In 24 seroconversion panels comprising 83 samples at the first site, the Elecsys Toxo IgM assay detected 64 samples from 66 samples which were found positive using a comparison test. 2 discordant negative sera were follow-up samples, taken more than 8 weeks after infection.

In 29 seroconversion panels (including 92 samples) at the second site, the Elecsys Toxo IgM assay detected 67 samples from 74 samples which were found positive by a second comparison test. 2 discordant negative sera from the very early phase of infection were also negative by another comparison test. In two panels (comprising 3 and 2 serial bleeds from the very early phase of infection) IgM was not detected, however seroconversion could be demonstrated by the Elecsys Toxo IgM assay.

In both panels discordant negative results for several samples were also found by two other commercial Toxoplasma IgM assays.

References

- Montoya JG, Liesenfeld O. Toxoplasmosis. *Lancet* 2004;363:1965-1976.
- Jones JL, Dubey JP. Foodborne toxoplasmosis. *Clin Infect Dis* 2012;55:845-851.
- Halonen SK, Weiss LM. Toxoplasmosis. *Handb Clin Neurol* 2013;114:125-145.
- Jones JL, Parise ME, Fiore AE. Neglected parasitic infections in the United States: toxoplasmosis. *Am J Trop Med Hyg* 2014;90:794-799.
- Luft BJ, Remington JS. Toxoplasmic encephalitis in AIDS. *Clin Infect Dis* 1992;15:211-222.
- Kieffer F, Wallon M. Congenital toxoplasmosis. *Handb Clin Neurol* 2013;112:1099-1101.
- Moncada PA, Montoya JG. Toxoplasmosis in the fetus and newborn: an update on prevalence, diagnosis and treatment. *Expert Rev Anti Infect Ther* 2012;10:815-828.
- Robert-Gangneux F, Dardé ML. Epidemiology of and diagnostic strategies for toxoplasmosis. *Clin Microbiol Rev* 2012;25:264-296.
- Murat JB, Hidalgo HF, Brenier-Pinchart MP, et al. Human toxoplasmosis: which biological diagnostic tests are best suited to which clinical situations? *Expert Rev Anti Infect Ther* 2013;11:943-956.
- Occupational Safety and Health Standards: Bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
- Meek B, van Gool T, Gilis H, et al. Dissecting the IgM antibody response during the acute and latent phase of toxoplasmosis. *Diagn Microbiol Infect Dis* 2001;41:131-137.
- Bobic B, Sibalic D, Djurkovic-Djakovic O. High levels of IgM Antibodies Specific for Toxoplasma gondii in Pregnancy 12 Years after Primary Toxoplasma Infection. *Gynecol Obstet Invest* 1991;31:182-184.
- Remington JS, McLeod R & Desmonts G 2001, Toxoplasmosis, 205-346, in J.S. Remington & J.O. Klein (ed.), *Infectious Diseases of the Fetus and Newborn Infant*, 5th ed. W.B. Saunders, Philadelphia, Pa.

Elecsys Toxo IgM

15 Thulliez P. Maternal and foetal infection: in Toxoplasmosis (eds D.H.M. Joynson, T.G. Wreghitt) Cambridge University Press, 2001:193-213 ISBN 0521 44328 8.

For further information, please refer to the appropriate user guide or operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

The Summary of Safety & Performance Report can be found here:
<https://ec.europa.eu/tools/eudamed>

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see navifyportal.roche.com for definition of symbols used):

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume for reconstitution
	Global Trade Item Number

Rx only For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

COBAS, NAVIFY, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2024, Roche Diagnostics



 Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com
 +800 5505 6606

