

| | | | |
|-------------|-------------|-----|----------------------------|
| REF | | | SYSTEM |
| 09289275190 | 09289275500 | 300 | cobas e 402 cobas e 801 |

English

System information

| | |
|------------|-------------------------------|
| Short name | ACN (application code number) |
| ACOV2S | 10230 |

Intended use

Immunoassay for the in vitro quantitative determination of total antibodies to the Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2) spike (S) protein receptor binding domain (RBD) in human serum and plasma. The test is intended as an aid to assess the adaptive humoral immune response, including neutralizing antibodies, to the SARS-CoV-2 S protein after natural infection with SARS-CoV-2 or in vaccine recipients.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.

Summary

SARS-CoV-2, the causative agent of Coronavirus Disease 2019 (COVID-19), is an enveloped, single-stranded RNA Betacoronavirus. 7 coronaviruses have been identified as agents of human infection, causing disease ranging from mild common cold to severe respiratory failure.¹

SARS-CoV-2 is transmitted primarily from person-to-person through respiratory droplets and aerosols.^{2,3} The incubation period from infection to detectable viral load in the host commonly ranges from 2 to 14 days.^{4,5} Detection of viral load can be associated with the onset of clinical signs and symptoms, although a considerable proportion of individuals remains asymptomatic or mildly symptomatic.^{6,7,8} The interval during which an individual with COVID-19 is infectious has not yet been clearly established, however, transmission from symptomatic, asymptomatic, and pre-symptomatic individuals has been well described.^{9,10,11}

Coronavirus genomes encode 4 main structural proteins: spike (S), envelope (E), membrane (M), and nucleocapsid (N). The S protein is a very large transmembrane protein that assembles into trimers to form the distinctive surface spikes of coronaviruses. Each S monomer consists of an N-terminal S1 subunit and a membrane-proximal S2 subunit. The virus gains entry to the host cell through binding of the S protein to the angiotensin-converting enzyme 2 (ACE2), which is present on the surface of numerous cell types including the alveolar type II cells of the lung and epithelial cells of the oral mucosa.^{12,13} Mechanistically, ACE2 acts as the virus receptor and is engaged by the receptor-binding domain (RBD) on the S1 subunit.^{14,15}

Upon infection with SARS-CoV-2, the host mounts an immune response against the virus, typically including production of specific antibodies against viral antigens. IgM and IgG antibodies against SARS-CoV-2 appear to arise nearly simultaneously in blood.¹⁶ There is significant inter-individual difference in the levels and chronological appearance of antibodies in COVID-19 patients, but median seroconversion has been observed at approximately 2 weeks.^{17,18,19,20,21} Also, titers after a resolved infection show considerable variance from patient to patient.²²

Antibodies against SARS-CoV-2 with strong neutralizing capacity, especially potent if directed against the RBD, have been identified.^{21,23,24} Competition of antibodies with binding of the RBD to ACE2 has been established as a reliable correlate for the assessment of the presence of neutralizing antibodies.²⁵ Numerous vaccines for COVID-19 are in development, many of which focus on eliciting an immune response to the RBD.^{26,27,28}

Serologic assays can play an important role in understanding viral epidemiology in the general population and identifying individuals who are apparently naive and thus presumably susceptible to the virus.

The Elecsys Anti-SARS-CoV-2 S assay uses a recombinant protein representing the RBD of the S antigen in a double-antigen sandwich assay format, which favors the quantitative determination of high affinity antibodies against SARS-CoV-2. Quantification of the antibody response can help to determine the specific antibody titer and aid in longitudinal monitoring of the dynamics of the antibody response in individual patients. The Elecsys Anti-SARS-CoV-2 S assay shows good agreement with direct and surrogate virus neutralization assays.

Test principle

Double-antigen sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: 12 µL of sample, biotinylated SARS-CoV-2 S-RBD-specific recombinant antigen and SARS-CoV-2 S-RBD-specific recombinant antigen labeled with a ruthenium complex^{a)} form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell II M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the **cobas** link.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Reagents - working solutions

The **cobas e** pack is labeled as ACOV2S.

- M Streptavidin-coated microparticles, 1 bottle, 16.0 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 SARS-CoV-2 S-Ag-biotin, 1 bottle, 18.8 mL:
Biotinylated RBD domain of SARS-CoV-2 S as recombinant antigen < 0.4 mg/L; HEPES^{b)} buffer 50 mmol/L, pH 7.4; preservative.
- R2 SARS-CoV-2 S-Ag-Ru(bpy)₃²⁺, 1 bottle, 18.8 mL:
RBD domain of SARS-CoV-2 S as recombinant antigen labeled with ruthenium complex < 0.4 mg/L; HEPES buffer 50 mmol/L, pH 7.4; preservative.

b) HEPES = [4-(2-hydroxyethyl)-piperazine]-ethane sulfonic acid

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Warning

H317 May cause an allergic skin reaction.

Prevention:

P261 Avoid breathing mist or vapours.

P272 Contaminated work clothing should not be allowed out of the workplace.

P280 Wear protective gloves.

Response:

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

For professional use.

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is available via the **cobas** link.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the **cobas e** pack **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

| Stability: | |
|--------------------|----------------------------------|
| unopened at 2-8 °C | up to the stated expiration date |
| on the analyzers | 16 weeks |

Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K₂-EDTA, K₃-EDTA and sodium citrate plasma.

Li-heparin and K₂-EDTA plasma tubes containing separating gel can be used.

Capillary blood collected in serum, Li-heparin plasma or K₂-EDTA plasma sampling tubes.

Criterion: Slope 1.00 ± 0.10 + bias at 0.8 U/mL ± 20 %.

For native samples collected in sodium citrated plasma: Slope 0.84 ± 0.10 .

For capillary blood derived samples in K₂-EDTA sampling tubes: negative samples: < 0.4 U/mL, reactive samples: recovery within 70-130 % of serum value.

Sampling devices containing liquid anticoagulants have a dilution effect resulting in lower values (U/mL) for individual patient specimens. In order to minimize dilution effects it is essential that respective sampling devices are filled completely according to manufacturer's instructions.

Stable for 14 days at 15-25 °C, 14 days at 2-8 °C, 3 months at -20 °C (± 5 °C). The samples may be frozen 3 times.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Specimens should not be subsequently altered with additives (e.g. biocides, anti-oxidants or substances that could possibly change the pH or ionic strength of the sample) in order to avoid erroneous findings.

Centrifuge samples containing precipitates and thawed samples before performing the assay.

Ensure the samples and calibrators are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples and calibrators on the analyzers should be analyzed/measured within 2 hours.

The performance of the Elecsys Anti-SARS-CoV-2 S assay has not been established with cadaveric samples or body fluids other than serum and plasma.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- [REF] 09289291190, CalSet Anti-SARS-CoV-2 S, for 4 x 1.0 mL
 - [REF] 09289313190, PreciControl Anti-SARS-CoV-2 S, 4 x 1.0 mL
 - [REF] 07299001190, Diluent Universal, 36 mL sample diluent
 - General laboratory equipment
 - **cobas e** analyzer
- Additional materials for **cobas e** 402 and **cobas e** 801 analyzers:
- [REF] 06908799190, ProCell II M, 2 x 2 L system solution
 - [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
 - [REF] 07485409001, Reservoir Cup, 8 cups to supply ProCell II M and CleanCell M
 - [REF] 06908853190, PreClean II M, 2 x 2 L wash solution
 - [REF] 05694302001, Assay Tip/Assay Cup tray, 6 magazines x 6 magazine stacks x 105 assay tips and 105 assay cups, 3 wasteliners
 - [REF] 07485425001, Liquid Flow Cleaning Cup, 2 adaptor cups to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning Detection Unit
 - [REF] 07485433001, PreWash Liquid Flow Cleaning Cup, 1 adaptor cup to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning PreWash Unit
 - [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use.

Place the cooled (stored at 2-8 °C) **cobas e** pack under the reagent manager. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the **cobas e** pack.

Calibration

Traceability: This method has been standardized against the internal Roche standard for anti-SARS-CoV-2 S.

Subsequently, it could be shown that the First WHO International Standard for anti-SARS-CoV 2 immunoglobulin (human), NIBSC code: 20/136, behaves identically to the internal Roche standard, with a Pearson correlation coefficient $r = 0.9996$ between Limit of Quantitation and 1000 BAU/mL. Hence, the numeric results in U/mL of the Elecsys Anti-SARS-CoV-2 S assay and BAU/mL are equivalent (e.g. 1 U/mL of the Elecsys Anti-SARS-CoV-2 S assay corresponds to 1 BAU/mL).

Note: Although the defined unit for the Elecsys Anti-SARS-CoV-2 S assay is identical to the binding antibody unit (BAU) defined by the WHO standard, the defined unit for the Elecsys Anti-SARS-CoV-2 S assay must not be used interchangeably with units of other assays. See also the section "Interpretation of results".

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

The predefined master curve is adapted to the analyzer using the relevant CalSet.

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the **cobas e** pack was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 42 days when using the same reagent lot
- after 14 days when using the same **cobas e** pack on the analyzer
- as required: e.g. quality control findings outside the defined limits

Quality control

Use Elecsys PreciControl Anti-SARS-CoV-2 S or other suitable controls for routine quality control procedures.

Elecsys Anti-SARS-CoV-2 S



Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per **cobas e** pack, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

The analyzer automatically calculates the analyte concentration of each sample in U/mL.

Interpretation of the results

| Result | Interpretation |
|-------------|--------------------------------|
| < 0.80 U/mL | Negative for anti-SARS-CoV-2-S |
| ≥ 0.80 U/mL | Positive for anti-SARS-CoV-2-S |

| Duplicate repeat results in cobas e flow ACOV2S DR | Interpretation |
|---|---|
| Both of the duplicate repeat tests < 0.80 U/mL | Negative for anti-SARS-CoV-2-S |
| One or both of the duplicate repeat tests ≥ 0.80 U/mL | Repeatedly reactive, positive for anti-SARS-CoV-2-S |

Note: Due to the diversity of the antibodies, the measured anti-SARS-CoV-2-S value can vary depending on the testing procedure used and the applied standard. Results obtained from a single sample using tests from different manufacturers can therefore differ. If there is a change in the assay procedure used during the monitoring of antibody titers, then the anti-SARS-CoV-2-S values obtained upon changing over to the new procedure must be confirmed by parallel measurements with both methods. For citrated plasma (1 part citrate solution + 9 parts blood), the dilution effect must be taken into account.

Limitations - interference

The effect of the following endogenous substances and pharmaceutical compounds on assay performance was tested. Interferences were tested up to the listed concentrations and no impact on results was observed.

Endogenous substances

| Compound | Concentration tested |
|--------------------|-------------------------------|
| Bilirubin | ≤ 1129 μmol/L or ≤ 66 mg/dL |
| Hemoglobin | ≤ 1000 mg/dL or ≤ 10 g/L |
| Intralipid | ≤ 2000 mg/dL |
| Biotin | ≤ 4912 nmol/L or ≤ 1200 ng/mL |
| Rheumatoid factors | ≤ 1200 IU/mL |
| IgG | ≤ 7.0 g/dL or ≤ 70 g/L |
| IgA | ≤ 1.6 g/dL or ≤ 16 g/L |
| IgM | ≤ 1.0 g/dL or ≤ 10 g/L |

Criterion: For concentrations of 1.0-20 U/mL, the deviation is ≤ 20 %. For concentrations > 20 U/mL, the deviation is ≤ 30 %. For concentrations < 1.0 U/mL, the deviation is ≤ 0.2 U/mL.

No false negative results due to a high-dose hook effect were found with the Elecsys Anti-SARS-CoV-2 S assay but occurrence of high-dose hook effect cannot be completely excluded.

Pharmaceutical substances

In vitro tests were performed on 17 commonly used pharmaceuticals. No interference with the assay was found.

Interference of itraconazole was tested up to the listed concentration and no impact on results was observed.

| Drug | Concentration tested |
|--------------|----------------------|
| Itraconazole | 15 mg/L |

In addition, the following special drugs were tested. No interference with the assay was found.

Antivirals

| Drug | Concentration tested |
|---------------------|----------------------|
| Interferon-alpha-2a | 14400 IU/mL |
| Interferon-alpha-2b | 1000 IU/mL |
| Zanamivir | 0.002 mg/mL |
| Ribavirin | 0.247 mg/mL |
| Oseltamivir | 0.030 mg/mL |
| Peramivir | 0.120 mg/mL |
| Lopinavir | 0.240 mg/mL |
| Ritonavir | 0.160 mg/mL |
| Arbidol | 0.040 mg/mL |
| Remdesivir | 0.040 mg/mL |

Antibiotics

| Drug | Concentration tested |
|--------------|----------------------|
| Levofloxacin | 0.1 mg/mL |
| Azithromycin | 0.1 mg/mL |
| Ceftriaxone | 0.8 mg/mL |
| Meropenem | 1.20 mg/mL |
| Tobramycin | 0.120 mg/mL |

Others

| Drug | Concentration tested |
|-----------------------|----------------------|
| Hydroxychloroquine | 0.16 mg/mL |
| Actemra (Tocilizumab) | 0.128 mg/mL |
| Ocrevus (Ocrelizumab) | 1.5 mg/mL |

Drugs potentially interfering with the ACE2-RBD interface²⁹

| Drug | Concentration tested |
|-------------------|----------------------|
| Risperidon | 0.03 mg/mL |
| Sitagliptin | 0.12 mg/mL |
| Baricitinib | 4.8 μg/mL |
| Silodosin | 9.6 μg/mL |
| Ebastin | 24 μg/mL |
| Indacaterolmaleat | 0.36 μg/mL |
| Regorafenib | 0.192 mg/mL |
| Omalizumab | 0.18 mg/mL |

Drug interferences are measured based on recommendations given in CLSI guidelines EP07 and EP37 and other published literature. Effects of concentrations exceeding these recommendations have not been characterized.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

A negative test result does not completely rule out the possibility of an infection with SARS-CoV-2. Serum or plasma samples from the very early (pre-seroconversion) phase can yield negative findings. Therefore, this test cannot be used to diagnose an acute infection. It

has also been reported that certain patients with confirmed infection do not develop SARS-CoV-2 antibodies.²¹ Furthermore, waning of antibody titers has been reported in some individuals within a range of months after infection, a feature which has also been reported for other coronaviruses.^{30,31,32}

Limits and ranges

Measuring range

0.40-250 U/mL (defined by the Limit of Quantitation and the maximum of the master curve). Values below the Limit of Quantitation are reported as < 0.40 U/mL. Values above the measuring range are reported as > 250 U/mL (or up to 2500 U/mL for 10-fold diluted samples).

Lower limits of measurement

Limit of Blank, Limit of Detection and Limit of Quantitation

Limit of Blank = 0.30 U/mL

Limit of Detection = 0.35 U/mL

Limit of Quantitation = 0.40 U/mL

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A2 requirements.

The Limit of Blank is the 95th percentile value from $n \geq 60$ measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is defined as the lowest amount of analyte in a sample that can be accurately quantified with a CV ≤ 20 %. It has been determined using samples with low concentration of anti-SARS-CoV-2-S.

Dilution

Samples with anti-SARS-CoV-2-S concentrations above the measuring range can be diluted with Diluent Universal. The recommended dilution range is 1:10 up to 1:100.

After dilution by the analyzers, the software automatically takes the dilution into account when calculating the sample concentration.

Note: Antibodies to SARS-CoV-2 are heterogeneous. In some isolated cases, this may lead to non-linear dilution behavior.

An optimized dilution algorithm can be performed automatically (see section "cobas e flows").

cobas e flows

cobas e flows are procedures programmed into the system to enable a fully automated sequence of measurements and the calculation of assay combinations to perform decision algorithms.

The **cobas e flow "ACOV2S D"** is available to automatically perform an initial 1:30 sample dilution. If the result of this measurement is within the extended measuring range (12-7500 U/mL), the result is reported. In case the initial result is found above the extended measuring range, another dilution (1:400) of the sample is automatically carried out to resolve titers up to 100000 U/mL. Results > 100000 U/mL are assigned the result message "above measuring range" with the numeric result set to 100000 U/mL.

In case the initial result is found below the measuring range associated with 1:30 dilution, another measurement is carried out without dilution of the sample and the result is reported.

The **cobas e flow "ACOV2S DR"** is available to measure a sample with the same automated dilution algorithm as in the **cobas e flow "ACOV2S D"**, followed by duplicate repeat measurement for samples with an initially "reactive" result (≥ 0.8 U/mL). Confirmation of the reactive status by one or both of the repeat measurements leads to the main result "repeatedly reactive". Lack of confirmation with both of the repeats leads to the qualitative interpretation of the sample being "non-reactive" reported as the main result. Relevant results of the individual determinations are provided as sub-results in addition to the main result.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents, samples and controls in a protocol (EP05-A3) of the CLSI (Clinical and Laboratory Standards Institute): 1 run per day with 5 replicates of each sample for 5 days ($n = 25$). The following results were obtained:

| cobas e 402 and cobas e 801 analyzers | | | | | |
|---------------------------------------|-----------|---------------|------|------------------------|------|
| Sample | Mean U/mL | Repeatability | | Intermediate precision | |
| | | SD U/mL | CV % | SD U/mL | CV % |
| HSP ^{c)} 1 | 0.483 | 0.014 | 2.9 | 0.014 | 2.9 |
| HSP 2 | 0.826 | 0.015 | 1.9 | 0.015 | 1.9 |
| HSP 3 | 5.69 | 0.121 | 2.1 | 0.136 | 2.4 |
| HSP 4 | 12.0 | 0.159 | 1.3 | 0.191 | 1.6 |
| HSP 5 | 54.8 | 0.743 | 1.4 | 0.770 | 1.4 |
| HSP 6 | 77.3 | 1.23 | 1.6 | 1.54 | 2.0 |
| HSP 7 | 184 | 1.69 | 0.9 | 2.63 | 1.4 |
| PC ^{d)} ACOV2S 1 | < 0.40 | - | - | - | - |
| PC ACOV2S 2 | 10.4 | 0.139 | 1.3 | 0.206 | 2.0 |

c) HSP = human specimen (serum/plasma)

d) PC = PreciControl: PC ACOV2S 1 is free of analyte and therefore consistently resulted below measuring range (< 0.40 U/mL) throughout the experiment, standard deviation and coefficient of variance could therefore not be determined.

Method comparison

A comparison of the Elecsys Anti-SARS-CoV-2 S assay, [REF] 09289275190 (**cobas e 402 analyzer**; y), with the Elecsys Anti-SARS-CoV-2 S assay, [REF] 09289275190 (**cobas e 801 analyzer**; x), gave the following correlations (U/mL):

Number of samples measured: 141

Passing/Bablok³³

$$y = 0.950x - 0.056$$

$$r = 0.998$$

The sample concentrations were between 0.047 and 241 U/mL.

Analytical specificity

1468 samples containing potentially cross-reacting analytes were tested with the Elecsys Anti-SARS-CoV-2 S assay. All samples were obtained before October 2019. No cross-reactivity was found. The resulting overall specificity was 100 %. Results are shown in the following tables:

SARS-CoV-2 related

| Indication | N | Reactive | Specificity % |
|--|-----|----------|---------------|
| MERS CoV (anti-S1 IgG+) | 51 | 0 | 100 |
| Common Coronavirus panel ^{e)} | 151 | 0 | 100 |

e) Pre-pandemic samples which showed serologic reactivity to at least 1 of the endemic Coronaviruses HKU1, NL63, 229E or OC43.

Infectious respiratory diseases

| Indication | N | Reactive | Specificity % |
|---------------------------------|----|----------|---------------|
| Bordetella pertussis | 39 | 0 | 100 |
| Chlamydia pneumoniae | 36 | 0 | 100 |
| Common cold panel ^{f)} | 21 | 0 | 100 |
| Enterovirus | 35 | 0 | 100 |
| Haemophilus influenzae B | 75 | 0 | 100 |
| Influenza A | 40 | 0 | 100 |
| Influenza B | 45 | 0 | 100 |

| Indication | N | Reactive | Specificity % |
|-----------------------------|----|----------|---------------|
| Influenza vaccinees | 25 | 0 | 100 |
| Mycoplasma pneumoniae | 46 | 0 | 100 |
| Parainfluenza | 82 | 0 | 100 |
| Respiratory syncytial virus | 51 | 0 | 100 |

f) 21 potentially cross-reactive samples from individuals with common cold symptoms, collected before October 2019

Other infectious diseases

| Indication | N | Reactive | Specificity % |
|---------------------------------|-----|----------|---------------|
| Adenovirus | 25 | 0 | 100 |
| Borrelia | 6 | 0 | 100 |
| Candida albicans | 13 | 0 | 100 |
| Chlamydia trachomatis | 12 | 0 | 100 |
| CMV acute (IgM+, IgG+) | 86 | 0 | 100 |
| E. coli (anti-E. coli-reactive) | 10 | 0 | 100 |
| EBV acute (IgM+, VCA IgG+) | 106 | 0 | 100 |
| Gonorrhea (tripper) | 5 | 0 | 100 |
| HAV acute (IgM+) | 10 | 0 | 100 |
| HAV late (IgG+) | 15 | 0 | 100 |
| HAV vaccinees | 15 | 0 | 100 |
| HBV acute | 12 | 0 | 100 |
| HBV chronic | 12 | 0 | 100 |
| HBV vaccinees | 15 | 0 | 100 |
| HCV | 50 | 0 | 100 |
| HEV | 12 | 0 | 100 |
| HIV | 10 | 0 | 100 |
| HSV acute (IgM+) | 24 | 0 | 100 |
| HTLV | 6 | 0 | 100 |
| Legionella (IgGAM+) | 7 | 0 | 100 |
| Listeria | 6 | 0 | 100 |
| Measles | 10 | 0 | 100 |
| Mumps | 14 | 0 | 100 |
| Parvovirus B19 | 30 | 0 | 100 |
| Plasmodium falciparum (malaria) | 8 | 0 | 100 |
| Rubella acute (IgM+, IgG+) | 12 | 0 | 100 |
| Toxoplasma gondii (IgM+, IgG+) | 8 | 0 | 100 |
| Treponema pallidum (syphilis) | 62 | 0 | 100 |
| VZV (varicella-zoster virus) | 30 | 0 | 100 |

Autoimmune diseases

| Indication | N | Reactive | Specificity % |
|-------------------------------------|----|----------|---------------|
| AMA (anti-mitochondrial antibodies) | 30 | 0 | 100 |
| ANA (anti-nuclear antibodies) | 17 | 0 | 100 |
| Hemophiliacs | 15 | 0 | 100 |
| RA (rheumatoid arthritis) | 10 | 0 | 100 |

| Indication | N | Reactive | Specificity % |
|------------------------------------|----|----------|---------------|
| SLE (systemic lupus erythematosus) | 10 | 0 | 100 |

Hepatic diseases

| Indication | N | Reactive | Specificity % |
|-------------------------------------|----|----------|---------------|
| Alcohol induced hepatitis/cirrhosis | 13 | 0 | 100 |
| Drug induced hepatitis/cirrhosis | 10 | 0 | 100 |
| Fatty liver | 10 | 0 | 100 |
| Liver cancer | 10 | 0 | 100 |
| Non-viral liver disease | 15 | 0 | 100 |

Clinical specificity

A total of 5991 samples were tested with the Elecsys Anti-SARS-CoV-2 S assay. All samples were obtained before October 2019. 1 false positive sample was detected.

The resulting overall specificity in the internal study was 99.98 %. The 95 % lower confidence limit was 99.91 %.

| Cohort | N | Reactive | Specificity % | 95 % lower confidence limit, % | 95 % upper confidence limit, % |
|-----------------------------|-------------|----------|---------------|--------------------------------|--------------------------------|
| Diagnostic routine (Europe) | 2528 | 0 | 100 | 99.85 | 100 |
| Blood donors (USA) | 2713 | 1 | 99.96 | 99.79 | 100 |
| Blood donors (Africa) | 750 | 0 | 100 | 99.51 | 100 |
| Overall | 5991 | 1 | 99.98 | 99.91 | 100 |

Sensitivity

A total of 1610 samples from 402 symptomatic patients (including 297 samples from 243 hospitalized patients) with a PCR confirmed SARS-CoV-2 infection were tested with the Elecsys Anti-SARS-CoV-2 S assay. 1 or more sequential samples from these patients were collected at various time points after PCR confirmation.

1423 of the tested samples had a sampling date of 14 days or later after diagnosis with PCR. 1406 of these 1423 samples were determined with ≥ 0.8 U/mL in the Elecsys Anti-SARS-CoV-2 S assay and hence considered positive, resulting in a sensitivity of 98.8 % (95 % CI: 98.1-99.3 %) in this sample cohort.

| U/mL | Days after diagnosis with positive PCR | | | | | |
|------------------|--|------|-------|-------|-------|------|
| | 0-6 | 7-13 | 14-20 | 21-27 | 28-34 | > 35 |
| < 0.4 | 4 | 16 | 7 | 3 | 0 | 0 |
| 0.4 - < 0.8 | 0 | 6 | 7 | 0 | 0 | 0 |
| 0.8 - < 1.5 | 2 | 3 | 4 | 1 | 0 | 0 |
| 1.5 - < 2.5 | 0 | 2 | 6 | 2 | 0 | 0 |
| 2.5 - < 5 | 3 | 10 | 9 | 12 | 10 | 40 |
| 5 - < 10 | 1 | 7 | 7 | 15 | 25 | 49 |
| 10 - < 20 | 0 | 11 | 19 | 32 | 25 | 62 |
| 20 - < 50 | 1 | 13 | 19 | 40 | 38 | 183 |
| 50 - < 100 | 3 | 9 | 11 | 34 | 48 | 232 |
| 100 - < 150 | 1 | 4 | 11 | 11 | 21 | 135 |
| 150 - < 200 | 2 | 4 | 2 | 5 | 11 | 95 |
| 200 - \leq 250 | 3 | 8 | 0 | 1 | 5 | 47 |
| > 250 | 15 | 59 | 28 | 20 | 14 | 77 |

| U/mL | Days after diagnosis with positive PCR | | | | | |
|---------------------------|--|------|-------------|-------|-------|------|
| | 0-6 | 7-13 | 14-20 | 21-27 | 28-34 | > 35 |
| ≥ 0.8 | 31 | 130 | 116 | 173 | 197 | 920 |
| Total | 35 | 152 | 130 | 176 | 197 | 920 |
| Sensitivity, % | 88.6 | 85.5 | 89.2 | 98.3 | 100 | 100 |
| CS ^{g)} , % | 86.1 | | 98.8 | | | |
| 95 % CI ^{h)} , % | 80.3 - 90.7 | | 98.1 - 99.3 | | | |

g) CS = Cumulated sensitivity

h) CI = confidence interval

Titer development was investigated with sequential samples from individual patients ranging up to 126 days following a reactive PCR result. None of the samples showed a decline of titer below the reactive range.

Titer development over time for patient samples ranging ≥ 100 days following a reactive PCR result is shown below.

| Donor | D* | D | D | D | D | D | D | D |
|-------|------------|------------|------------|------------|-------------|------------|------------|-------------|
| | U/mL | U/mL | U/mL | U/mL | U/mL | U/mL | U/mL | U/mL |
| 1 | 20 20.4 | 23 22.2 | 27 30.5 | 33 47.4 | 36 51.7 | 61 73.5 | 82 87.7 | 103 114 |
| 2 | 21 36.1 | 24 44.3 | 31 32.4 | 34 48.5 | 37 51.4 | 62 63.1 | 83 73.2 | 104 71.9 |
| 3 | 26 139 | 34 223 | 38 186 | 41 153 | 45 150 | 67 198 | 87 147 | 106 155 |
| 4 | 21 32.3 | 30 95.3 | 33 151 | 36 315 | 41 374 | 62 293 | 83 244 | 107 214 |
| 5 | 30 33.0 | 35 29.5 | 38 31.2 | 42 41.2 | 112 59.9 | | | |
| 6 | 20 7.88 | 30 32.6 | 38 26.6 | 62 39.2 | 71 35.7 | 76 40.3 | 86 36.0 | 107 42.1 |
| 7 | 19 20.7 | 22 40.4 | 25 101 | 29 149 | 39 115 | 48 97.7 | 59 115 | 104 175 |
| 8 | 15 22.1 | 22 14.2 | 30 37.1 | 37 166 | 40 136 | 55 226 | 79 124 | 107 96.9 |
| 9 | 34 181 | 41 148 | 45 148 | 52 165 | 67 152 | 74 154 | 87 125 | 106 119 |
| 10 | 26 4.42 | 29 4.79 | 32 4.83 | 35 5.21 | 42 4.67 | 52 5.95 | 73 7.28 | 103 7.69 |
| 11 | 16 305 | 42 296 | 78 371 | 106 408 | | | | |
| 12 | 28 139 | 31 162 | 40 114 | 44 166 | 47 141 | 62 93.0 | 86 69.5 | 103 59.1 |
| 13 | 24 33.9 | 31 45.6 | 38 63.7 | 46 53.4 | 59 47.4 | 74 41.8 | 92 41.9 | 102 42.8 |
| 14 | 25 79.8 | 28 86.4 | 33 120 | 41 117 | 47 103 | 59 108 | 76 97.1 | 109 105 |
| 15 | 36 255 | 52 165 | 68 126 | 77 94.8 | 92 122 | 96 107 | 106 141 | 126 162 |
| 16 | 30 425 | 44 246 | 51 379 | 58 298 | 73 215 | 85 169 | 90 173 | 104 147 |
| 17 | 29 220 | 32 205 | 40 177 | 48 141 | 55 136 | 76 122 | 95 116 | 101 101 |
| 18 | 31 63.6 | 39 66.9 | 43 53.4 | 53 43.4 | 64 57.3 | 68 48.9 | 92 69.7 | 102 58.8 |
| 19 | 32 94.5 | 46 79.5 | 53 84.3 | 60 71.8 | 68 92.1 | 74 73.6 | 94 78.9 | 102 75.8 |

| Donor | D* | D | D | D | D | D | D | D |
|-------|------------|------------|------------|------------|-------------|------------|------------|-------------|
| | U/mL | U/mL | U/mL | U/mL | U/mL | U/mL | U/mL | U/mL |
| 20 | 38 56.4 | 46 84.2 | 68 104 | 74 106 | 82 114 | 99 141 | 106 152 | 110 146 |
| 21 | 31 9.4 | 38 10.1 | 48 8.7 | 52 9.0 | 57 8.0 | 71 8.8 | 92 10.4 | 106 10.4 |
| 22 | 44 54.3 | 49 51.0 | 61 59.2 | 70 56.9 | 117 99.8 | | | |
| 23 | 35 524 | 42 451 | 55 416 | 74 386 | 81 392 | 109 345 | | |
| 24 | 44 669 | 48 685 | 51 584 | 58 605 | 63 582 | 73 562 | 90 591 | 104 570 |
| 25 | 36 64.0 | 49 83.5 | 56 78.6 | 69 83.9 | 82 100 | 89 103 | 105 121 | |

* Days after initial positive PCR

Detection of antibodies induced by active immunization with vaccines against SARS-CoV-2

Vaccines comprising the RBD of the SARS-CoV-2 Spike protein as an immunogen are expected to induce antibodies in vaccinated individuals that can be quantified with the Elecsys Anti-SARS-CoV-2 S assay. Roche performed internal studies using the Elecsys Anti-SARS-CoV-2 S assay to evaluate the determination of antibody titers induced by the Moderna vaccine Spikevax (mRNA-1273) and the Pfizer-BioNTech vaccine Comirnaty (BNT162b2) following the respectively approved 2-dose vaccination scheme.

Following vaccination with Spikevax or Comirnaty, seroconversion was observed for all participants that had been seronegative at baseline. Titer assessment at the 3 indicated time-points determined rapidly rising titers indicating a strong humoral immune response to vaccination.

Anti-RBD titers induced by Spikevax in seronegative individuals, results are given in U/mL as determined with the Elecsys Anti-SARS-CoV-2 S assay.

| ACOVS2S results in U/mL = BAU/mL | Pre-Vaccination (baseline) | Pre 2nd vaccination (21 days post 1st vaccination) | 14 days post 2nd vaccination* |
|----------------------------------|----------------------------|--|-------------------------------|
| Minimum | < 0.4 | 0.973 | 223 |
| 5. Percentile | n.a.** | 0.978 | 269 |
| 25. Percentile (lower quartile) | n.a.** | 14.15 | 2701 |
| Median | < 0.4 | 95.55 | 6792 |
| 75. Percentile (upper quartile) | n.a.** | 221 | 11044 |
| 95. Percentile | n.a.** | 661 | 17755 |
| Maximum | < 0.4 | 680 | 18169 |
| Inter-quartile range | n.a.** | 679 | 17946 |
| Geometric mean (GM) | < 0.4 | 57.3 | 4559 |
| 95 % CI of GM | n.a.** | (25.4-129) | (2745-7574) |
| Number of values (n) | 24 | 24 | 24 |

Elecsys Anti-SARS-CoV-2 S

Anti-RBD titers induced by Comirnaty in seronegative individuals, results are given in U/mL as determined with the Elecsys Anti-SARS-CoV-2 S assay.

| ACOV2S results in U/mL = BAU/mL | Pre-Vaccination (baseline) | Pre 2nd vaccination (21 days post 1st vaccination) | 14 days post 2nd vaccination* |
|---------------------------------|----------------------------|--|-------------------------------|
| Minimum | < 0.4 | 2.63 | 562 |
| 5. Percentile | n.a.** | 9.28 | 1024 |
| 25. Percentile (lower quartile) | n.a.** | 48.2 | 2064 |
| Median | < 0.4 | 96.8 | 2728 |
| 75. Percentile (upper quartile) | n.a.** | 147 | 3660 |
| 95. Percentile | n.a.** | 790 | 8328 |
| Maximum | < 0.4 | 1070 | 13491 |
| Inter-quartile range | n.a.** | 1067 | 12929 |
| Geometric mean (GM) | < 0.4 | 80.9 | 2833 |
| 95 % CI of GM | n.a.** | (56.2-116) | (2396-3350) |
| Number of values (n) | 31 | 39 | 48 |

* indicates the blood draws from exactly 14 days post 2nd vaccination or the closest available time-point.

** result distribution not assessed, as all pre-vaccination results were non-reactive and below detection limit.

Correlation of assay results to detection of SARS-CoV-2 inhibitory antibodies

534 samples from patients with PCR confirmed SARS-CoV-2 infection covering a range of 6 to 210 days post reactive PCR were used. The samples included cohorts from patients with severe disease requiring hospitalization (n = 122) and mild disease following quarantine at home (n = 412).

Assay results were compared to the result obtained with a commercially available qualitative IVD to detect SARS-CoV-2 inhibitory antibodies (cPass SARS-CoV-2 Neutralization Antibody Detection Kit, GenScript, China). 30 % or higher inhibition of RBD-ACE2 binding in this test indicates the presence of SARS-CoV-2 neutralizing antibodies.³⁴

Application of the medical decision point of the Elecsys Anti-SARS-CoV-2 S assay at 0.8 U/mL (differentiating non-reactive and reactive results) led to the following correlation:

| | | cPass SARS-CoV-2 Surrogate Virus Neutralization Test | | |
|---------------------------------|---------------------------|--|--------------------------------------|-------|
| | | Neutralizing (≥ 30 % inhibition) | Non-neutralizing (< 30 % inhibition) | Total |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 0.8 U/mL (reactive) | 470 | 39 | 509 |
| | < 0.8 U/mL (non-reactive) | 2 | 23 | 25 |
| | Total | 472 | 62 | 534 |

| | Point estimate | 95 % CI ⁱ⁾ |
|----------------------------------|----------------|-----------------------|
| PPA (positive percent agreement) | 99.58 % | 98.48 - 99.95 % |
| NPA (negative percent agreement) | 37.10 % | 25.16 - 50.31 % |
| PPV (positive predictive value) | 92.34 % | 90.87 - 93.59 % |

| | Point estimate | 95 % CI ⁱ⁾ |
|----------------------------------|----------------|-----------------------|
| NPV (negative predictive value)* | n.a. | n.a. |

i) CI = confidence interval

* The analysis focused on PPV only, all included samples were derived from patients with PCR-confirmed SARS-CoV-2 infection. Therefore, NPV is not applicable.

The application of a threshold of 15 U/mL to the results of the Elecsys Anti-SARS-CoV-2 S assay further improved the PPV:

| | | SARS-CoV-2 Surrogate Virus Neutralization Test | | |
|---------------------------------|-----------|--|--------------------------------------|-------|
| | | Neutralizing (≥ 30 % inhibition) | Non-neutralizing (< 30 % inhibition) | Total |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 15 U/mL | 430 | 13 | 443 |
| | < 15 U/mL | 42 | 49 | 91 |
| | Total | 472 | 62 | 534 |

| | Point estimate | 95 % CI |
|------|----------------|-----------------|
| PPA | 91.10 % | 88.16 - 93.51 % |
| NPA | 79.03 % | 66.82 - 88.34 % |
| PPV | 97.07 % | 95.32 - 98.17 % |
| NPV* | n.a. | n.a. |

* The analysis focused on PPV only, all included samples were derived from patients with PCR-confirmed SARS-CoV-2 infection. Therefore, NPV is not applicable.

This study showed that samples with a result of ≥ 15 U/mL had a likelihood of 97.07 % to contain SARS-CoV-2 inhibitory antibodies as determined with the reference assay for detection of inhibitory antibodies.

Correlation of assay results to serum neutralization capacity

The Elecsys Anti-SARS-CoV-2 S assay was compared to a VSV (Vesicular Stomatitis Virus)-based pseudo-neutralization assay.³⁵ The results for 15 clinical samples from individual patients are summarized in the following table:

| | | Pseudo-neutralization assay | | |
|---------------------------------|------------|-----------------------------|---------------|----------|
| | | Positive | Indeterminate | Negative |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 0.8 U/mL | 12 | 0 | 0 |
| | < 0.8 U/mL | 1 | 1 | 1 |

Positive agreement rate: 92.3 %

Calculation of predictive values was not performed due to low sample numbers and resulting lack of statistical significance.

In a randomized, placebo controlled clinical trial on use of Tocilizumab in hospitalized patients with severe COVID-19 pneumonia³⁶, samples were analyzed for virus neutralization capacity by a functional in vitro whole virus neutralization assay (Viroclinics, Netherlands) and antibody titers to the RBD of SARS-CoV-2 S1 (Elecsys Anti-SARS-CoV-2 S assay). The obtained neutralization results were compared to the results of the Elecsys Anti-SARS-CoV-2 S assay. The comparison was performed for the placebo group only to avoid any potential confounding with putative treatment effects.

The sample cohort comprised 206 samples from 111 hospitalized patients with PCR confirmed SARS-CoV-2 infection and severe COVID-19 pneumonia. Up to 3 samples were collected from each patient covering baseline visit (median 11 days from symptom onset, range 2 to 30 days) and 28 or 60 days after enrollment. Presence of 80 % neutralization (NT80) at a sample dilution of 1:8 or higher identified functional virus neutralization in vitro. Comparison to results of the Elecsys Anti-SARS-CoV-2 S assay was realized by application of two different qualitative thresholds, one representing the decision point to identify presence of RBD specific antibodies (0.8 U/mL, medical decision point of the assay to define reactive results) and one based on optimized correlation with detection of inhibitory effects (15 U/mL).

Elecsys Anti-SARS-CoV-2 S



| | | Whole virus NT | | |
|---------------------------------|------------------------------|------------------------------|------------------|-------|
| | | Neutralizing (NT80 ≥ 1:8) | Non-neutralizing | Total |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 0.8 U/mL (reactive) | 187 | 1 | 188 |
| | < 0.8 U/mL (non-reactive) | 6 | 12 | 18 |
| | Total | 193 | 13 | 206 |

| | Point estimate | 95 % CI |
|-------|----------------|---------------|
| PPA | 96.9 % | 93.4 - 98.9 % |
| NPA | 92.3 % | 64.0 - 99.8 % |
| PPV | 99.5 % | 97.1 - 100 % |
| NPV** | n.a. | n.a. |

** Predicting absence of neutralization based on absence of RBD-specific antibodies is not recommended because neutralizing antibodies may also be directed to other proteins besides the RBD. Therefore, NPV is not applicable.

The application of a threshold of 15 U/mL to the results of the Elecsys Anti-SARS-CoV-2 S assay resulted in a PPV of 100 %:

| | | Whole virus NT | | |
|---------------------------------|--------------|------------------------------|------------------|-------|
| | | Neutralizing (NT80 ≥ 1:8) | Non-neutralizing | Total |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 15 U/mL | 164 | 0 | 164 |
| | < 15 U/mL | 29 | 13 | 42 |
| | Total | 193 | 13 | 206 |

| | Point estimate | 95 % CI |
|-------|----------------|---------------|
| PPA | 85.0 % | 79.1 - 89.7 % |
| NPA | 100 % | 75.3 - 100 % |
| PPV | 100 % | 97.8 - 100 % |
| NPV** | n.a. | n.a. |

** Predicting absence of neutralization based on absence of RBD-specific antibodies is not recommended because neutralizing antibodies may also be directed to other proteins besides the RBD. Therefore, NPV is not applicable.

In this study, samples with a result of ≥ 15 U/mL had a likelihood of 100 % to confer in vitro neutralization to SARS-CoV-2 as determined with the applied whole virus NT method.

In a study of the Vitalant Research Institute (CA, USA) investigating COVID-19 convalescent plasma for neutralization capacity, plasma donations from convalescent donors after SARS-CoV-2 infection were analyzed for whole virus neutralizing potential in vitro (BROAD Institute plaque reducing neutralization assay (PRNT), USA). Presence of 50 % neutralization (NT50) at a sample dilution of > 1:20 identified functional virus neutralization in vitro.

390 donations, including cross-sectional and longitudinal sample panels, were analyzed and compared to the obtained Elecsys Anti-SARS-CoV-2 S assay results. Comparison to results of the Elecsys Anti-SARS-CoV-2 S assay was realized by application of two different thresholds, one representing the decision point to identify presence of RBD specific antibodies (0.8 U/mL, medical decision point of the assay to define reactive results) and one based on optimized correlation with detection of inhibitory effects (15 U/mL).

| | | BROAD PRNT | | |
|---------------------------------|------------------------------|-------------------------------|------------------|-------|
| | | Neutralizing (NT50 ≥ 1:20) | Non-neutralizing | Total |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 0.8 U/mL (reactive) | 356 | 4 | 360 |
| | < 0.8 U/mL (non-reactive) | 2 | 28 | 30 |
| | Total | 358 | 32 | 390 |

| | Point estimate | 95 % CI |
|-------|----------------|---------------|
| PPA | 99.4 % | 98.0 - 99.9 % |
| NPA | 87.5 % | 71.0 - 96.5 % |
| PPV | 98.9 % | 97.2 - 99.7 % |
| NPV** | n.a. | n.a. |

** Predicting absence of neutralization based on absence of RBD-specific antibodies is not recommended because neutralizing antibodies may also be directed to other proteins besides the RBD. Therefore, NPV is not applicable.

The application of a threshold of 15 U/mL to the results of the Elecsys Anti-SARS-CoV-2 S assay resulted in a PPV of 100 % (95 % CI: 98.9-100 %):

| | | BROAD PRNT | | |
|---------------------------------|--------------|-------------------------------|------------------|-------|
| | | Neutralizing (NT50 ≥ 1:20) | Non-neutralizing | Total |
| Elecsys Anti-SARS-CoV-2 S assay | ≥ 15 U/mL | 331 | 0 | 331 |
| | < 15 U/mL | 27 | 32 | 59 |
| | Total | 358 | 32 | 390 |

| | Point estimate | 95 % CI |
|-------|----------------|---------------|
| PPA | 92.5 % | 89.2 - 95.0 % |
| NPA | 100 % | 89.1 - 100 % |
| PPV | 100 % | 98.9 - 100 % |
| NPV** | n.a. | n.a. |

** Predicting absence of neutralization based on absence of RBD-specific antibodies is not recommended because neutralizing antibodies may also be directed to other proteins besides the RBD. Therefore, NPV is not applicable.

In this study on convalescent plasma, samples with a result of ≥ 15 U/mL had a likelihood of 100 % (95 % CI: 98.9-100 %) to confer in vitro neutralization to SARS-CoV-2 as determined with the applied PRNT method.

Screening for convalescent plasma for the treatment of hospitalized patients with COVID-19

The Elecsys Anti-SARS-CoV-2 S assay has been included in the emergency use approval (EUA) granted by US FDA for the emergency use of convalescent plasma for the treatment of hospitalized patients with COVID-19.³⁷ The assay has been approved to be used for the purpose of qualifying high titer COVID-19 convalescent plasma in the manufacture of COVID-19 convalescent plasma. US FDA defined ≥ 132 U/mL as the titer cutoff for qualification of high titer COVID-19 convalescent plasma.

This EUA will be effective until the declaration that circumstances exist justifying the authorization of the emergency use of drugs and biological products during the COVID-19 pandemic is terminated under Section 564(b)(2) of the Act or the EUA is revoked under Section 564(g) of the Act. Please refer to the US FDA website for current status.

References

- Ye Z-W, Yuan S, Yuen K-S, et al. Zoonotic origins of human coronaviruses. *Int J Biol Sci* 2020 Mar 15;16(10):1686-1697.
- Transmission of SARS-CoV-2: implications for infection prevention precautions [Internet]. 2020 [cited 2020 Jul 14]. Available from: <https://www.who.int/news-room/commentaries/detail/transmission-of-sars-cov-2-implications-for-infection-prevention-precautions>

- 3 Zhu N, Zhang D, Wang W, et al. A Novel Coronavirus from Patients with Pneumonia in China, 2019. *N Engl J Med* 2020 20;382(8):727-733.
- 4 Chan JF-W, Yuan S, Kok K-H, et al. A familial cluster of pneumonia associated with the 2019 novel coronavirus indicating person-to-person transmission: a study of a family cluster. *Lancet* 2020 15;395(10223):514-523.
- 5 Lauer SA, Grantz KH, Bi Q, et al. The Incubation Period of Coronavirus Disease 2019 (COVID-19) From Publicly Reported Confirmed Cases: Estimation and Application. *Ann Intern Med* 2020 Mar 10.
- 6 Zhou R, Li F, Chen F, et al. Viral dynamics in asymptomatic patients with COVID-19. *International Journal of Infectious Diseases* 2020 Jul 1;96:288-290.
- 7 He X, Lau EHY, Wu P, et al. Temporal dynamics in viral shedding and transmissibility of COVID-19. *Nature Medicine* 2020 May;26(5):672-675.
- 8 Mizumoto K, Kagaya K, Zarebski A, et al. Estimating the asymptomatic proportion of coronavirus disease 2019 (COVID-19) cases on board the Diamond Princess cruise ship, Yokohama, Japan, 2020. *Euro Surveill* 2020 Mar 12;25(10).
- 9 Gao M, Yang L, Chen X, et al. A study on infectivity of asymptomatic SARS-CoV-2 carriers. *Respir Med* 2020 Aug;169:106026.
- 10 Yu P, Zhu J, Zhang Z, et al. A Familial Cluster of Infection Associated With the 2019 Novel Coronavirus Indicating Possible Person-to-Person Transmission During the Incubation Period. *J Infect Dis* 2020 11;221(11):1757-1761.
- 11 Liu Z, Chu R, Gong L, et al. The assessment of transmission efficiency and latent infection period on asymptomatic carriers of SARS-CoV-2 infection. *International Journal of Infectious Diseases* 2020 Jun 13.
- 12 Letko M, Marzi A, Munster V. Functional assessment of cell entry and receptor usage for SARS-CoV-2 and other lineage B betacoronaviruses. *Nat Microbiol* 2020;5(4):562-569.
- 13 Xu H, Zhong L, Deng J, et al. High expression of ACE2 receptor of 2019-nCoV on the epithelial cells of oral mucosa. *Int J Oral Sci* 2020 Feb 24;12(1):1-5.
- 14 Wrapp D, Wang N, Corbett KS, et al. Cryo-EM structure of the 2019-nCoV spike in the prefusion conformation. *Science* 2020 13;367(6483):1260-1263.
- 15 Hoffmann M, Kleine-Weber H, Schroeder S, et al. SARS-CoV-2 Cell Entry Depends on ACE2 and TMPRSS2 and Is Blocked by a Clinically Proven Protease Inhibitor. *Cell* 2020 16;181(2):271-280.e8.
- 16 Centers for Disease Control and Prevention. Interim Guidelines for COVID-19 Antibody Testing [Internet]. Centers for Disease Control and Prevention. 2020 [cited 2020 Jun 4]. Available from: <https://www.cdc.gov/coronavirus/2019-ncov/lab/resources/antibody-tests-guidelines.html>
- 17 Long Q-X, Liu B-Z, Deng H-J, et al. Antibody responses to SARS-CoV-2 in patients with COVID-19. *Nat Med* 2020 Apr 29.
- 18 Lou B, Li T-D, Zheng S-F, et al. Serology characteristics of SARS-CoV-2 infection since exposure and post symptom onset. *Eur Respir J* 2020 May 19;2000763.
- 19 Zhao J, Yuan Q, Wang H, et al. Antibody responses to SARS-CoV-2 in patients of novel coronavirus disease 2019. *Clin Infect Dis* 2020 Mar 28.
- 20 Tuaille E, Bolloré K, Pisoni A, et al. Detection of SARS-CoV-2 antibodies using commercial assays and seroconversion patterns in hospitalized patients. *Journal of Infection* 2020 Jun 3.
- 21 Luchsinger LL, Ransegnola B, Jin D, et al. Serological Analysis of New York City COVID-19 Convalescent Plasma Donors [Internet]. *Infectious Diseases (except HIV/AIDS)*; 2020 Jun [cited 2020 Jul 23]. Available from: <http://medrxiv.org/lookup/doi/10.1101/2020.06.08.20124792>
- 22 Salazar E, Kuchipudi SV, Christensen PA, et al. Relationship between Anti-Spike Protein Antibody Titers and SARS-CoV-2 In Vitro Virus Neutralization in Convalescent Plasma [Internet]. *Immunology*; 2020 Jun [cited 2020 Jun 13]. Available from: <http://biorxiv.org/lookup/doi/10.1101/2020.06.08.138990>
- 23 Klasse P, Moore JP. Antibodies to SARS-CoV-2 and their potential for therapeutic passive immunization. *Giamarellos-Bourboulis EJ, van der Meer JW, editors. eLife*. 2020 Jun 23;9:e57877.
- 24 Premkumar L, Segovia-Chumbez B, Jadi R, Martinez DR, Raut R, Markmann AJ, et al. The receptor-binding domain of the viral spike protein is an immunodominant and highly specific target of antibodies in SARS-CoV-2 patients. *Science Immunology* 2020 Jun 11;5(48).
- 25 Tan CW, Chia WN, Qin X, et al. A SARS-CoV-2 surrogate virus neutralization test based on antibody-mediated blockage of ACE2-spike protein-protein interaction. *Nat. Biotechnol* 2020 doi:10.1038/s41587-020-0631-z.
- 26 Mukherjee R. Global efforts on vaccines for COVID-19: Since, sooner or later, we all will catch the coronavirus. *J Biosci* 2020;45.
- 27 Graham BS. Rapid COVID-19 vaccine development. *Science* 2020 29;368(6494):945-946.
- 28 Hotez PJ, Corry DB, Bottazzi ME. COVID-19 vaccine design: the Janus face of immune enhancement. *Nature Reviews Immunology* 2020 Jun;20(6):347-348.
- 29 Awad IE, Abu-Saleh AAA, Sharma S, et al. High-throughput virtual screening of drug databanks for potential inhibitors of SARS-CoV-2 spike glycoprotein. *J Biomol Struct Dyn* 2022 Mar;40(5):2099-2112. doi: 10.1080/07391102.2020.1835721
- 30 Liu A, Wang W, Zhao X, et al. Disappearance of antibodies to SARS-CoV-2 in a Covid-19 patient after recovery. *Clinical Microbiology and Infection* 2020 Jul 8;0(0).
- 31 Long Q-X, Tang X-J, Shi Q-L, et al. Clinical and immunological assessment of asymptomatic SARS-CoV-2 infections. *Nature Medicine* 2020 Jun 18;1-5.
- 32 Wu L-P, Wang N-C, Chang Y-H, et al. Duration of Antibody Responses after Severe Acute Respiratory Syndrome - Volume 13, Number 10 - October 2007 - Emerging Infectious Diseases journal - CDC. [cited 2020 Jul 16]; Available from: https://wwwnc.cdc.gov/eid/article/13/10/07-0576_article
- 33 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.
- 34 Nanjing GenScript Biotech Co. Ltd. cPass™ SARS-CoV-2 Neutralization Antibody Detection Kit, Instruction for Use. January 2021.
- 35 Meyer B, Torriani G, Yerly S, et al. Validation of a commercially available SARS-CoV-2 serological Immunoassay. medRxiv. 2020. <https://doi.org/10.1101/2020.05.02.20080879>.
- 36 Rosas IO, Bräu N, Waters M, et al. 2020. Tocilizumab in Hospitalized Patients With COVID-19 Pneumonia, medRxiv 2020.08.27.20183442; doi: <https://doi.org/10.1101/2020.08.27.20183442>
- 37 FDA Updates Emergency Use Authorization for COVID-19 Convalescent Plasma to Reflect New Data Convalescent Plasma EUA Letter of Authorization

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

| | |
|---|---|
| CONTENT | Contents of kit |
| SYSTEM | Analyzers/Instruments on which reagents can be used |
| REAGENT | Reagent |
| CALIBRATOR | Calibrator |
|  | Volume for reconstitution |

Elecsys Anti-SARS-CoV-2 S



GTIN

Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2023, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

