

GC Europe N.V.  
European Head Office



**EC DECLARATION OF CONFORMITY**  
**According to Annex VII of MDD 93/42/EEC**

TF08 26 Febr. 2007  
GMDN 16704

**GC EUROPE N.V.**  
**Research Park**  
**Interleuvenlaan 13**  
**B-3001 Leuven**  
**Belgium**

We ensure and declare under our sole responsibility that the product :

**GC FUJI IX GP EXTRA**

to which this declaration relates is in conformity with the following standards or other normative documents :

**Company specifications AB-15-Q-301-432**  
**ISO 9917-1:2003 Dentistry - Water-based cements - Part 1: Powder/liquid acid-base cements**

and meets the provisions of Council Directive 93/42/EEC concerning Medical Devices which apply to it, and is manufactured in accordance with the technical documentation.

This product is Class IIa according to rule 5 of annex IX of the Council Directive.

Leuven, .....  
Date

.....  
J. Van Bragt  
Quality Systems Manager  
On behalf of GC EUROPE N.V.

Interleuvenlaan 13  
B-3001 Leuven Belgium

tel: +32 16 39.80.60

BE 452.407.307  
fax: +32 16.40.26.84