



## Statement

For needle 327350, specification is as follows;

| Ref    | Gauge | Outer diameter (mm) | Inner diameter (mm) |
|--------|-------|---------------------|---------------------|
| 327350 | 17G   | 1.49                | 1.06                |

For Double Lumen Needle 377350 the specification is:

| Ref    | Gauge | Outer Needle        |                     | Inner Needle        |                     |
|--------|-------|---------------------|---------------------|---------------------|---------------------|
|        |       | Outer diameter (mm) | Inner diameter (mm) | Outer diameter (mm) | Inner diameter (mm) |
| 377350 | 17G   | 1.49                | 1.28                | 1.00                | 0.90                |

Quality Control Manager

Signed

Dated 12th day of April 2018