

**Nuo:** RMED Product Support <product.support@radiometer.dk>  
**Išsiųsta:** antradienis 2022 m. lapkritis 22 12:48  
**Kam:** Skaistė Kraujalytė  
**Tema:** RE: SDS CRM:0177437

**Svarba:** Aukštas

Dear Skaistė

You are correct.

SDS is required for chemical substances and mixtures according to Article 31 of REACH Regulation. The items listed below are articles, so we do not create SDSs for them as there are no legal obligations for that.

Kind Regards,  
Bettina

----- Original Message -----

**From:** Skaistė Kraujalytė <konkursai@vitrolab.lt>;  
**Received:** Mon Nov 21 2022 10:53:41 GMT+0100 (Central European Standard Time)  
**To:** TS product.support <product.support@radiometer.dk>; <product.support@radiometer.dk>;  
**Subject:** SDS

**This Message Is From an External Sender**

This message came from outside your organization. Use caution when opening.

Deal All,

Could you write a confirmation that the products listed below don't have Safety Data Sheets (SDS)?

945-603 E1001 Ref electrode  
945-612 E788 pCO2 electrode  
945-613 E799 pO2 electrode  
945-614 pH electrode  
945-615 E722 K electrode  
945-616 E733 Ca electrode  
945-617 E744 Cl electrode  
945-618 E755 Na electrode  
945-619 E7077 Lactate electrode  
945-620 E7066 Glucose electrode

906-020 Clot Catcher

Skaistė Kraujalytė  
Public tender administrator