

Declaration of Conformity

SpofaDental a.s., Markova 238, 506 46 Jičín, Czech Republic

IČO(Company Identification Number): 63999447
DIČ(VAT Taxpayer Number): CZ 63999447

According to DIR 93/42/EEC:2007, Sect. 13 Par. 2 of Law 22/1997 Coll. as amended - of Technical Requirements for Products, according to Sect. 4 of Law 346/2003 Coll. as amended - of Medical Devices and according to the Decree of the Government 336/2004 Coll. determining the Technical Requirements for Medical Devices,

the manufacturer hereby declares that

CARYOSAN - zinc oxide-eugenol dental cement - powder/liquid, class IIa
(according to Annex IX of the DIR 93/42/EEC:2007)

Original package: 60 g (30 g plv. Normal, 30 g plv. Rapid and 25 g liq.)

conforms to the basic requirements and with ordinary use, it is safe for its purpose of use. The manufacturer has implemented measures to guarantee conformity of all medical devices introduced to the market, to their technical documents and to the basic requirements.

To prove conformity, the following regulations and harmonised Czech standards have been applied:

DIR 93/42/EEC:2007 concerning medical devices
DIR 2004/12/EC on packaging and packaging waste
DIR 67/548/EEC The classification, packaging and labelling of dangerous substances
ČSN EN 1641/2010 Dentistry - Medical devices for dentistry - Materials
ČSN EN ISO 10993-1/2010 Biological evaluation of medical devices - Part 1:
Evaluation and testing within a risk management process
ČSN EN ISO 14971/2009 Medical devices - Application of risk management to medical devices
ČSN EN 1041/2009 Information supplied by the manufacturer of medical devices
ČSN EN 980/2009 Symbols for use in the labelling of medical devices
ČSN EN ISO 3107/2011 Dentistry - Zinc oxide/eugenol cements and zinc oxide/non-eugenol cements

To evaluate the basic properties of the product in a specified manner, procedure according to Annex II of the DIR 93/42/EEC:2007 standard was applied, with the participation of the Notified Body TÜV NORD - certificate No. 44 232 09 365388-001

SpofaDental a.s.
Markova 238, CZ - 506 46 Jičín
IČO: 63999447
DIČ: CZ63999447
-17-

27-01-2012

.....
Mgr. Jaroslav Adolf
Quality Assurance and Quality
Control Director

.....
Date