

## EC - DECLARATION OF CONFORMITY

**Name of manufacturer** : gbo Medizintechnik AG

**Address** : Kleiststrasse 6  
D - 64668 Rimbach  
Germany

*We hereby declare under our sole responsibility that the product described below conforms in design and make as well as in the versions delivered to the corresponding safety and protection requirements defined in the applicable EC regulations.*

*Harmonized standards have been used for all conformity checks, national standards have not been applied.*

*Any change to the product design that is not validated by us will render this declaration invalid.*

**Type of product** : Mikrowave Therapy Unit

**Label** : RADIOTHERM 1006 for 230 V Part-No. 020-0-1006-230  
RADIOTHERM 1006 for 115 V Part-No. 020-0-1006-115

<b>Accessories</b>	<b>Description</b>	<b>Part-No.</b>
	Cradle radiator	011-0-0011
	Focus radiator	011-0-0008
	Longitudinal field radiator	011-0-0009
	Circular field radiator	011-0-0012
	Radiator cable	45-39-342EH729

**Corresponding EC regulations:** EC medical devices directive (93/42/EEC)

**Conformity assessment procedure** : Annex II (without section 4) of the directive 93/42/EEC

**Classification** : IIa (according to MDD, appendix IX)

**Name, address and registration no. of the notified body** : TÜV SÜD Product Service GmbH  
Riedlerstrasse 65, 80339 München  
registration no. 0123

**Additional information** : none

**Date** : April-03, 2014

**Name of persons responsible** : Dr. Eberhard Keck

**Title/Function** : CEO

**Signature** : 