

Graduated pipettes, glass

Cat. No. Capacity Tolerance Multipack Unit

Class B

amber stain graduation, DIN 12 696 (0.1 and 0.2 ml: DIN 12 689)

▶ 19 000 00	0.1 : 0.001 ml ("ln")	± 0.002 ml	39 x 24	24
▶ 19 000 10	0.2 : 0.001 ml ("ln")	± 0.003 ml	39 x 24	24
▶ 19 000 30	0.5 : 0.01 ml	± 0.010 ml	39 x 24	24
▶ 19 000 40	1 : 0.01 ml	± 0.010 ml	39 x 24	24
▶ 19 000 50	1 : 0.1 ml	± 0.010 ml	39 x 24	24
▶ 19 000 60	2 : 0.01 ml	± 0.020 ml	39 x 24	24
▶ 19 000 80	2 : 0.1 ml	± 0.020 ml	39 x 24	24
▶ 19 000 90	5 : 0.05 ml	± 0.050 ml	55 x 12	12
▶ 19 001 00	5 : 0.1 ml	± 0.050 ml	55 x 12	12
▶ 19 001 10	10 : 0.1 ml	± 0.10 ml	55 x 12	12
▶ 19 001 20	20 : 0.1 ml	± 0.20 ml	10 x 12	12
▶ 19 001 30	25 : 0.1 ml	± 0.20 ml	10 x 10	10
▶ 19 001 40	50 : 0.2 ml	± 0.40 ml	10 x 5	5

Cat. No. Capacity Tolerance Multipack Unit

Class AS

amber stain graduation, DIN 12 697 (0.1 and 0.2 ml: DIN 12 689)

19 020 00	0.1 : 0.001 ml	± 0.001 ml	39 x 24	24
19 020 10	0.2 : 0.001 ml	± 0.002 ml	39 x 24	24
19 020 30	0.5 : 0.01 ml	± 0.005 ml	39 x 24	24
▶ 19 020 40	1 : 0.01 ml	± 0.006 ml	39 x 24	24
▶ 19 020 50	1 : 0.1 ml	± 0.006 ml	39 x 24	24
▶ 19 020 60	2 : 0.01 ml	± 0.010 ml	39 x 24	24
19 020 70	2 : 0.02 ml	± 0.010 ml	39 x 24	24
19 020 80	2 : 0.1 ml	± 0.010 ml	39 x 24	24
▶ 19 020 90	5 : 0.05 ml	± 0.030 ml	55 x 12	12
▶ 19 021 00	5 : 0.1 ml	± 0.030 ml	55 x 12	12
▶ 19 021 10	10 : 0.1 ml	± 0.050 ml	55 x 12	12
19 021 20	20 : 0.1 ml	± 0.10 ml	10 x 12	12
▶ 19 021 30	25 : 0.1 ml	± 0.10 ml	10 x 10	10
19 021 40	50 : 0.2 ml	± 0.20 ml	10 x 5	5

Class AS, conformity certified, with batch certificate

amber stain graduation, DIN EN ISO 835

(waiting time reduced to 5 seconds)

▶ 19 030 40	1 : 0.01 ml	± 0.006 ml	39 x 24	24
▶ 19 030 70	2 : 0.02 ml	± 0.010 ml	39 x 24	24
▶ 19 030 90	5 : 0.05 ml	± 0.030 ml	55 x 12	12
▶ 19 031 00	5 : 0.1 ml	± 0.030 ml	55 x 12	12
▶ 19 031 10	10 : 0.1 ml	± 0.050 ml	55 x 12	12
19 031 20	20 : 0.1 ml	± 0.10 ml	10 x 12	12
19 031 30	25 : 0.1 ml	± 0.10 ml	10 x 10	10

xx xxx x1 For blue graduation please amend Cat. No. accordingly



Do not mouth pipette! For recommended pipetting aids please see pages 89 and 90.

