



# Shaoxing GANGFENG Hospital Products Co., Ltd.

TEL.: 865758085388  
✉ [medical@mail.sxptt.zj.cn](mailto:medical@mail.sxptt.zj.cn)

Block No.818, Gaobei Indl Zone, Gaobu,  
Shaoxing county, zhejiang 312035, China

## LAB REPORT

INV. NO. 24GFEM245

BATCH NO. : 24GFEM245

GOODS: GAUZE ROLL PACKED / NON -STERILE GAUZE BANDAGES / NON STERILE GAUZE BANDAGES  
PACKED APIECE / STERILE GAUZE BANDAGES PACKED APIECE / STERILE GAUZE SWABS /NON STERILE  
GAUZE BALLS

SAMPLING DATE:JAN 14,2025 REPORT DATE:JAN 14,2025

TEST ITEMS	BP 2000	RESULTS
GAUZE ROLL PACKED:100MX90CM, 20MX90CM,10MX90CM, 5MX90CM,3MX90CM,1MX90CM		
NON -STERILE GAUZE BANDAGES:7CMX5M,10CMX5M,14CMX7M		
NON STERILE GAUZE BANDAGES PACKED APIECE :7CMX5M,10CMX5M,14CMX7M		
STERILE GAUZE BANDAGES PACKED APIECE:7CMX5M,10CMX5M,14CMX7M		
STERILE GAUZE SWABS:40CMX45CM-4P,45CMX70CM-4P		
THREAD COUNT (WARP X WEFT)		26 X 18
YARN COUNT (WARP X WEFT)		40'S X 40'S
WHITENESS		82.3A <sup>0</sup>
WEIGHT(g/sq.m)		23.0g/m <sup>2</sup>
BREAKING LOAD/WARP	≥100 N/10CM	124N
BREAKING LOAD/WEFT	≥60 N/10CM	85N
FLUORESCENCE	NO	NO
LOSS ON DRYING	≤8%	6.3%
ABSORBENCY (SINKING TIME)	≤10 sec.	3sec.
ACIDITY OR ALKALINITY	NO PINK COLOR	NO
STARCH AND DEXTRIN	NO RED, VIOLET OR BLUE COLOR	NO
RESIDUE ON IGNITION	≤0.4%	0.26%
ALCOHOL-SOLUBLE DYES	NO BLUE/GREEN COLOR	NO
WATER-SOLUBLE SUBSTANCES	≤0.5%	0.31%
SURFACE-ACTIVE SUBSTANCES (FROTH HEIGHT)	≤2mm	O.K.
ETHER-SOLUBLE SUBSTANCES	≤0.5%	0.25%

### STERILE PRODUCTS:

### STERILIZATION TEST RESULT:

Coliform Bacteria NO

Aerobic Bacteria NO

Molds and Yeasts NO

Conclusion: Quality conforms to the requirements for medical use.

TESTING LAB: SHAOXING GANGFENG HOSPITAL PRODUCTS CO.,LTD      TECHNICIAN : Ms Jianhong Xi

绍兴港峰医用品有限公司

SHAOXING GANGFENG HOSPITAL PRODUCTS CO., LTD.

七二零