

TRSF2

Tina-quant Transferrin ver.2

Order information

cobas[®]
Specific proteins

REF	CONTENT	Analyzer(s) on which cobas c pack(s) can be used
03015050 122	Tina-quant Transferrin ver.2 (100 tests)	System-ID 07 6567 8 COBAS INTEGRA 400 plus COBAS INTEGRA 800
11355279 216	Calibrator f.a.s. Proteins (5 × 1 mL)	System-ID 07 6557 0
11355279 160	Calibrator f.a.s. Proteins (5 × 1 mL, for USA)	System-ID 07 6557 0
10557897 122	Precinorm Protein (3 × 1 mL)	System-ID 07 9105 9
10557897 160	Precinorm Protein (3 × 1 mL, for USA)	System-ID 07 9105 9
11333127 122	Precipath Protein (3 × 1 mL)	System-ID 07 9106 7
11333127 160	Precipath Protein (3 × 1 mL, for USA)	System-ID 07 9106 7
05117003 190	PreciControl ClinChem Multi 1 (20 × 5 mL)	System-ID 07 7469 3
05947626 190	PreciControl ClinChem Multi 1 (4 × 5 mL)	System-ID 07 7469 3
05947626 160	PreciControl ClinChem Multi 1 (4 × 5 mL, for USA)	System-ID 07 7469 3
05117216 190	PreciControl ClinChem Multi 2 (20 × 5 mL)	System-ID 07 7470 7
05947774 190	PreciControl ClinChem Multi 2 (4 × 5 mL)	System-ID 07 7470 7
05947774 160	PreciControl ClinChem Multi 2 (4 × 5 mL, for USA)	System-ID 07 7470 7
20756350 322	NaCl Diluent 9 % (6 × 22 mL)	System-ID 07 5635 0

English

System information

Test TRSF2, test ID 0-567

Intended use

In vitro test for the quantitative immunological determination of human transferrin in serum and plasma on COBAS INTEGRA systems.

Summary^{1,2,3,4,5}

Transferrin is a glycoprotein with a molecular weight of 79570 daltons. It consists of a polypeptide strand with two N-glycosidically linked oligosaccharide chains and exists in numerous isoforms. The rate of synthesis in the liver can be altered in accordance with the body's iron requirements and iron reserves.

Transferrin is the iron transport protein in serum. In cases of iron deficiency, the degree of transferrin saturation appears to be an extremely sensitive indicator of functional iron depletion. The ferritin levels are depressed when there is a deficiency of storage iron. In sideropenia, an iron deficiency can be excluded if the serum transferrin concentration is low, as in inflammations or - less commonly - in cases of ascorbic acid deficiency. In screening for hereditary hemochromatosis, transferrin saturation provides a better indication of the homozygous genotype than does ferritin. The treatment of anemia with erythropoietin in patients with renal failure is only effective when sufficient depot iron is present. The best monitoring procedure is to determine transferrin saturation during therapy. Transferrin saturation in conjunction with ferritin gives a conclusive prediction of the exclusion of iron overloading in patients with chronic liver disease.

A variety of methods are available for determining transferrin including radial immunodiffusion, nephelometry and turbidimetry. The Roche transferrin assay is based on the immunological agglutination principle.

Test principle

Immunoturbidimetric assay^{6,7,8}

Human transferrin forms a precipitate with a specific antiserum which is determined turbidimetrically at 340 nm.

Reagents - working solutions

- R1** Phosphate buffer 55 mmol/L, pH 7.2; NaCl 25 mmol/L; polyethyleneglycol (PEG) 5 %; preservative.
- SR** Anti-human transferrin antibodies (rabbit), dependent on titer; NaCl 100 mmol/L; preservative.
- R1 is in position B and SR is in position C.

Precautions and warnings

Pay attention to all precautions and warnings listed in Section 1 / Introduction of this Method Manual.

For USA: For prescription use only.

Reagent handling

Ready for use

Storage and stability

Shelf life at 2-8 °C

See expiration date on **cobas c** pack label

COBAS INTEGRA 400 plus system

On-board in use at 10-15 °C

8 weeks

COBAS INTEGRA 800 system

On-board in use at 8 °C

8 weeks

Specimen collection and preparation

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum. Collect serum using standard sampling tubes.

Plasma: Heparin (Li-, Na-, NH₄⁺) plasma.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Samples and controls are automatically prediluted 1:21 (1 + 20) with NaCl solution by the instrument.

Stability:⁹

8 days at 20-25 °C

8 days at 4-8 °C

6 months at -20 °C

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

NaCl Diluent 9 %, Cat. No. 20756350322, system-ID 07 5635 0 for automatic sample and standard serial dilutions. NaCl Diluent 9 % is placed in its predefined rack position and is stable for 4 weeks on-board COBAS INTEGRA 400 plus/800 analyzers.

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Application for serum and plasma

COBAS INTEGRA 400 plus test definition

Measuring mode	Absorbance
Abs. calculation mode	Endpoint
Reaction mode	D-R1-S-SR
Reaction direction	Increase
Wavelength A	340 nm
Calc. first/last	33/50
Typical prozone effect	> 17 g/L (> 214 µmol/L or > 1700 mg/dL)
Antigen excess check	No
Predilution factor	21
Unit	g/L

Pipetting parameters

		Diluent (H ₂ O)
R1	140 µL	
Sample	12.5 µL	5 µL
SR	30 µL	5 µL
Total volume	192.5 µL	

COBAS INTEGRA 800 test definition

Measuring mode	Absorbance
Abs. calculation mode	Endpoint
Reaction mode	D-R1-S-SR
Reaction direction	Increase
Wavelength A	340 nm
Calc. first/last	44/82
Typical prozone effect	> 17 g/L (> 214 µmol/L or > 1700 mg/dL)
Antigen excess check	No
Predilution factor	21
Unit	g/L

Pipetting parameters

		Diluent (H ₂ O)
R1	140 µL	
Sample	12.5 µL	5 µL
SR	30 µL	5 µL
Total volume	192.5 µL	

Calibration

Calibrator	Calibrator f.a.s. Proteins
Calibration dilution ratio	1:10.5, 1:21, 1:42, 1:83, 1:170, and 0 g/L performed automatically by the instrument
Calibration mode	Logit/log 5
Calibration replicate	Duplicate recommended
Calibration interval	Each lot and as required following quality control procedures.

Enter the assigned lot-specific transferrin value of the undiluted calibrator, indicated in the method sheet of the Calibrator f.a.s. Proteins.

Traceability: This method has been standardized with regard to the IFCC/BCR/CAP reference preparation CRM 470 (RPPHS 91/0619) for 14 serum proteins.¹⁰

Quality control

Reference range	Precinorm Protein or PreciControl ClinChem Multi 1
Pathological range	Precipath Protein or PreciControl ClinChem Multi 2
Control interval	24 hours recommended
Control sequence	User defined
Control after calibration	Recommended

For quality control, use control materials as listed in the "Order information" section. In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

COBAS INTEGRA analyzers automatically calculate the analyte concentration of each sample. For more details, please refer to Data Analysis in the Online Help (COBAS INTEGRA 400 plus/800 analyzers).

Conversion factors:	g/L × 100 = mg/dL
	g/L × 12.6 = µmol/L

Limitations - interference

Criterion: Recovery within ± 10 % of initial value.

Icterus:¹¹ No significant interference up to an I index of 60 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 1026 µmol/L or 60 mg/dL).

Hemolysis:¹¹ No significant interference up to an H index of 1000 (approximate hemoglobin concentration: 621 µmol/L or 1000 mg/dL).

Lipemia (Intralipid):¹¹ No significant interference up to an L index of 1500. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Drugs: No interference was found at therapeutic concentrations using common drug panels.^{12,13}

Rheumatoid factors: No significant interference.

Other: Interference by type IgM gammopathy (Waldenström's macroglobulinemia) is recognized by the "High Activity" check. Should samples be flagged "High Act", correct results can be obtained after postdilution.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

ACTION REQUIRED

Special Wash Programming: The use of special wash steps is mandatory when certain test combinations are run together on COBAS INTEGRA analyzers. Refer to the CLEAN Method Sheet for further instructions and for the latest version of the Extra wash cycle list.

Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.

Limits and ranges

Measuring range

0.1-5.2 g/L (1.26-65.5 µmol/L or 10-520 mg/dL) (typical measuring range)

The upper and lower limits of the measuring range depend on the actual calibrator value.

Determine samples having higher concentrations via the rerun function. Dilution of samples via the rerun function is a 1:2 dilution. Results from samples diluted using the rerun function are automatically multiplied by a factor of 2.

Lower limits of measurement

Lower detection limit of the test:
0.1 g/L (1.26 µmol/L or 10 mg/dL)

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying 3 standard deviations above that of a zero sample (zero sample + 3 SD, repeatability, n = 21).

Expected values¹⁴

2.0-3.6 g/L (25.2-45.4 µmol/L or 200-360 mg/dL)

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the COBAS INTEGRA analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using human samples and controls in an internal protocol with repeatability (n = 21) and intermediate precision (3 aliquots per run, 1 run per day, 10 days). The following results were obtained:

Repeatability	Mean	CV
Level 1	1.35 g/L (17.0 µmol/L or 135 mg/dL)	0.86 %
Level 2	3.36 g/L (42.3 µmol/L or 336 mg/dL)	0.77 %

Intermediate precision	Mean	CV
Level 1	1.32 g/L (16.6 µmol/L or 132 mg/dL)	1.8 %
Level 2	3.70 g/L (46.6 µmol/L or 370 mg/dL)	1.9 %

Method comparison

Transferrin values for human serum samples obtained on a COBAS INTEGRA 700 analyzer with the COBAS INTEGRA Tina-quant Transferrin ver.2 reagent (y) were compared to the same reagent on a COBAS INTEGRA 400 analyzer (x) and to those of an alternative manufacturer's automated system (nephelometric determination) (x).

COBAS INTEGRA 400 analyzer

Sample size	(n)	74
Corr. coefficient	(r)	0.998
Lin. regression		$y = 0.97x + 0.04 \text{ g/L}$
Passing/Bablok ¹⁵		$y = 0.97x + 0.03 \text{ g/L}$

The sample concentrations were between 0.35 and 4.82 g/L (4.41 and 60.7 µmol/L or 35 and 482 mg/dL).

Alternative system

Sample size	(n)	50
Corr. coefficient	(r)	0.960
Lin. regression		$y = 1.03x - 0.04 \text{ g/L}$
Passing/Bablok ¹⁵		$y = 1.04x - 0.08 \text{ g/L}$

References

- Wick M, Pinggera W, Lehmann P, eds. Iron Metabolism, Diagnosis and Therapy of Anemias. 5th ed. Vienna/New York: Springer-Verlag 1999.
- Haupt H, Baudner S. Behring Inst Mitt 1990;86:16-19.
- Mancini G, Carbonara AO, Heremans JF. Immunochemical quantitation of antigens by single radial Immunodiffusion Immunochemistry 1965;2:235-243.
- Buffone GJ, Lewis SA, Josefson M, et al. Chemical and immunochemical measurement of total iron-binding capacity compared. Clin Chem 1978;24:1788-1791.

- Kreutzer HJH. An immunological turbidimetric method for serum transferrin determination. J Clin Chem Clin Biochem 1976;14:401-406.
- Lizana J, Hellsing K. Manual immunonephelometric assay of proteins, with use of polymer enhancement. Clin Chem 1974;20:1181-1186.
- Tietz NW. Clinical Guide to Laboratory Tests. 2nd ed. Philadelphia, Pa: WB Saunders Co 1976;278-280.
- Heidelberger M, Kendall FE. A quantitative theory of the precipitin reaction. J Exp Med 1935;62:697-720.
- Guder WG, Narayanan S, Wisser H, et al. List of Analytes; Preanalytical Variables. Brochure in: Samples: From the Patient to the Laboratory. Darmstadt: GIT-Verlag 1996.
- Whicher JT, Ritchie RF, Johnson AM, et al. New international reference preparation for proteins in human serum (RPPHS). Clin Chem 1994;40:934-938.
- Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. Clin Chem 1986;32:470-475.
- Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". Eur J Clin Chem Clin Biochem 1996;34:385-386.
- Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem 2001;38:376-385.
- Schumann G, Dati F. Vorläufige Referenzbereiche für 14 Proteine im Serum (für Erwachsene) nach Standardisierung immunochemischer Methoden unter Bezug auf das internationale Referenzmaterial CRM 470. Lab Med 1995;19:401-403.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT	Contents of kit
→	Volume after reconstitution or mixing
GTIN	Global Trade Item Number

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS INTEGRA, COBAS C, TINA-QUANT, PRECICONTROL, PRECINORM and PRECIPATH are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

Distribution in USA by:
Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336

