

REF	CONTENT	Analyzer(s) on which <b>cobas c</b> pack(s) can be used
07190794 190	Creatine Kinase (200 tests)	System-ID 07 7485 5 Roche/Hitachi <b>cobas c</b> 311, <b>cobas c</b> 501/502
10759350 190	Calibrator f.a.s. (12 x 3 mL)	Code 401
10759350 360	Calibrator f.a.s. (12 x 3 mL, for USA)	Code 401
12149435 122	Precinorm U plus (10 x 3 mL)	Code 300
12149435 160	Precinorm U plus (10 x 3 mL, for USA)	Code 300
12149443 122	Precipath U plus (10 x 3 mL)	Code 301
12149443 160	Precipath U plus (10 x 3 mL, for USA)	Code 301
10171743 122	Precinorm U (20 x 5 mL)	Code 300
10171735 122	Precinorm U (4 x 5 mL)	Code 300
10171778 122	Precipath U (20 x 5 mL)	Code 301
10171760 122	Precipath U (4 x 5 mL)	Code 301
11447378 122	Precinorm CK-MB (4 x 3 mL)	Code 320
04358210 190	Precipath CK-MB (4 x 3 mL, not available in the USA)	Code 356
05117003 190	PreciControl ClinChem Multi 1 (20 x 5 mL)	Code 391
05947626 190	PreciControl ClinChem Multi 1 (4 x 5 mL)	Code 391
05947626 160	PreciControl ClinChem Multi 1 (4 x 5 mL, for USA)	Code 391
05117216 190	PreciControl ClinChem Multi 2 (20 x 5 mL)	Code 392
05947774 190	PreciControl ClinChem Multi 2 (4 x 5 mL)	Code 392
05947774 160	PreciControl ClinChem Multi 2 (4 x 5 mL, for USA)	Code 392
04489357 190	Diluent NaCl 9 % (50 mL)	System-ID 07 6869 3

## English

## System information

For **cobas c** 311/501 analyzers:

**CK2:** ACN 550

For **cobas c** 502 analyzer:

**CK2:** ACN 8550

## Intended use

In vitro test for the quantitative determination of creatine kinase (CK) in human serum and plasma on Roche/Hitachi **cobas c** systems.

## Summary

Creatine kinase (CK) is a dimeric enzyme occurring in four different forms: a mitochondrial isoenzyme and the cytosolic isoenzymes CK-MM (skeletal muscle type), CK-BB (brain type) and CK-MB (myocardial type).<sup>1</sup>



The determination of CK and CK-isoenzyme activities is utilized in the diagnosis and monitoring of myocardial infarction and myopathies such as the progressive Duchenne muscular dystrophy. Following injury to the myocardium, such as occurs with acute myocardial infarction<sup>1</sup>, CK is released from the damaged myocardial cells. In early cases, a rise in the CK-activity can be found just 4 hours after an infarction.<sup>1,2</sup> The CK activity reaches a maximum after 12-24 hours and then falls back to the normal range after 3-4 days.<sup>1,2</sup>

The assay method using creatine phosphate and ADP was first described by Oliver<sup>3</sup>, modified by Rosalki<sup>4</sup> and further improved for optimal test conditions by Szasz et al.<sup>5</sup> CK is rapidly inactivated by oxidation of the sulfhydryl groups in the active center. The enzyme can be reactivated by the addition of acetylcysteine (NAC).<sup>5</sup> Interference by adenylate kinase is prevented by the addition of diadenosine pentaphosphate<sup>6</sup> and AMP.<sup>5,6</sup>

Standardized methods for the determination of CK with activation by NAC were recommended by the German Society for Clinical Chemistry (DGKC)<sup>6</sup> in 1977 and the International Federation of Clinical Chemistry (IFCC)<sup>7</sup> in 1991. In 2002 the IFCC confirmed their recommendation and extended it to 37 °C.<sup>8,9</sup> The method described here is derived from the formulation recommended by the IFCC and was optimized for performance and stability.

## Test principle

UV-test



Equimolar quantities of NADPH and ATP are formed at the same rate. The photometrically measured rate of formation of NADPH is directly proportional to the CK activity.

## Reagents - working solutions

**R1** Imidazole buffer: 123 mmol/L, pH 6.5 (37 °C); EDTA: 2.46 mmol/L; Mg<sup>2+</sup>: 12.3 mmol/L; ADP: 2.46 mmol/L; AMP: 6.14 mmol/L; diadenosine pentaphosphate: 19 μmol/L; NADP<sup>+</sup> (yeast): 2.46 mmol/L; N-acetylcysteine: 24.6 mmol/L; HK (yeast): ≥ 36.7 μkat/L; G6PDH (E. coli): ≥ 23.4 μkat/L; preservative; stabilizers; additives.

**R2** CAPSO\* buffer: 20 mmol/L, pH 8.8 (37 °C); glucose: 120 mmol/L; EDTA: 2.46 mmol/L; creatine phosphate: 184 mmol/L; preservative; stabilizers.

\*CAPSO: 3-(cyclohexylamine)-2-hydroxy-1-propanesulfonic acid

R1 is in position B and R2 is in position C.

## Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

For USA: For prescription use only.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

**CK****Creatine Kinase**

Danger

H360D May damage the unborn child.

**Prevention:**

P201 Obtain special instructions before use.

P202 Do not handle until all safety precautions have been read and understood.

P280 Wear protective gloves/ protective clothing/ eye protection/ face protection.

**Response:**

P308 + P313 IF exposed or concerned: Get medical advice/attention.

**Storage:**

P405 Store locked up.

**Disposal:**

P501 Dispose of contents/container to an approved waste disposal plant.

Product safety labeling primarily follows EU GHS guidance.

Contact phone: all countries: +49-621-7590, USA: 1-800-428-2336

**Reagent handling**

Ready for use

**Storage and stability****CK**

Shelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 8 weeks

**Diluent NaCl 9 %**

Shelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 12 weeks

**Specimen collection and preparation**

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum: Nonhemolyzed serum is the specimen of choice and also recommended by IFCC.

Plasma: Li-heparin, K<sub>2</sub>-, K<sub>3</sub>-EDTA plasma.

Please note: Differences in the degree of hemolysis resulting from the blood sampling procedure used can lead to deviating results in serum and plasma.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Stability in serum:<sup>10</sup> 2 days at 20-25 °C  
7 days at 4-8 °C

4 weeks at -20 °C

Stability in EDTA/heparin plasma:<sup>11</sup> 2 days at 15-25 °C

7 days at 2-8 °C

4 weeks at (-15)-(-25) °C

**Materials provided**

See "Reagents – working solutions" section for reagents.

**Materials required (but not provided)**

See "Order information" section

General laboratory equipment

**Assay**

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

**Application for serum and plasma****cobas c 311 test definition**

Assay type	Rate A		
Reaction time / Assay points	10 / 16-29		
Wavelength (sub/main)	546/340 nm		
Reaction direction	Increase		
Units	U/L (μkat/L)		
Reagent pipetting	Diluent (H <sub>2</sub> O)		
R1	100 μL	–	
R2	20 μL	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	2.8 μL	–	–
Decreased	2.8 μL	15 μL	150 μL
Increased	2.8 μL	–	–

**cobas c 501 test definition**

Assay type	Rate A		
Reaction time / Assay points	10 / 24-44		
Wavelength (sub/main)	546/340 nm		
Reaction direction	Increase		
Units	U/L (μkat/L)		
Reagent pipetting	Diluent (H <sub>2</sub> O)		
R1	100 μL	–	
R2	20 μL	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	2.8 μL	–	–
Decreased	2.8 μL	15 μL	150 μL
Increased	2.8 μL	–	–

**cobas c 502 test definition**

Assay type	Rate A		
Reaction time / Assay points	10 / 24-44		
Wavelength (sub/main)	546/340 nm		
Reaction direction	Increase		
Units	U/L (μkat/L)		

Reagent pipetting		Diluent (H <sub>2</sub> O)	
R1	100 µL	–	
R2	20 µL	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	2.8 µL	–	–
Decreased	2.8 µL	15 µL	150 µL
Increased	5.6 µL	–	–

**Calibration**

Calibrators	S1: H <sub>2</sub> O S2: C.f.a.s.
Calibration mode	Linear
Calibration frequency	2-point calibration • after reagent lot change • as required following quality control procedures

Traceability: This method has been standardized against the IFCC Method for Creatine Kinase.<sup>8</sup>

**Quality control**

For quality control, use control materials as listed in the "Order information" section.

In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

**Calculation**

Roche/Hitachi **cobas c** systems automatically calculate the analyte activity of each sample.

Conversion factor: U/L x 0.0167 = µkat/L

**Limitations - interference**

Criterion: Recovery within ± 10 % of initial value at a creatine kinase activity of 140 U/L (2.34 µkat/L).

Icterus:<sup>12</sup> No significant interference up to an I index of 60 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 1026 µmol/L or 60 mg/dL).

Hemolysis:<sup>12</sup> No significant interference up to an H index of 100 (approximate hemoglobin concentration: 62.1 µmol/L or 100 mg/dL). The level of interference may be variable depending on the exact content of erythrocytes.

Lipemia (Intralipid):<sup>12</sup> No significant interference up to an L index of 1000. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration. Highly lipemic specimens (L index > 1000) may cause high absorbance flagging. Choose diluted sample treatment for automatic rerun.

Drugs: No interference was found at therapeutic concentrations using common drug panels.<sup>13,14</sup>

Cyanokit (Hydroxocobalamin) at therapeutic concentrations interferes with the test.

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.<sup>15</sup>

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

**ACTION REQUIRED**

**Special Wash Programming:** The use of special wash steps is mandatory when certain test combinations are run together on Roche/Hitachi **cobas c** systems. The latest version of the carry-over evasion list can be found with the NaOHD-SMS-SmpCln1+2-SCCS Method Sheets. For further

instructions refer to the operator's manual. **cobas c** 502 analyzer: All special wash programming necessary for avoiding carry-over is available via the **cobas** link, manual input is not required.

**Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.**

**Limits and ranges****Measuring range**

7-2000 U/L (0.12-33.4 µkat/L)

Determine samples having higher activities via the rerun function. Dilution of samples via the rerun function is a 1:11 dilution. Results from samples diluted by the rerun function are automatically multiplied by a factor of 11.

**Lower limits of measurement****Limit of Blank, Limit of Detection and Limit of Quantitation**

Limit of Blank = 7 U/L (0.12 µkat/L)

Limit of Detection = 7 U/L (0.12 µkat/L)

Limit of Quantitation = 7 U/L (0.12 µkat/L)

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A2 requirements.

The Limit of Blank is the 95<sup>th</sup> percentile value from n ≥ 60 measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the limit of blank with a probability of 95 %).

The Limit of Quantitation is the lowest analyte concentration that can be reproducibly measured with a precision of 20 % CV. It has been determined using low concentration creatine kinase samples.

**Expected values**

Reference intervals strongly depend on the patient group and the specific clinical situation.

For healthy people, according to Klein et al.:<sup>16</sup>

CK	U/L	µkat/L
Men	39-308	0.65-5.14
Women	26-192	0.43-3.21

Consensus values:<sup>17</sup>

CK	U/L	µkat/L
Men	< 190	< 3.20
Women	< 170	< 2.85

Consensus values:<sup>17</sup>

CK-MB	U/L	µkat/L
Men/women	< 25	< 0.42

Myocardial infarction: There is a high probability of myocardial damage when the following three conditions are fulfilled:<sup>18</sup>

	U/L	µkat/L
1 CK <sub>men</sub>	> 190	> 3.17
CK <sub>women</sub>	> 167	> 2.79
2 CK-MB	> 24	> 0.40
3 The CK-MB activity accounts for 6-25 % of the total CK-activity.		

According to Tietz:<sup>19</sup>

CK	U/L	µkat/L
Adult males > 19 years	20-200	0.33-3.34
Adult females > 19 years	20-180	0.33-3.01

## Creatine Kinase

The reference values according to Klein et al. are based on the 95th percentile of a group of healthy persons (202 men and 217 women) not involved in high-intensity athletic activities.

In order to ensure high sensitivity in the diagnosis of heart diseases the values given by Tietz are recommended. The loss of diagnostic specificity thereby incurred can be compensated for by additionally determining CK-MB and/or troponin T. When myocardial infarction is suspected the diagnostic strategy proposals in the consensus document of European and American cardiologists should in general be followed.<sup>20</sup>

If despite the suspicion of myocardial infarction the values found remain below the stated limits, a fresh infarction may be involved. In such cases, the determinations should be repeated after 4 hours.

CK varies with physical activity level and race in healthy individuals.<sup>19,21</sup>

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

### Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

### Precision

Repeatability and intermediate precision were determined using human samples and controls in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP5 requirements (2 aliquots per run, 2 runs per day, 21 days). The following results were obtained:

Repeatability	Mean U/L (μkat/L)	SD U/L (μkat/L)	CV %
Human serum 1	18.7 (0.31)	0.6 (0.01)	3.0
Human serum 2	137 (2.29)	0.8 (0.01)	0.6
Human serum 3	477 (7.97)	3.0 (0.05)	0.6
Human serum 4	946 (15.8)	5.3 (0.09)	0.6
Human serum 5	1816 (30.3)	9.4 (0.16)	0.5
PCCC Multi 1*	154 (2.57)	0.9 (0.02)	0.6
PCCC Multi 2	301 (5.02)	1.3 (0.02)	0.4

Intermediate precision	Mean U/L (μkat/L)	SD U/L (μkat/L)	CV %
Human serum 1	18.7 (0.31)	0.6 (0.01)	3.2
Human serum 2	137 (2.29)	1.1 (0.02)	0.8
Human serum 3	477 (7.97)	3.1 (0.05)	0.6
Human serum 4	946 (15.8)	5.8 (0.10)	0.6
Human serum 5	1816 (30.3)	10 (0.17)	0.6
PCCC Multi 1	154 (2.57)	1.7 (0.03)	1.1
PCCC Multi 2	301 (5.02)	2.6 (0.04)	0.9

\*PCCC = PreciControl ClinChem

### Method comparison

Creatine kinase values for human serum and plasma samples obtained on a Roche/Hitachi **cobas c** 501 analyzer (y) were compared with those determined using the CKL reagent on a COBAS INTEGRA 800 analyzer (x).

Sample size (n) = 132

Passing/Bablok<sup>22</sup>

$y = 1.021x + 5.88 \text{ U/L}$

$r = 0.980$

Linear regression

$y = 1.006x + 13.6 \text{ U/L}$

$r = 0.999$

The sample activities were between 7.59 and 1946 U/L (0.13 and 32.5 μkat/L).

### References




- 1 Thomas L, ed. Labor und Diagnose, 8th ed. Bd 1:TH-Books Verlagsgesellschaft 2012.
- 2 Stein W. Laboratory Diagnosis of Acute Myocardial Infarction. Darmstadt: GIT Verlag 1988;34-37.
- 3 Oliver IT. A spectrophotometric method for the determination of creatine phosphokinase and myokinase. Biochem J 1955;61:116-122.
- 4 Rosalki SB. An improved procedure for serum creatine phosphokinase determination. J Lab Clin Med 1967;69:696-705.
- 5 Szasz G, Gruber W, Bernt E. Creatine kinase in serum: 1. Determination of optimum reaction conditions. Clin Chem 1976;22(5):650-656.
- 6 Standard method for the determination of creatine kinase activity. J Clin Chem Clin Biochem 1977;15:249-260.
- 7 Hørdler M, Elser RC, Gerhardt M, et al. Approved Recommendation on IFCC Methods for the Measurement of Catalytic Concentration of Enzymes. Part 7. IFCC Method for Creatine Kinase. Eur J Clin Chem Clin Biochem 1991;29:435-456.
- 8 Schumann G, Bonora R, Ceriotti F, et al. IFCC Primary Reference Procedures for the Measurement of Catalytic Activity Concentrations of Enzymes at 37 °C – Part 2. Reference Procedure for the Measurement of Catalytic Concentration of Creatine Kinase. Clin Chem Lab Med 2002;40(6):635-642.
- 9 Klauke R, Schmidt E, Lorentz K. Recommendations for carrying out standard ECCLS procedures (1988) for the catalytic concentrations of creatine kinase, aspartate aminotransferase, alanine aminotransferase and γ-glutamyltransferase at 37 °C. Eur J Clin Chem Clin Biochem 1993;31:901-909.
- 10 Guder WG, Narayanan S, Wisser H, et al. List of Analytes; Preanalytical Variables. Brochure in: Samples: From the Patient to the Laboratory. Darmstadt: GIT-Verlag 1996.
- 11 Data on file at Roche Diagnostics.
- 12 Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. Clin Chem 1986;32:470-475.
- 13 Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". Eur J Clin Chem Clin Biochem 1996;34:385-386.
- 14 Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem 2001;38:376-385.
- 15 Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med 2007;45(9):1240-1243.
- 16 Klein G, Berger A, Bertholf R, et al. Abstract: Multicenter Evaluation of Liquid Reagents for CK, CK-MB and LDH with Determination of Reference Intervals on Hitachi Systems. Clin Chem 2001;47:Suppl. A30.
- 17 Thomas L, Müller M, Schumann G, et al. Consensus of DGKL and VDGH for interim reference intervals on enzymes in serum. J Lab Med 2005; 29(5):301-308.
- 18 Stein W. Strategie der klinischen-chemischen Diagnostik des frischen Myokardinfarktes. Med Welt 1985;36:572-577.
- 19 Wu AHB, editor. Tietz Clinical Guide to Laboratory Tests, 4th edition. St. Louis (MO): Saunders Elsevier 2006;306-307.
- 20 Myocardial Infarction Redefined - A Consensus Document of the Joint European Society of Cardiology/ American College of Cardiology Committee for the Redefinition of Myocardial Infarction. Eur Heart J 2000;21:1502-1513.
- 21 Black HR, Quallich H, Gareleck CB. Racial differences in serum creatine kinase levels. Am J Med 1986;81:479-487.
- 22 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

**CK****Creatine Kinase****cobas®**

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

**Symbols**

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard:

	Contents of kit
	Volume after reconstitution or mixing
	Global Trade Item Number

**FOR US CUSTOMERS ONLY: LIMITED WARRANTY**

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C, PRECICONTROL, PRECINORM and PRECIPATH are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2016, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)



Distribution in USA by:  
Roche Diagnostics, Indianapolis, IN  
US Customer Technical Support 1-800-428-2336