

TPLA Control Set

SEKISUI

REF 04955188 190

→ 2 x 1 mL Level A

→ 2 x 1 mL Level B

English

System information

For use on Roche/Hitachi MODULAR and **cobas c** analyzers the control code is 119 for Level A and 120 for Level B.

Please note: For technical reasons it is necessary to specify dummy units (mg/dL) on Roche/Hitachi analyzers. The values can be converted to the correct units via the host computer.

Intended use

The TPLA Control Set is used for quality control of the Sekisui "Mediace TPLA" immunoturbidimetric assay (Cat.No. 04955137).

Summary

The TPLA Control Set contains 2 x 2 lyophilized controls based on human serum.

The controls are used for monitoring accuracy and precision.

Reagents – working solutions

Reactive components in the lyophilizate:

Human serum and bovine serum albumin (BSA) with chemical additives and material of biological origin as specified. The origin of the biological additive is as follows:

Analyte	Origin
Anti-Treponema pallidum antibodies	human

Non-reactive components in the lyophilizate:

BSA, gelatin, stabilizers and preservatives

The concentrations of the control components are lot-specific. The exact target values are given on the labels of each vial and in the electronically available or enclosed value sheets.

The exact ranges are given in the electronically available or enclosed value sheets.

The values are also encoded in the enclosed control barcode sheets for Roche/Hitachi MODULAR analyzers.

For the **cobas c** analyzers (except for the **cobas c** 111 analyzer) the values are encoded in electronic files sent via the **cobas** link to the analyzers.

Target values and ranges

The target values were determined using the Sekisui Mediacce TPLA immunoturbidimetric method. Results should be within the defined ranges. Each laboratory should establish corrective measures to be taken if values fall outside the range.

A clinically insignificant difference may be seen between the value(s) listed on the value sheet and the value(s) obtained from the instrument readable data. This is caused by:

- the rounding of value(s) during conversion from the unit in the instrument readable data to the unit that is being used.
- the calculation of the ranges by the analyzer using the percentage values for the ranges encoded in the barcodes.

The traceability of the target value is given in the respective Method Sheets for the system reagents to be used in combination with the recommended calibrator.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level

of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{1,2}

Handling

Carefully open 1 bottle avoiding the loss of lyophilizate, and pipette in exactly 1 mL of distilled/deionized water. Carefully close the bottle and dissolve the contents completely by occasional gentle swirling within 30 minutes. Avoid the formation of foam.

The enclosed barcoded labels are intended exclusively for the Roche/Hitachi MODULAR analyzers and **cobas c** systems to identify the control. Attach the barcoded labels to the tubes carrying the sample cups containing the control material.

Storage and stability³

Store at 2-8 °C.

Criterion for the stability data stated:

Recovery within ± 20 % of initial value.

Stability:

Stability of the lyophilized control:	up to the stated expiration date at 2-8 °C
---------------------------------------	--

Stability of the reconstituted control:	24 hours at 15-25 °C
	7 days at 2-8 °C

	2 weeks at (-15)-(-25) °C
--	---------------------------

Store controls tightly capped when not in use.

Materials provided

- See "Reagents – working solutions" section
- Barcoded labels

Materials required (but not provided)

- System reagents and clinical chemistry analyzers
- General laboratory equipment

Assay

Dispense the required volume into a sample cup and analyze in the same way as patient samples.

The controls should be run daily in parallel with patient samples and after every calibration. Control intervals must be adapted to individual laboratory's requirements.

Follow the applicable government regulations and local guidelines for quality control.

References

- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
- Data on file at Sekisui.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT



Contents of kit

Volume after reconstitution or mixing

COBAS, COBAS C and MODULAR are trademarks of Roche. MEDIACE and TPLA are trademarks of Sekisui. All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics

TPLA Control Set

SEKISUI



Sekisui Medical Co. Ltd.
13-5, Nihombashi 3-chome, Chuo-ku
Tokyo, 103-0027, Japan



Medical Device Safety Service GmbH (MDSS)
Schiffgraben 41
D-30175 Hannover, Germany

European Distributor:
Roche Diagnostics GmbH
Sandhofer Strasse 116
D-68305 Mannheim, Germany

