

Cyclosporine

Cyclosporine

cobas®

REF		SYSTEM
05889014 190	100	Elecsys 2010 MODULAR ANALYTICS E170 cobas e 411 cobas e 601 cobas e 602

English

Intended use

Immunoassay for the in vitro quantitative determination of cyclosporine in human whole blood. The assay is used as an aid in the management of heart, liver, kidney, lung and bone marrow transplant patients receiving cyclosporine therapy.

The **electrochemiluminescence immunoassay** "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

Summary

Cyclosporine is a cyclic undecapeptide of fungal origin and a potent immunosuppressive agent. The introduction of cyclosporine in human kidney transplantation in the late 1970s was a major step forward in transplantation medicine, and substantially improved patient and graft survival in patients receiving e.g. heart, kidney, liver, pancreas, lung or bone marrow transplants.^{1,2,3}

Cyclosporine was the first drug identified to specifically and reversibly inhibit the activation and proliferation of lymphocytes and represents the prototype of a class of drugs called calcineurin inhibitors.⁴

The main mechanism through which cyclosporine exerts its immunosuppressive effect is believed to be via the inhibition of T cell activation and proliferation. Intracellular cyclosporine binds to both cyclophilin A and B and these complexes then inhibit the enzymatic activity of calcineurin.^{3,5,6}

The inhibition of calcineurin restricts the dephosphorylation and nuclear translocation of nuclear factor of activated T cells (NFAT), which regulates transcription of several cytokines, including IL-2, IL-4, TNF- α , and interferon- γ , and therefore limits lymphocyte activation and proliferation.^{7,8,9,10,11,12}

Cyclosporine is highly lipophilic and absorption from the gastrointestinal tract is incomplete and variable. Approximately 90 % of the cyclosporine within the plasma is bound to proteins.¹³

The bioavailability and metabolism of cyclosporine are predominantly influenced by the activity of the cytochrome P450 isozymes CYP3A4 and CYP3A5, as well as the efflux pump p-glycoprotein, which show significant inter- and intra-individual variability in expression and function.^{14,15,16}

Cyclosporine displays significant inter- and intra-patient pharmacokinetic variability, as well as potentially severe side effects from doses that are either too low or too high. Inadequate cyclosporine concentrations might result in rejection of the transplanted organ. High levels may lead to severe adverse effects. The most significant and well recognized side effect of cyclosporine therapy is nephrotoxicity, which can manifest as both reversible acute manifestations and as irreversible chronic manifestations.^{3,17} The use of cyclosporine is also associated with renal dysfunction, tremor, hirsutism, hypertension, and gum hyperplasia.¹³

Therefore the application of therapeutic drug monitoring (TDM) and concentration-controlled dosing in order to maintain each patient's drug exposure within a narrow therapeutic window is clearly required and part of standard clinical practice for many years.^{18,19,20}

Monitoring is most effective when there is a measurement that is a good surrogate for total drug exposure (measured as area under the time-concentration curve AUC 0-12). Advantages of monitoring cyclosporine concentrations based on predose trough (C0) concentrations versus two hours after administration (C2) are still in discussion and more multicenter studies are required to demonstrate a clinical benefit of C2 monitoring.^{18,21}

Test principle

Manual precipitation:

Before testing with the Elecsys Cyclosporine assay, samples, calibrators and controls must be **pretreated** with Elecsys ISD Sample Pretreatment.

The reagent lyses the cells, extracts cyclosporine, and precipitates most of the blood proteins. The **pretreated** samples are centrifuged, and an aliquot of the resulting supernatant containing cyclosporine is then assayed using the Elecsys Cyclosporine assay.

Competition principle. Total duration of assay: 18 minutes.

- 1st incubation: 20 μ L of pretreated sample is incubated with a cyclosporine-specific biotinylated antibody and a ruthenium complex^{a)} labeled cyclosporine-derivate. Depending on the analyte concentration in the sample and the formation of the respective immune complex, the labeled antibody binding site is occupied in part with sample analyte and in part with ruthenylated hapten.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Reagents - working solutions

The reagent rackpack is labeled as CSA.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-cyclosporine Ab-biotin (gray cap), 1 bottle, 9 mL:
Biotinylated monoclonal anti-cyclosporine antibody (mouse) 25 μ g/L;
phosphate buffer 50 mmol/L, pH 6.0; preservative.
- R2 Cyclosporine~Ru(bpy)₃²⁺ (black cap), 1 bottle, 9 mL:
Cyclosporine labeled with ruthenium complex 5 μ g/L; phosphate
buffer 50 mmol/L, pH 6.0; preservative.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date

Cyclosporine

Cyclosporine



Stability:	
after opening at 2-8 °C	84 days
on the analyzers	56 days

Specimen collection and preparation

Only the specimens listed below were tested in a sufficient number and found acceptable.

K₂- and K₃-EDTA whole blood.

Sample concentration ranges from trough (C0) to peak (C2) samples are suitable.

Specimens collected in EDTA tubes may be stored for up to 5 days at 15-25 °C or 7 days at 2-8 °C prior to being tested. If testing will be delayed by more than 7 days, store frozen at -20 °C or lower for up to 6 months.

Freeze only once. Specimens must be mixed thoroughly after thawing to ensure consistency of the results.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Mix thawed specimens thoroughly by hand or on a roller mixer or rocker. Visually inspect the specimens. If layering or stratification is observed, continue mixing until specimens are visibly homogeneous.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to pretreatment.

Use caution when handling patient specimens to prevent cross contamination. Use of disposable pipettes or pipette tips is recommended.

Pretreated samples can be stored in closed tubes for up to 4 hours at 20-25 °C.

Due to evaporation effects, pretreated samples should be analyzed/measured within 30 minutes after opening the vials and loading the samples on the analyzer. Avoid delays between loading and measurement to ensure the 30 minute stability of pretreated samples.

A re-run requires repeating of the manual pretreatment procedure.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- [REF] 05889073190, ISD Sample Pretreatment, 1 x 30 mL
- [REF] 05889022190, Cyclosporine CalSet, for 6 x 1 mL
- [REF] 05889081190, PreciControl ISD, for 3 x 3 mL each of PreciControl ISD 1, 2 and 3
- [REF] 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF] 03183971122, Diluent Universal, 2 x 36 mL sample diluent
- [REF] 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
- General laboratory equipment
- Precision pipettes (use only positive displacement pipettes for ISD Sample Pretreatment reagent handling)
- Microcentrifuge tubes (2.0 mL capacity)
- Microcentrifuge (at least 10000 g)
- Vortex mixer
- Roller mixer or rocker
- Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** analyzer

Accessories for Elecsys 2010 and **cobas e** 411 analyzers:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean

- [REF] 11706802001, Elecsys 2010 AssayCup, 60 x 60 reaction vessels

- [REF] 11706799001, Elecsys 2010 AssayTip, 30 x 120 pipette tips

Accessories for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
- [REF] 12102137001, AssayTip/AssayCup Combimagazine M, 48 magazines x 84 reaction vessels or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Accessories for all analyzers:

- [REF] 11298500316, Elecsys SysClean, 5 x 100 mL system cleaning solution

Manual specimen pretreatment

Follow the steps listed below to pretreat calibrators, controls and/or specimens. **The technical notes are an essential part of the instructions and must be read thoroughly before completing each step.** Follow Steps 1 through 7 to pretreat calibrators, controls and/or specimens.

Steps	Technical notes
1. Equilibrate all reagents, calibrators, controls and specimens to 20-25 °C. Mix all calibrators, controls and specimens gently but thoroughly just before use.	Do not vortex. The liquids may be mixed by hand or on a roller mixer or rocker. The calibrators and controls are a whole-blood hemolysate and may be slightly different in appearance from whole-blood samples.
2. Label one microcentrifuge tube for each calibrator, control and/or specimen to be pretreated.	none
3. Using a precision pipette, transfer 300 µL of each calibrator, control and/or specimen to the appropriately labeled micro-centrifuge tube.	Use a fresh pipette tip for each calibrator, control and/or specimen.
4. Using a precision pipette, add 300 µL of ISD Sample Pretreatment reagent to each microcentrifuge tube. Immediately cap each tube and immediately proceed to step 5.	Note: ISD Sample Pretreatment is highly volatile. Keep tightly closed when not in use to prevent evaporation.
5. Vortex each microcentrifuge tube for at least 10 seconds. Failure to perform this step may result in a supernatant that appears red. See Step 6, technical note.	Note: Failure to vortex each tube immediately after addition of the ISD Sample Pretreatment reagent will lead to erroneous assay results. Sample and reagent mixture should be completely homogeneous immediately after vortexing. Visual inspection is required.

Cyclosporine

Cyclosporine

Steps	Technical notes
6. Centrifuge the samples for at least 4 minutes in a micro-centrifuge (≥ 10000 g).	The centrifuged samples should have well-defined pellets and clear supernatant. The supernatant should not appear cloudy or red. If the supernatant is red, discard and replace it with a newly extracted sample.
7. Transfer each supernatant directly into an appropriate vial and immediately cap each vial. The samples are ready to be assayed.	<p>Pretreated samples can be stored in closed tubes for up to 4 hours at 20-25 °C.</p> <p>Please note: Due to evaporation effects, pretreated samples should be analyzed/measured within 30 minutes after opening the vials and loading the samples on the system. Avoid delays between loading and measurement to ensure the 30 minutes stability of pretreated samples.</p> <p>This is supported by running the cyclosporine samples in batch mode: Based on average system sample processing time, no more than 35 cyclosporine samples may be loaded per calibrated measuring cell onto the analyzers at the same time.</p>

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

MODULAR ANALYTICS E170, **cobas e 601** and **cobas e 602** analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

Calibration

Traceability: This method has been standardized against reference standards traceable to cyclosporine reference material by weight.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

Cyclosporine CalSet must be pretreated freshly before calibration.

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer). Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

Quality control

For quality control, use PreciControl ISD.

PreciControl ISD must be pretreated freshly before measurement.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in ng/mL, nmol/L, µg/L).

Conversion factors: $\text{ng/mL} \times 1.0 = \mu\text{g/L}$
 $\text{ng/mL} \times 0.832 = \text{nmol/L}$

Limitations - interference

The effect of the following endogenous substances, pharmaceutical compounds and clinical conditions on assay performance was tested. Interferences were tested up to the listed concentrations and no impact on results was observed.

Criterion: Recovery within ± 18.0 ng/mL (concentration range ≤ 90.0 ng/mL) or within $\pm 20\%$ (concentration range > 90.0 ng/mL) of initial value.

Endogenous substances:

Compound	Concentration tested
Albumin	≤ 12.0 g/dL
Bilirubin	≤ 1026 µmol/L or ≤ 60.0 mg/dL
Biotin	< 30.0 ng/mL or < 123 nmol/L
Cholesterol	≤ 500 mg/dL
HAMA	≤ 50.0 µg/mL
Hematocrit	15-60 %
IgG	≤ 12.0 g/dL
Intralipid	≤ 1500 mg/dL
Rheumatoid factors	up to 500 IU/mL
Uric acid	≤ 20.0 mg/dL

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. > 5 mg/day) until at least 8 hours following the last biotin administration.

Pharmaceutical compounds:

In vitro tests were performed on 15 commonly used pharmaceutical compounds. No interference with the assay was found.

Criterion: Recovery within ± 18.0 ng/mL (concentration range ≤ 90.0 ng/mL) or within $\pm 20\%$ (concentration range > 90.0 ng/mL) of initial value.

27 special drugs were additionally tested. An interaction with Itraconazole (INN international nonproprietary name) was found. Do not use samples from patients under Itraconazole treatment.

Drug	Concentration tested
Acyclovir	3.2 µg/mL
Amphotericin B	5.8 µg/mL
Ciprofloxacin	7.4 µg/mL
K ₂ -EDTA	6 mg/mL
K ₃ -EDTA	6 mg/mL
Erythromycin	20 mg/dL
Everolimus	60 ng/mL
Fluconazole	30 µg/mL
Flucytosine	40 µg/mL
Gancyclovir	1000 µg/mL
Gentamicin	12 mg/dL

Cyclosporine

Cyclosporine



Drug	Concentration tested
Itraconazole	50 µg/mL
Kanamycin	100 µg/mL
Ketoconazole	50 µg/mL
Lidocaine	6 mg/dL
MPA (mycophenolic acid) glucuronide	1800 µg/mL
Mycophenolic acid	500 µg/mL
Nitrofurantoin	6 µg/mL
Phenobarbital	15 mg/dL
Rifampicin	5 mg/dL
Sirolimus	60 ng/mL
Spectinomycin	100 µg/mL
Sulfomethoxazole	200 µg/mL
Tacrolimus	60 ng/mL
Tobramycin	2 mg/dL
Trimethoprim	40 µg/mL
Vancomycin	6 mg/dL

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Limits and ranges

Measuring range

30.0-2000 ng/mL (defined by the Limit of Detection and the maximum of the master curve). Values below the Limit of Detection are reported as < 30.0 ng/mL. Values above the measuring range are reported as > 2000 ng/mL.

Lower limits of measurement

Limit of Blank (LoB), Limit of Detection (LoD) and Limit of Quantitation (LoQ)

Limit of Blank = 20.0 ng/mL

Limit of Detection = 30.0 ng/mL

Limit of Quantitation = 50.0 ng/mL with a total allowable error of ≤ 20 %

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A requirements.

The Limit of Blank is the 95th percentile value from $n \geq 60$ measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is defined as the lowest amount of analyte in a sample that can be accurately quantitated with a total allowable error of ≤ 20 %.

Dilution

Samples with cyclosporine concentrations above the measuring range can be manually diluted 1:3 with Diluent Universal prior to the manual pretreatment procedure. The concentration of the diluted sample must be > 500 ng/mL.

After manual dilution, multiply the result by the dilution factor.

Expected values

No firm therapeutic range exists for cyclosporine in whole blood. The complexity of the clinical state, individual differences in sensitivity to immunosuppressive and nephrotoxic effects of cyclosporine,

coadministration of other immunosuppressants, type of transplant, time post-transplant, and a number of other factors contribute to different requirements for optimal blood levels of cyclosporine. Individual cyclosporine values cannot be used as the sole indicator for making changes in the treatment regimen. Each patient should be thoroughly evaluated clinically before treatment adjustments are made, and each assay user must establish his or her ranges based on clinical experience.

These ranges will vary according to the commercial in vitro diagnostic test used. Ranges must be established for each commercial test used.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents, samples and controls in a protocol (EP5-A2) of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplication each for 21 days ($n = 84$). The following results were obtained:

Elecsys 2010 and cobas e 411 analyzers					
		Repeatability		Intermediate precision	
Sample	Mean ng/mL	SD ng/mL	CV %	SD ng/mL	CV %
HSP ^{b)} 1	63.3	2.45	3.9	5.80	9.2
HSP 2	146	4.57	3.1	10.2	7.0
HSP 3	391	13.9	3.5	23.3	6.0
HSP 4	951	29.4	3.1	44.8	4.7
HSP 5	1830	59.1	3.2	89.9	4.9
PC ^{c)} ISD1	65.0	1.96	3.0	4.87	7.5
PC ISD2	317	7.81	2.5	13.3	4.2
PC ISD3	1210	39.2	3.2	53.8	4.4

b) HSP = Human Sample Pool

c) PC = PreciControl

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers					
		Repeatability		Intermediate precision	
Sample	Mean ng/mL	SD ng/mL	CV %	SD ng/mL	CV %
HSP 1	64.0	2.75	4.3	4.12	6.4
HSP 2	146	4.36	3.0	7.29	5.0
HSP 3	400	12.5	3.1	19.1	4.8
HSP 4	973	23.0	2.4	40.4	4.2
HSP 5	1820	48.2	2.6	105	5.8
PC ISD1	69.0	2.82	4.1	3.63	5.3
PC ISD2	326	6.45	2.0	10.1	3.1
PC ISD3	1230	38.3	3.1	53.8	4.4

Method comparison

a) A comparison of the Elecsys Cyclosporine assay (y) with an automated immunoassay (x) using clinical samples gave the following correlations:

Number of samples measured: 339

Passing/Bablok²²

$y = 1.01x - 15.5$

$r = 0.857$

Weighted linear regression

$y = 0.946x - 8.95$

$r = 0.977$

The sample concentrations were between approximately 30.7 and 1770 ng/mL.

b) A comparison of the Elecsys Cyclosporine assay (y) with an LC-MS-MS method (x) using clinical samples gave the following correlations:

Cyclosporine

Cyclosporine

Number of samples measured: 352

Passing/Bablok²²

$$y = 1.091x + 2.08$$

$$r = 0.900$$

Weighted linear regression

$$y = 1.092x + 1.87$$

$$r = 0.997$$

The sample concentrations were between approximately 30.7 and 1912 ng/mL.

Analytical specificity

A study was performed with the Elecsys Cyclosporine assay based on guidance from the CLSI document EP7-A2.

Metabolite	Maximum concentration of metabolite added ng/mL	Cross-reactivity %
AM1	2000	2
AM19	2000	n. d. ^{d)}
AM1c	2000	n. d.
AM1c9	2000	n. d.
AM4n	2000	2
AM9	2000	6

d) n. d. = not detectable

Cross-reactivity was designated as „not detectable“ if the obtained value was less than the sensitivity of the assay.

References

- Kahan BD. Cyclosporine. *New Engl J Med* 1989;321:1725-1738.
- Kahan BD. Cyclosporine: a revolution in transplantation. *Transplant Proc* 1999;31(1-2A):14S-15S.
- Naesens M, Kuypers DRJ, Sarwal M. Calcineurin inhibitor nephrotoxicity. *Clin J Am Soc Nephrol* 2009;4:481-508.
- Borel JF, Feurer C, Gubler HU, et al. Biological effects of cyclosporine A: a new antilymphocytic agent. *Agents Actions* 1994;6:468-475.
- Takahashi N, Hayano T, Suzuki M. Peptidyl-prolyl cis-trans isomerase is the cyclosporine A-binding protein cyclophilin. *Nature* 1989;337:473-475.
- Fischer G, Wittmann-Liebold B, Lang K, et al. Cyclophilin and peptidyl-prolyl cis-trans isomerase are probably identical proteins. *Nature* 1989;337:476-478.
- Flanagan WM, Corthesy B, Bram RJ, et al. Nuclear association of a T-cell transcription factor blocked by FK-506 and cyclosporin A. *Nature* 1991;352:803-807.
- Jain J, McCaffrey PG, Miner Z, et al. The T-cell transcription factor NFATp is a substrate for calcineurin and interacts with Fos and Jun. *Nature* 1993;365:352-355.
- Shaw KT, Ho AM, Raghavan A, et al. Immunosuppressive drugs prevent a rapid dephosphorylation of transcription factor NFAT1 in stimulated immune cells. *Proc Natl Acad Sci USA* 1995;92:11205-11209.
- Clipstone NA, Crabtree GR. Identification of calcineurin as a key signalling enzyme in T-lymphocyte activation. *Nature* 1992;357:695-697.
- O'Keefe SJ, Tamura J, Kincaid RL, et al. FK-506 and CsA-sensitive activation of the interleukin-2 promoter by calcineurin. *Nature* 1992;357:692-694.
- Emmel EA, Verweij CL, Durand DB, et al. Cyclosporin A specifically inhibits function of nuclear proteins involved in T cell activation. *Science* 1989;246:1617-1620.
- Novartis. Sandimmune Package Insert. Available at: http://www.accessdata.fda.gov/drugsatfda_docs/label/2010/050625s048,050573s034,050574s04-2lbl.pdf [Last accessed 01/03/2012].
- Cummins CL, Jacobsen W, Benet LZ. Unmasking the dynamic interplay between intestinal P-glycoprotein and CYP3A4. *J Pharmacol Exp Ther* 2002;300:1036-1045.

- Benet LZ, Cummins CL, Wu CY. Unmasking the dynamic interplay between intestinal efflux transporters and metabolic enzymes. *Int J Pharm* 2004;277:3-9.
- Anglicheau D, Legendre C, Beaune P, et al. Cytochrome P450 3A polymorphisms and immunosuppressive drugs: an update. *Pharmacogenomics* 2007;8(7):835-849.
- Myers BD, Ross J, Newton L, et al. Cyclosporine-associated chronic nephropathy. *N Engl J Med* 1984;311(11):699-705.
- Oellerich M, Armstrong VW, Schütz E, et al. Therapeutic drug monitoring of cyclosporine and tacrolimus. Update on Lake Louise Consensus Conference on cyclosporine and tacrolimus. *Clin Biochem* 1998;31:309-316.
- Kahan, BD. Therapeutic drug monitoring of cyclosporine: 20 years of progress. *Transplant Proc* 2004;36:378S-391S.
- Schiff J, Cole E, Cantarovich M. Therapeutic monitoring of calcineurin inhibitors for the nephrologist. *Clin J Am Soc Nephrol* 2007;2:374-384.
- Zhang Y, Zhang XD, Wang Y. Efficacy and safety of changing from cyclosporine C0 to C2 monitoring in stable recipients following renal transplantation: a prospective cohort study. *Transplant Proc* 2011;43(10):3697-3701.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
→	Volume after reconstitution or mixing

COBAS, COBAS E, ELECSYS, MODULAR and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

