

No.	Component	Requested parameter	Quantity	Name of the offered product, name of the manufacturer, compliance with the technical requirements.
1.	Invitation screen (in posts)	<p>Operating principle:</p> <p>1.1. The invitation screen must receive signals from call centers in the wards without the use of any interconnected wires.</p> <p>1.2. At least 4 invitations are displayed at a time.</p> <p>Operating distance:</p> <p>1.3. Buildings must be located at least 100 meters away.</p> <p>1.4. It must be possible to extend the operating range by using additional equipment.</p> <p>Screen:</p> <p>1.5. A touch-sensitive display that informs you of a call and lets you control the entire system.</p> <p>1.6. The screens are wall mounted and must be supplied with wall brackets.</p> <p>1.7 Screen size at least 9.7 inches.</p> <p>Certificates:</p> <p>1.8 The device must have a CE declaration of conformity.</p>	3	<p>Yes</p> <p>Yes. More then 4</p> <p>Yes. More then 100 m</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes. 9.7 inches</p> <p>Yes</p>
2.	Call panel (for patients)	<p>Operating distance:</p> <p>2.1. Buildings must be located at least 100 meters away.</p> <p>Housing:</p> <p>2.2. The body of the wireless call panel must be made of a durable plastic coated antibacterial coating that can be cleaned with antibacterial cleaners.</p> <p>Standards:</p> <p>2.3. Wireless call centers must comply with the applicable European Union safety standards for radio equipment and must have a CE declaration of conformity.</p> <p>Feed:</p> <p>2.4. The wireless call panel must be powered by standard batteries.</p> <p>Attachment:</p> <p>2.5. Wireless call panels must be equipped with wall brackets that allow the panel itself to be mounted on the wall without the use of any additional tools.</p> <p>2.6. The kit should include a twisted extension cord to allow the patient to hold the call button in their hand.</p>	38	<p>Yes. More then 100</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>

				Yes
3.	Call panel (for toilets / showers)	<p>Operating distance: 3.1. Buildings must be located at least 100 meters away.</p> <p>Housing: 3.2. The body of the wireless call panel must be made of a durable plastic coated antibacterial coating that can be cleaned with antibacterial cleaners.</p> <p>3.3. The call is initiated by pulling the call cord.</p> <p>Standards: 3.4. Wireless call centers must comply with the applicable European Union safety standards for radio equipment and must have a CE declaration of conformity.</p> <p>Feed: 3.5. The wireless call panel must be powered by standard batteries.</p> <p>Attachment: 3.6. Call panels must be mounted on the wall.</p>	17	<p>Yes. More then 100 m</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>
4.	System software	<p>Program features: 4.1. The software must be installed on a server owned by the institution.</p> <p>4.2. Must be supported: 4.2.1 at least 800 call centers; 4.2.2 at least 1000 users (employees); 4.2.3 not less than 800 beds; 4.2.4 at least 500 wards.</p> <p>4.3. It must be possible to change the number of any panel in the program, assign a ward number.</p> <p>4.4. Ability to remotely administer call panels and screens.</p> <p>4.5. Ability to initiate a call for hospital staff on the Buyer's computer and on the invitation screen.</p> <p>4.6. The calls must be displayed in the invitation screen and at the same time an acoustic signal must be activated, the level and sound type of which must be adjustable.</p> <p>4.7. It must be possible to invite (transmit) a message to at least 1000 users.</p> <p>4.8 The Buyer must be provided with a mobile application for informing the hospital staff of the summons, the invited user must</p>	1	<p>Yes</p> <p>Yes</p> <p>Yes. More then 800</p> <p>Yes. More then 1000 users</p> <p>Yes. More then 800 beds</p> <p>Yes. More then 500 wards</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes. More then 1000 users</p> <p>Yes</p>

		<p>see the announcement of the summons indicating the time and place of the summons.</p> <p>4.9. Call time and response to the message should be recorded and reported.</p> <p>4.10. Reporting by wards, posts, call and waiting times must be possible;</p> <p>4.11. Notices of malfunctioning hardware must be displayed on tablets and / or computer</p>		<p>Yes</p> <p>Yes</p> <p>Yes</p>
5.	Corridor Lights	<p>5.1. The supplier shall install corridor lights above each ward for light and audible alarm.</p> <p>5.2. Colors: red, green.</p> <p>5.3. Power supply at least 12V.</p>	18	<p>Yes</p> <p>Yes</p> <p>Yes. 24 V.</p>
6.	Ward call and cancel call control	<p>Operating distance:</p> <p>6.1. Buildings must be located at least 100 meters away.</p> <p>Housing:</p> <p>6.2. The body of the wireless call panel must be made of a durable plastic coated antibacterial coating that can be cleaned with antibacterial cleaners.</p> <p>Standards:</p> <p>6.3. Wireless call centers must comply with the applicable European Union safety standards for radio equipment and must have a CE declaration of conformity.</p> <p>Feed:</p> <p>6.4. The wireless call panel must be powered by standard batteries.</p> <p>Functions:</p> <p>6.5. The red LED button activates the call.</p> <p>6.6. The green LED button cancels the call</p>	18	<p>Yes. More then 100 m</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>
7.	Source of audible signal	<p>7.1 The personnel call shall be accompanied by an audible signal.</p> <p>7.2. Sound sources must be installed in premises agreed with the customer.</p> <p>7.3. It must be possible to adjust the volume.</p>	6	<p>Yes</p> <p>Yes</p> <p>Yes</p>

Confirmed by manufacturer:
EUROTRONIK Kranj d. o. o.
 Stirnova 8, 4000 Kranj,
 Slovenia

EUROTRONIK
 Eurotronik Kranj d.o.o.
 Stirnova 8, SI-4000 Kranj, Slovenia
 T: +386 4 2 350 350
 F: +386 4 2 350 351
 info@eurotronik.si
 www.eurotronik.si

Kranj